

Laboratory Central Laboratory, Hasdeo Area, SECL, B-Seam Mine, P.O.- West Jhagrakhand, Dist.- Korla, Chhattisgarh

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number TC-7887 (in lieu of T-4264)

Page 1 of 1

Validity 23.11.2018 to 22.11.2020

Last Amended on --

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
-----	----------------------------	-------------------------	---	--

CHEMICAL TESTING

I.	SOLID FUELS			
1.	Coal	Proximate Analysis		
		Ash	IS 1350 (Part 1)	1 % to 50 %
		Moisture	IS 1350 (Part 1)	0.5 % to 40 %
		Volatile Matter	IS 1350 (Part 1)	0.5 % to 40 %
		Gross calorific Value.	ASTM D 5865	1000 kcal/kg to 8300 kcal/kg