

Laboratory Meter Testing Laboratory, Torrent Power Limited, Torrent House,
Station Road, Badri Road, Surat, Gujarat

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number TC-5178

Page 1 of 2

Validity 22.07.2019 to 12.01.2021

Last Amended on --

"In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020"

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
-----	----------------------------	-------------------------	---	--

ELECTRICAL TESTING

I.	MEASURING INSTRUMENTS			
1.	AC Static Energy Meter Class 0.2S, 0.5S, 1 & 2	Limits of error	IS 13779 IS 14697	40V to 300 V 10 mA to 100 A 45Hz to 55 Hz 1 to 0.5 lead / lag
		No load condition	IS 13779/IS 14697	40V to 300V
		Influence Quantities Frequency Variation	IS 13779 IS 14697	40V to 300V 10 mA to 100 A 45Hz to 55 Hz 1 to 0.5 lead / lag
		Influence Quantities Reverse Phase Sequence	IS 13779 IS 14697	40V to 300V 10 mA to 100 A 45Hz to 55 Hz 1 to 0.5 lead / lag
		Influence Quantities Voltage Unbalance	IS 13779 IS 14697	40V to 300 V 10 mA to 100 A 45Hz to 55Hz 1 to 0.5 lead / lag
		Influence Quantities Voltage Variation	IS 13779 IS 14697	40V to 300 V 10mA to 100A 45Hz to 55Hz 1 to 0.5 lead / lag
		Meter Constant	IS 13779 IS 14697	40V to 300 V 10 mA to 100 A 45Hz to 55 Hz 1 to 0.5 lead / lag
		Repeatability of Error	IS 13779 IS 14697	40V to 300 V 10 mA to 100 A 45Hz to 55 Hz 1 to 0.5 lead / lag

Prachi Kukreti
Convenor

Mallika Gope
Program Manager

Laboratory Meter Testing Laboratory, Torrent Power Limited, Torrent House,
Station Road, Badri Road, Surat, Gujarat

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number TC-5178

Page 2 of 2

Validity 22.07.2019 to 12.01.2021

Last Amended on --

"In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020"

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
		Starting current	IS 13779 IS 14697	40V to 300 V 10 mA to 100 A 45Hz to 55 Hz 1 to 0.5 lead / lag

Prachi Kukreti
Convenor

Mallika Gope
Program Manager