

Laboratory **Power System Testing Laboratory, C.T.I. Nayagoan, Rampur,
Jabalpur, Madhya Pradesh**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **TC-7657**

Page 1 of 1

Validity **06.08.2018 to 05.08.2020**

Last Amended on --

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
-----	----------------------------	-------------------------	---	--

ELECTRICAL TESTING

I.	ELECTRICAL INDICATING & RECORDING INSTRUMENTS			
1.	AC Static Energy Meter	Limits of Error	IS 13779:1999 Clause no. 11.1	40 mV to 600 V 1 mA to 120 A 50 Hz±2.5

Gaurav Saini
Convenor

Anuja Anand
Program Manager