

Laboratory **Micro Small and Medium Enterprises Testing Station, MSME DI
Campus, Government of India, Ministry of MSME, Rajajinagar,
Industrial Estate, Bangalore, Karnataka**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **TC-7939**

Page 1 of 3

Validity **26.09.2018 to 25.09.2020**

Last Amended on --

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
-----	----------------------------	-------------------------	---	--

ELECTRICAL TESTING

I.	CABLES AND ACCESSORIES				
1.	PVC Insulated Cables for Working Voltages Upto and Including 1100 V	Annealing - Copper	IS 694:2010/IS 5831:1984 IS 8130:2013 IS 10810 (Part 1): 1984	100 N to 2500 N 20 mm to 200 mm	
		Tensile - Aluminium	IS 10810 (Part 2): 1984	100 N to 2500 N	
		Wrapping - Aluminium	IS 10810 (Part 3): 1984	Qualitative	
		Conductor Resistance - Copper & Aluminium	IS 10810 (Part 5): 1984	1 mΩ to 1800 Ω	
		Overall Dimension and thickness of Insulation	IS 10810 (Part 6): 1984	0.02 mm to 25 mm	
		Physical test for Insulation & Sheath			
		- Tensile Strength - Elongation at break	IS 10810 (Part 7): 1984	100 N to 2500 N 20 mm to 200 mm 1 % to 300 %	
		Loss of Mass	IS 10810 (Part 10): 1984	25 °C to 200 °C 10 mg to 200 g	
		Ageing in air oven	IS 10810 (Part 11): 1984	25 °C to 200 °C 20 mm to 200 mm	
		Shrinkage	IS 10810 (Part 12): 1984	25 °C to 200 °C 20 mm to 300 mm	
		Heat Shock	IS 10810 (Part 14): 1984	25 °C to 200 °C	
		Hot-deformation	IS 10810 (Part 15): 1984	25 °C to 200 °C 0.02 mm to 25mm	
		Insulation Resistance	IS 10810 (Part 43): 1984	Upto 200 GΩ at 500 V _{dc}	
High Voltage - Water Immersion	IS 10810 (Part 45): 1984	0.5 kV to 10 kV Ambient to 80 °C/240 hrs			
Flammability	IS 10810 (Part 53): 1984	10 mm to 1000 mm			

Laboratory

Micro Small and Medium Enterprises Testing Station, MSME DI
Campus, Government of India, Ministry of MSME, Rajajinagar,
Industrial Estate, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number TC-7939

Page 2 of 3

Validity 26.09.2018 to 25.09.2020

Last Amended on --

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection	
2.	PVC Insulated (Heavy Duty) Electric Cables	Annealing - Copper	IS 1554 (Part 1): 1988 IS 8130: 2013 IS 10810 (Part 2): 1984	100 N to 2500 N 20 mm to 200 mm	
		Tensile - Aluminium	IS 10810 (Part 3): 1984	100 N to 2500 N	
		Wrapping - Aluminium	IS 10810 (Part 4): 1984	Qualitative	
		Conductor Resistance - Copper & Aluminium	IS 10810 (Part 5): 1984	1 mΩ to 1800 Ω	
		Thickness of Insulation and sheath	IS 10810 (Part 6): 1984	0.02 mm to 25 mm	
		Armoring wires strips	IS 3975: 1999 IS 5831: 1984	Qualitative	
		Tensile	IS 10810 (Part 37): 1984	100 N to 2500 N	
		Torsion on Galvanized steel	IS 10810 (Part 38): 1984	Qualitative	
		Resistivity at 20 °C for Armour wires - Round wires only	IS 10810 (Part 42): 1984	0.1X10 ⁶ Ω-cm to 14.5X10 ⁶ Ω-cm	
		Mass of Zinc coating on steel armour	IS 10810 (Part 41): 1984	10 mg to 200 g	
		Uniformity of Zinc coating on steel armour	IS 10810 (Part 40): 1984	Qualitative	
		Dimension, thickness & width of armour	IS 10810 (Part 36): 1984	0.02 mm to 200 mm	
		Physical Test for Insulation and Outer Sheath			
			- Tensile strength - Elongation at break	IS 10810 (Part 7): 1984	100 N to 2500 N 1 % to 300 %
	Loss of Mass	IS 10810 (Part 10): 1984	25 °C to 200 °C 10 mg to 200 g		
	Ageing in air oven - Tensile strength - Elongation	IS 10810 (Part 11): 1984	25 °C to 200 °C 50 N to 2500 N		

Sreeram Pinnamaraju
Convenor

Alok Jain
Program Manager

Laboratory **Micro Small and Medium Enterprises Testing Station, MSME DI
Campus, Government of India, Ministry of MSME, Rajajinagar,
Industrial Estate, Bangalore, Karnataka**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **TC-7939**

Page 3 of 3

Validity **26.09.2018 to 25.09.2020**

Last Amended on --

SI.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
		Shrinkage	IS 10810 (Part 12): 1984	25 °C to 200 °C 0.2 mm to 300 mm
		Heat shock	IS 10810 (Part 14): 1984	25 °C to 200 °C
		Hot deformation	IS 10810 (Part 15): 1984	25 °C to 200 °C 0.02 mm to 25 mm
		Insulation Resistance	IS 10810 (Part 43): 1984	1 MΩ to 200 GΩ at 500 V _{dc}
		High Voltage -Water Immersion	IS 10810 (Part 45): 1984	0.5 kV to 10 kV Ambient to 80 °C
		Flammability	IS 10810 (Part 53): 1984	20 mm to 1000 mm 10 mg to 200 g