

Laboratory Diligence Bio Pvt. Ltd., R.S. No. 28/2, Voluntariat Road, Thuthipet, Pondicherry

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number TC-8062

Page 1 of 1

Validity 03.11.2018 to 02.11.2020

Last Amended on --

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
-----	----------------------------	-------------------------	---	--

**BIOLOGICAL TESTING**

I.	MEDICAL ACCESSORIES & SURGICAL PRODUCTS			
1.	<b>Toxicology</b>			
a.	<b>Medical Devices (Catheter, Syringes, Needles, Surgical Gloves, Elastic Bandages, Intraocular Lenses, Implants &amp; Vascular Devices) Agrochemicals &amp; Pesticides</b>	AMES Test	OECD 471 Adopted 21 <sup>st</sup> July 1997	Qualitative (Background lawn observed or Background lawn not observed)  3 cfu/plate to 400 cfu/plate