

Laboratory Central Coal and Gas Laboratory, South Eastern Coalfields Ltd.,
Chirimiri Area, Haldibadi, Chirimiri, Dist. Koriya, Chhattisgarh

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number TC-8152

Page 1 of 1

Validity 30.11.2018 to 29.11.2020

Last Amended on --

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
-----	----------------------------	-------------------------	---	--

CHEMICAL TESTING

I.	SOLID FUELS			
1.	Coal	Moisture (Air Dry Basis)	IS 1350 (Part 1)	1.0 % to 20 %
		Ash	IS 1350 (Part 1)	5.0 % to 60.0 %
		Volatile Matter	IS 1350 (Part 1)	10.0 % to 50.0 %
		Gross Calorific Value (GCV)	IS 1350 (Part 2)	1000 kcal/kg to 8000 kcal/kg

Naveen Jangra
Convenor

Anuja Anand
Program Manager