Laboratory N.G. Logistics & Quality Coal Co., MHADA Colony, Quarter No. L-96,

Chandrapur, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number TC-8239 Page 1 of 1

Validity 01.01.2019 to 31.12.2020 Last Amended on --

"In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020"

SI.	Product / Material	Specific Test Performed	Test Method Specification	Range of Testing /
	of Test	-	against which tests are	Limits of Detection
			performed	

CHEMICAL TESTING

I.	SOLID FUELS			
1.	Coal	Moisture	IS 1350 (Part 1)	1.0 % to 20.0 %
		Ash	IS 1350 (Part 1)	5.0 % to 52.0 %
		Volatile Matter	IS 1350 (Part 1)	10.0 % to 35.0 %
		Total Moisture	IS 1350 (Part 1)	5.0 % to 20.0 %
		Gross Calorific Value (GCV)	IS 1350 (Part 2)	1500 kcal/kg to 7000 kcal/kg

Deepak Kumar Sharma Convenor Anuja Anand Program Manager