

<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>1 of 16</b>

<b>Quantity Measured/ Instrument</b>	<b>Range / Frequency</b>	<b>*Calibration Measurement Capability (<math>\pm</math>)</b>	<b>Remarks</b>
<b>I. DIMENSION</b>			
<b>1. CALIPER<sup>S</sup> (Vernier, Dial, Electronic) L. C.: 0.005 mm<sup>Φ</sup></b>			
	Upto 600 mm	13.0 $\mu$ m	Using Caliper Checker& Length Bar
	Upto 1000 mm	15.0 $\mu$ m	
<b>Gear Tooth Caliper L.C.: 0.02 mm<sup>Φ</sup></b>	Upto 200 mm	18.0 $\mu$ m	
<b>2. HEIGHT GAUGE<sup>S</sup> (Vernier, Dial , Electronic) L.C. 0.01 mm<sup>Φ</sup></b>			
	Upto 600 mm	13.0 $\mu$ m	Using Caliper Checker & Length Bars
	Upto 1000 mm	15.0 $\mu$ m	
<b>3. DEPTH GAUGE<sup>S</sup> (Dial , Digital, Vernier) L.C. 0.01 mm<sup>Φ</sup></b>			
	Upto 300 mm	10.0 $\mu$ m	Using Slip Gauge Set & Holding Fixture
<b>4. EXTERNAL MICROMETER<sup>S</sup> L. C. : 0.001 mm<sup>Φ</sup> L.C.: 0.01 mm</b>			
	0 to 150 mm	1.6 $\mu$ m	Using Slip Gauge Set & Length Bars
	> 150 mm to 300 mm	5.0 $\mu$ m	
	> 300 mm to 600 mm	8.0 $\mu$ m	
	> 600 mm to 1000mm	11.0 $\mu$ m	
<b>5. V- ANVIL MICROMETER<sup>S</sup> L.C. 0.001 mm<sup>Φ</sup></b>			
	Upto 100 mm	6.5 $\mu$ m	Using Cylindrical Setting Masters

<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>2 of 16</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
6. INTERNAL MICROMETER-2 POINTS <sup>S</sup> L.C. : 0.001 mm <sup>Φ</sup>	25mm (Transverse of micrometer head) Upto 2100 mm (With interchangeable setting rod not more than 400mm)	8.9 $\mu$ m 8.0 $\mu$ m	Using Slip Gauge Set Accessories & Caliper Checker
7. INSIDE DIAL CALIPER-2 POINTS <sup>S</sup> L.C. 0.001 mm <sup>Φ</sup>	Upto 150 mm	5.0 $\mu$ m	Using Slip Gauge Set Accessories & Caliper Checker
8. GROOVE MICROMETER <sup>S</sup> L.C. 0.001 mm <sup>Φ</sup>	Upto 150 mm	5.0 $\mu$ m	Using Slip Gauge Set Accessories & Caliper Checker
9. DEPTH MICROMETER <sup>S</sup> L.C. : 0.001 mm <sup>Φ</sup>	Upto 300 mm	7.0 $\mu$ m	Using Slip Gauge Set & Holding Fixture
10. DIAL GAUGE - PLUNGER TYPE <sup>S</sup> L.C. 0.001 mm L.C.0.002 mm L.C. 0.01 mm <sup>Φ</sup> L.C. 0.001 mm <sup>Φ</sup>	Upto 2 mm Upto 5 mm Upto 25 mm Upto 50 mm	1.0 $\mu$ m 2.0 $\mu$ m 3.0 $\mu$ m 2.0 $\mu$ m	Using Calibration Tester & Slip gauge set Using Universal Length Measuring Machine SIP 302m

<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>3 of 16</b>

<b>Quantity Measured/ Instrument</b>	<b>Range / Frequency</b>	<b>*Calibration Measurement Capability (±)</b>	<b>Remarks</b>
11. <b>DIAL GAUGE – LEVER TYPE<sup>S</sup> L.C. 0.001 mm<sup>Φ</sup></b>	Upto 2 mm	1.1 μm	Using Dial Calibration Tester
12. <b>BORE GAUGE<sup>S</sup> (For transmission accuracy check only)</b>	Upto 1 mm	3.0 μm	Using Dial Calibration Tester
13. <b>THICKNESS GAUGE <sup>\$</sup> (Dial &amp; Electronic) L.C. 0.001 mm L.C. 0.01 mm<sup>Φ</sup></b>	Upto 1 mm Upto 10 mm	1.0 μm 3.2 μm	Using Slip Gauge Set
14. <b>PISTOL CALIPER<sup>S</sup> L.C.0.01 mm<sup>Φ</sup></b>	Upto 65 mm	58.0 μm	Using Slip Gauge Set
15. <b>DIAL SNAP GAUGE<sup>S</sup></b>	Upto 200 mm	3.0 μm	Using Slip Gauge Set & Optical Flat
16. <b>COMPARATOR WITH STAND / COMPARATOR DIAL / ELECTRONIC PROBE WITH DRO<sup>\$</sup> L.C. 0.00001 mm<sup>Φ</sup></b>	Upto 25 mm	0.4 μm	Using Slip Gauge Set

<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>4 of 16</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
17. COMPARATOR STAND <sup>s</sup> (FLATNESS OF WORK (Flatness of Work Table)	Upto Dia.100mm	0.3 $\mu\text{m}$	Using Optical Flat
	Upto 200 mm X 200 mm	1.0 $\mu\text{m}$	Using Electronic Comparator
	Upto 300 mm X 300 mm	4.5 $\mu\text{m}$	
18. DIAL CALIBRATION TESTER <sup>s</sup> L.C 0.0001 mm <sup>Φ</sup>	0 to 25 mm	0.75 $\mu\text{m}$	Using Electronic Comparator
19. MEASURING SCALE <sup>s</sup>	Upto 1000 mm	144 $\mu\text{m}$	Using Scale & Tape Calibrator
	1000 mm to 2000 mm	204 $\mu\text{m}$	
20. MEASURING TAPE <sup>s</sup>	Upto 1 m	144 $\mu\text{m}$	Using Scale & Tape Calibrator
	>1 m to 50 m with step of 1 m	$144 \sqrt{\frac{L}{1000}} \mu\text{m}$ Where L is in mm	
21. PI TAPE <sup>s</sup>	>1 m to 50 m with step of 1 m	$58 \sqrt{\frac{L}{1000}} \mu\text{m}$ Where L is in mm	Using Scale & Tape Calibrator
22. MICROMETER HEAD <sup>s</sup> L.C.: 0.0002 mm <sup>Φ</sup>	0 to 25 mm	0.7 $\mu\text{m}$	Using Electronic Comparator

<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>5 of 16</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
23. VERTICAL SINGLE AXIS MEASURING MACHINE / ELECTRONIC HEIGHT GAUGE <sup>s</sup> L.C.: 0.0001mm <sup>Φ</sup>	Upto 1000 mm	$2.5 + \frac{L}{1000} \mu\text{m}$ Where L is in mm	Using Length Bars
24. COATING THICKNESS METER <sup>s</sup>	Upto 2 mm	4.5 $\mu\text{m}$	Using Master Foils
25. ULTRASONIC THICKNESS GAUGE <sup>s</sup>	Upto 200 mm	8.3 $\mu\text{m}$	Using Step Master
26. WELDING FILLET GAUGES <sup>s</sup>	Length parameters Angle parameters	5.6 $\mu\text{m}$ 186 as	Using Vision Measuring Machine
27. FLOATING CARRIAGE DIA. MEASURING MACHINE <sup>s</sup>  L.C. 0.0001 mm <sup>Φ</sup>	0 to 200 mm	1.6 $\mu\text{m}$	Using Cylindrical Setting Master, Electronic Comparator, Optical Flat, Measuring Pin and Slip Gauges
28. HEIGHT MICROMETER / HEIGHT MASTER <sup>s</sup> L.C.0.0005 mm <sup>Φ</sup>	0 to 350 mm	4.5 $\mu\text{m}$	Using Slip gauge blocks
29. LINEAR GLASS SCALE (GRATICULE) / GLASS GRID <sup>s</sup> L.C. 0.001mm <sup>Φ</sup>	Upto 200 mm	4.4 $\mu\text{m}$	Using Vision Measuring Machine (based on edge detection)

<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>6 of 16</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
30. TAPER SCALE / TAPER BORE GAUGE <sup>s</sup>	Upto 60 mm	5.6 $\mu\text{m}$	Using Vision Measuring Machine
31. HEGMAN GAUGE <sup>s</sup>	Upto 1 mm	6.5 $\mu\text{m}$	Using 2 D Height Gauge
32. LENGTH MEASURING MACHINE <sup>s</sup> (Horizontal) L.C 0.0001 mm <sup>o</sup>	Upto 300 mm	1.4 $\mu\text{m}$	Using Slip Gauge Blocks
33. SURFACE ROUGHNESS TESTER <sup>s</sup>	Ra Rz Rmax	7.50% 7.50% 7.50%	Using Surface Roughness Masters
34. WET FILM THICKNESS GAUGE <sup>s</sup>	Upto 5 mm	7.1 $\mu\text{m}$	Using Vision Measuring Machine
35. SCALE & TAPE CALIBRATION MACHINE <sup>s</sup> L.C.0.0001 mm <sup>o</sup>	Upto 1000 mm	6.0 $\mu\text{m}$	Using Slip Gauge Blocks & Length Bars
36. ORIFICE DIAMETER & MAJOR DIAMETER OF VISCOSITY CUP <sup>s</sup>	Upto 50 mm	9.5 $\mu\text{m}$	Using Linear Height Direct Gauge
37. STRAIGHT EDGE <sup>s</sup>	Upto 5000 mm	$2.9 \sqrt{\frac{L}{120}}$ Where L is in mm	Using Electronic Level

**Laboratory**                      **Accurate Calibration & Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra**

**Accreditation Standard**    **ISO/IEC 17025: 2005**

**Discipline**                      **Mechanical Calibration**                      **Issue Date**    **25.09.2014**

**Certificate Number**        **C-0503**    **Valid Until**    **24.09.2016**

**Last Amended on**            **29.09.2014**    **Page**            **7 of 16**

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
38. ENGINEER'S PARALLEL <sup>§</sup> (PARALLELISM)	Upto 350 mm	3.5 µm	Using Electronic Comparator
39. ENGINEER'S SQUARE <sup>§</sup>	Upto 600 mm	8.6 µm (For Squareness)	Using Comparator & Granite-L square
40. MASTER CYLINDER / GRANITE L-SQUARE <sup>§</sup>	Upto 600 mm	3.5 µm (For Squareness)	Using Master Granite L-Square
41. ANGLE PLATE / BOX ANGLE PLATE <sup>§</sup> (Parallelism / Flatness / Squareness)	Upto 300mm	8.6 µm (For Parallelism / Flatness)  8.6 µm (For Squareness)	Using Master Granite L-Square & Comparator For Angle plate
42. V-BLOCK <sup>§</sup> (Parallelism / Symetricity/ Squareness)	Upto 150mm	6.8 µm (For Parallelism / Symmetricity)  8.3 µm (For Squareness)	Using Master Cylinders/ Master Granite L-Square & Comparator
43. GAUGE BLOCK <sup>§</sup>	0.5 mm to 25 mm > 25 mm to 50 mm > 50 mm to 75 mm > 75 mm to 100 mm	0.10 µm 0.11 µm 0.13 µm 0.15 µm	Using Gauge Block Comparator & Reference "K" Grade Gauge Block Set

<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>8 of 16</b>

<b>Quantity Measured/ Instrument</b>	<b>Range / Frequency</b>	<b>*Calibration Measurement Capability (±)</b>	<b>Remarks</b>
44. MICROMETER SETTING STANDARD / LENGTH BAR (Inspection & Work Shop Grade) / LONG GAUGE BLOCK <sup>§</sup>	0 to 150 mm	0.8 µm	Using Slip Gauge Set & Elec. Comparator
	> 150 mm to 300 mm	1.42 µm	
	> 300 mm to 400 mm	1.5 µm	
	> 400 mm to 700 mm	3.0 µm	
	> 700 mm to 1000 mm	3.5 µm	
45. FEELER GAUGE / SHIMS (FOILS) OF COATING THICKNESS GAUGE / THICKNESS STANDARD <sup>§</sup>	Upto 2 mm	1.5 µm	Using Electronic Comparator with stand
46. MEASURING PINS / THREAD MEASURING WIRES / THREE WIRE UNIT <sup>§</sup>	0.17 mm to 20 mm	0.8 µm	Using Electronic Comparator with stand
47. THREAD MEASURING PRISMS <sup>§</sup>	A/B/C/D	1.0 µm	Using Electronic Comparator
48. CYLINDRICAL SETTING MASTER <sup>§</sup>	Upto 100 mm	1.6 µm	Using Slip Gauge Set & Electronic Comparator
	> 100 mm to 200 mm	2.2 µm	
49. PLAIN RING GAUGES / SETTING RING GAUGE <sup>§</sup>	Upto 100 mm	2.3 µm	Using Universal Length Measuring Machine
	> 100 mm to 200 mm	2.5 µm	
	>200 mm to 300mm	5.0 µm	Using Linear Height Gauge
	> 300 mm to 400mm	6.7 µm	



<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>9 of 16</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
50. CALIPER CHECKER / STEP GAUGE / INTERNAL MICRO CHECKER <sup>\$</sup>	0 to 370 mm 0 to 1000 mm	4.5 $\mu$ m 6.1 $\mu$ m	Using Linear Height Gauge & Length Bars
50. DEPTH MICRO CHECKER <sup>\$</sup>	0 to 300 mm	3.5 $\mu$ m	Using Linear Height & Length Bars
51. RISER BLOCK	Upto 300 mm	3.0 $\mu$ m	Using Electronic Comparator & Slip Gauge
52. RADIUS GAUGE SET / RADIUS <sup>\$</sup>	Upto 200 mm	5.6 $\mu$ m	Using Vision Measuring Machine
53. STANDARD WIRE GAUGE <sup>\$</sup>	Upto 100 mm	7.2 $\mu$ m	Using Vision Measuring Machine
54. THREAD PITCH GAUGE / THREAD PITCH <sup>\$</sup>	55°, 60° Flank angle 0.3 mm to 8.0 mm Pitch	9 as 5.6 $\mu$ m	Using Vision Measuring Machine
FLANK ANGLE / HELIX ANGLE OF THREADS <sup>\$</sup>	Upto 90°	9 as	
55. FORM GAUGES / TEMPLATES <sup>\$</sup>	Linear upto 200 mm Angle upto 360°	4.1 $\mu$ m 45 as	Using Vision Measuring Machine Direct Comparison Method
56. PLAIN PLUG GAUGE / SETTING PLUG GAUGE / WIDTH GAUGE <sup>\$</sup>	0 to 100 mm > 100 mm to 200 mm > 200 mm to 300 mm	1.6 $\mu$ m 2.2 $\mu$ m 3.0 $\mu$ m	Using Slip Gauge Set & Electronic Comparator
57. PLAIN SNAP GAUGE <sup>\$</sup>	Upto 300 mm	3.0 $\mu$ m	Using Slip Gauge Set

**Avijit Das**  
Program Manager

**Bibin Philip**  
Convenor

<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>10 of 16</b>

<b>Quantity Measured/ Instrument</b>	<b>Range / Frequency</b>	<b>*Calibration Measurement Capability (<math>\pm</math>)</b>	<b>Remarks</b>
<b>58. THREAD PLUG GAUGE<sup>\$</sup> (Effective Diameter Only)</b>	Upto 100 mm	3.6 $\mu$ m	Using FCDM, Cylindrical Setting Masters & Thread Measuring Wires Using Universal Length Measuring Machine & thread measuring wires
	> 100 mm to 200 mm	2.5 $\mu$ m	
	> 200 mm to 300 mm	3.0 $\mu$ m	
<b>59. TAPER THREAD PLUG GAUGE<sup>\$</sup> (Effective Diameter Only)</b>	Upto 100 mm	4.0 $\mu$ m	Using FCDM, Cylindrical Setting Masters & Thread Measuring Wires
	> 100 mm to 200 mm	2.5 $\mu$ m	Using Universal Length Measuring Machine & thread measuring wires
<b>60. THREAD RING GAUGE<sup>\$</sup> (Effective Diameter Only)</b>	3 mm to 100 mm	2.2 $\mu$ m	Using Universal Length Measuring Machine
	>100 mm to 200 mm	2.8 $\mu$ m	
<b>61. TAPER THREAD RING GAUGE<sup>\$</sup> (Effective Diameter Only)</b>	Upto 110 mm	2.3 $\mu$ m	Using Universal Length Measuring Machine
<b>62. PLAIN TAPER PLUG GAUGE<sup>\$</sup></b>	Diameter upto 150 mm	2.6 $\mu$ m	Using Universal Length Measuring Machine
	Half include Angle Upto 60°	9 as	
<b>63. PLAIN TAPER RING GAUGE<sup>\$</sup></b>	Diameter upto 150 mm	2.6 $\mu$ m	Using Universal Length Measuring Machine
	Half include Angle Upto 60°	40.8 as	

<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>11 of 16</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
64. SPLINE GAUGE <sup>\$</sup> (INTERNAL)	Upto 200mm	2.1 $\mu$ m	Using Slip Gauges & Measuring Pins
65. SPLINE GAUGE <sup>\$</sup> (EXTERNAL)	Upto 100mm	3.8 $\mu$ m	Using F.C.D.M. & Measuring Pins
	> 100mm to 200mm	3.8 $\mu$ m	Using Universal Length Measuring Machine & Measuring Pins
66. ANGLE MEASUREMENT <sup>\$</sup>	0° to 360°	6.4 as	Using Vision Measuring Machine
67. BEVEL PROTRACTOR <sup>\$</sup> L.C:1' <sup>Φ</sup>	0° -90°-0°-90°	4 a min	Using Angle Gauge Blocks
68. DEGREE PROTRACTOR <sup>\$</sup> L.C. 1°	0° to 360°	43 a min	Using Angle Gauge Blocks
69. COMBINATION SET <sup>\$</sup> L.C. 1°	0° -90°-0°	43 a min	Using Angle Gauge Blocks
70. SPIRIT LEVEL / FRAME LEVEL <sup>\$</sup> L.C.0.01mm	Base Length Upto 300 mm	2.9 $\mu$ m/m	Using Electronic Comparator
ELECTRONIC LEVEL L.C.: 0.001 mm/m <sup>Φ</sup>			

<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>12 of 16</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
71. SINE BAR / SINE CENTER <sup>§</sup>	100 mm/200mm/300 mm		Using Vertical measuring machine, Slip Gauge Blocks Grade-0, Angle Gauges & electronic comparator
Linear		4.2 $\mu$ m	
Angular		11.2 as	
72. ANGLE GAUGES <sup>§</sup>	Upto 90°	18.3 as	Using Slip Gauge Blocks Grade-0, Sine Bar, Electronic Comparator
73. ANGULAR GRATICULE <sup>§</sup> (Scale)	0° to 360°	8.5 as	Using Vision Measuring Machine
74. VICKERS INDENTOR, KNOOP INDENTOR, ROCKWELL DIAMOND CONE INDENTOR <sup>§</sup>	Angle 136°, 130°, 172°-30', 120°  Radius 0.2 mm	76 as  4.5 $\mu$ m	Using Vision Measuring Machine
75. SURFACE ROUGHNESS SPECIMEN <sup>§</sup>	Ra Rz Rmax	6.4% 6.5% 6.5%	Using Surface Roughness Tester
76. OPTICAL FLAT / OPTICAL PARALLEL <sup>§</sup>			
- FLATNESS	Upto 100 mm	0.25 $\mu$ m	Using Optical Flat
- PARALLELISM	Upto 100 mm	0.9 $\mu$ m	Using Electronic Comparator
77. TEST SIEVES <sup>§</sup>	32 $\mu$ m to 50mm > 50 mm to 125mm	5.6 $\mu$ m 37.5 $\mu$ m	Using Vision Measuring Machine & Digital Caliper

<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>13 of 16</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>78. SHORE-A Hardness Tester INDENTOR EXTENSION &amp; SHAPE<sup>§</sup> ( 0 to 100 shore A)</b>			
<b>i) Shaft Diameter of indenter d2</b>	1.25 mm	5.6 $\mu$ m	Using Vision Measuring Machine for d2, $\alpha$ & d1
<b>ii) Cone angle of indenter <math>\alpha</math></b>	35°	5 as	
<b>iii) Hole diameter of pressure foot d1</b>	3 mm	5.6 $\mu$ m	
<b>iv) Diameter of pressure foot D</b>	18 mm	27.1 $\mu$ m	Using Digital Caliper
<b>v) Depth of indenter-L</b>	Upto 2.5 mm	2.6 $\mu$ m	Using Digital Dial Gauge
<b>79. SHORE-D Hardness Tester INDENTOR EXTENSION &amp; SHAPE<sup>§</sup> (0 to 100 Shore D)</b>			
<b>i) Shaft Diameter of indenter d2</b>	1.25 mm	5.6 $\mu$ m	Using Vision Measuring Machine for d2, $\alpha$ & d1
<b>ii) Cone angle of indenter <math>\alpha</math></b>	30°	5 as	
<b>iii) Hole diameter of pressure foot d1</b>	3 mm	5.6 $\mu$ m	
<b>iv) Diameter of pressure foot D</b>	18 mm	27.4 $\mu$ m	Using Digital Caliper
<b>v) Depth of indenter-L</b>	Upto 2.5 mm	2.6 $\mu$ m	Using Digital Dial Gauge

<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>14 of 16</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
80. SURFACE PLATE* - GRANITE - CAST IRON	Upto 4000 mm X 4000 mm	$1.0\sqrt{\frac{L+W}{120}}$ where L is in mm, W is in mm	Using Electronic Level
81. PROFILE PROJECTOR / OPTICAL MICROSCOPE / TOOL MAKERS MICROSCOPE / BRINELL MICROSCOPE / VISION MEASURING MACHINE*			
a) Linear X,Y Axis L.C. 0.0001 mm <sup>Φ</sup>	0 to 200 mm	2.7 μm	Using Linear Glass Scale
b) Angular	0° to 360° 0° to 360°	5 as 8.9 as	Using Angle Gauges Using Angular Graticule
c) Magnification	upto 1000X	0.45 %	Using Linear Glass scale & Digital Caliper
82. LENGTH MEASURING MACHINE* (Horizontal) L.C 0.0001 mm <sup>Φ</sup>	Upto 300 mm	1.4 μm	Using Slip Gauge Blocks
83. VERTICAL SINGLE AXIS MEASURING MACHINE / ELECTRONIC HEIGHT GAUGE* L.C. 0.0001mm <sup>Φ</sup>	Upto 1000 mm	2.5+(L/1000) Where L in mm	Using Length Bars

<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>15 of 16</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
84. STRAIGHT EDGE*	Upto 5000 mm	$2.9\sqrt{\frac{L}{120}}$ Where L is in mm	Using Electronic Level
<b>II. HARDNESS</b>			
1. CALIBRATION OF SHORE 'A' AND SHORE 'D' HARDNESS TESTER <sup>s</sup>	0 to 100 Shore Unit 0 to 100 Shore Unit	0.9 Shore 'A' 0.7 Shore 'D'	Using Shore Hardness Tester Calibrator as per ASTM D 2240
2. VERIFICATION OF FORCE OF SHORE 'A' AND SHORE 'D' HARDNESS TESTER AND SHORE 'A', SHORE 'D' HARDNESS TESTER CALIBRATOR <sup>s</sup>	0-100 Shore 'A' 0-100 Shore 'D'	0.15 Shore Unit	Using Load Cell with Indicator as per ASTM D 2240
3. VERIFICATION OF ROCKWELL HARDNESS TESTER BY INDIRECT METHOD*	HRA HRB HRC	1.9 HRA 1.0 HRB 1.9 HRC	Using Standard Hardness Test Blocks as per IS 1586 (Part 2); 2012
4. VERIFICATION OF BRINELL HARDNESS TESTER BY INDIRECT METHOD*	10/3000 HBW 5/750 HBW	2.0 % 2.0 %	Using Standard Hardness Test Blocks as per IS 2281:2005
<b>III. ACOUSTIC</b>			
1. SOUND LEVEL METER <sup>s</sup>	94 dB & 114 dB	0.51 dB	Using Sound Level Meter

<b>Laboratory</b>	<b>Accurate Calibration &amp; Testing Center, 168, B-Wing, 1<sup>st</sup> Floor, Jai Ganesh Vision, Akurdi, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Mechanical Calibration</b>	<b>Issue Date</b>	<b>25.09.2014</b>
<b>Certificate Number</b>	<b>C-0503</b>	<b>Valid Until</b>	<b>24.09.2016</b>
<b>Last Amended on</b>	<b>29.09.2014</b>	<b>Page</b>	<b>16 of 16</b>

<b>Quantity Measured/ Instrument</b>	<b>Range / Frequency</b>	<b>*Calibration Measurement Capability (±)</b>	<b>Remarks</b>
--	--------------------------	--	----------------

#### IV. ACCELERATION AND SPEED

<b>TACHOMETER / RPM METER / SPEED INDICATORS <sup>§</sup> (Non-Contact)</b>	10 rpm to 100 rpm > 100 rpm to 1000 rpm > 1000 rpm to 10000 rpm > 10000 rpm to 99900 rpm	1.3 % 0.1 % 0.017 % 0.020 %	Using RPM Tester
---	---	--------------------------------------	------------------

#### V. TORQUE

<b>1. TORQUE WRENCHES<sup>§</sup> Type – I Class A,B,C,D,E Type- II Class A,B,C,D,E,F,G</b>	0 to 1000 Nm	1.20 % 1.29 %	Using Digital Torque Calibration System with four torque transducers As per ISO 6789:2003
---	--------------	------------------	--

\* Measurement Capability is expressed as an uncertainty (±) at a confidence probability of 95%

<sup>§</sup> Only in Permanent Laboratory

\* Only for Site Calibration

<sup>Φ</sup> Laboratory can also calibrate instruments/devices of coarser resolution / least count within the accredited range using same reference standard/ master equipment under the scope of accreditation.