

Laboratory	Calibration & Standards Laboratory (Electro Thermal), Controllerate Of Quality Assurance (Electronics), JC Nagar P.O., Bangalore, Karnataka		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	07.11.2016
Certificate Number	C-0838	Valid Until	06.11.2018
Last Amended on	-	Page	1 of 4

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>SOURCE</u>			
1. DC VOLTAGE^s	1 mV to 100 mV	0.047% to 0.0014%	Using Multi Function Calibrator Fluke 5720A by Direct Method
	100 mV to 1 V	0.0014% to 0.0007%	
	1 V to 10 V	0.0007% to 0.0005%	
	10 V to 1000 V	0.0005% to 0.0009%	
2. DC CURRENT^s	10 μ A to 100 μ A	0.077% to 0.006%	Using Multi Function Calibrator Fluke 5720A by Direct Method
	100 μ A to 1 mA	0.006% to 0.005%	
	1 mA to 10 mA	0.005% to 0.0045%	
	10 mA to 100 mA	0.0045% to 0.006%	
	100 mA to 2 A	0.006% to 0.011%	
	2 A to 100 A	0.01% to 0.047%	Using Multi Function Calibrator Fluke 5720A with Transconductance Amplifier by Direct Method
3. AC VOLTAGE^s	20 Hz to 40 Hz		Using Multi Function Calibrator Fluke 5720A by Direct Method
	20 mV to 20 V	0.097% to 0.031%	
	40 Hz to 1 kHz		
	2 mV to 20 mV	0.22% to 0.035%	
	20 mV to 200 mV	0.037% to 0.017%	
	200 mV to 2 V	0.017% to 0.013%	
	2 V to 200 V	0.013%	
	200 V to 750 V	0.013% to 0.037%	
	1 kHz to 50 kHz		
	2 mV to 20 mV	0.28% to 0.1%	
	50 kHz to 300 kHz		
	200 mV to 20 V	0.2 % to 0.042%	
300 kHz to 1 MHz			
2 V to 20 V	0.26 %		

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory	Calibration & Standards Laboratory (Electro Thermal), Controllerate Of Quality Assurance (Electronics), JC Nagar P.O., Bangalore, Karnataka		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	07.11.2016
Certificate Number	C-0838	Valid Until	06.11.2018
Last Amended on	-	Page	2 of 4

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
4. AC CURRENT^{\$}	40 Hz to 1 kHz		
	10 μ A to 200 μ A	0.11% to 0.02%	Using Multi Function Calibrator Fluke 5720A by Direct Method
	200 μ A to 100 mA	0.02% to 0.027%	
	100 mA to 1 A	0.027% to 0.04%	
	1 kHz		
	1 A to 20 A	0.04% to 0.13%	Using Multi Function Calibrator Fluke 5720A with Transconductance Amplifier by Indirect Method
5. RESISTANCE^{\$} (in steps of 1-1.9-10)	1 Ω to 1.9 Ω	0.011%	Using Multi Function Calibrator Fluke 5720A by Direct Method
	1.9 Ω to 10 Ω	0.011 %to 0.0027%	
	10 Ω to 19 k Ω	0.0027% to 0.0011%	
	19 k Ω to 1 M Ω	0.0011% to 0.0046%	
6. OSCILLOSCOPE^{\$} VERTICAL CHANNEL			
a) DC Voltage	3 mV to 50 V (1M Ω)	2.32%	Using Multi Function Calibrator Fluke 5720A by Direct Method
	3 mV to 5 V (50 Ω)	4.43%	
b) Square wave	1 kHz		
	6 mV to 60 V (1M Ω)	0.9%	
	1 kHz		
	30 mV to 5V (50 Ω)	1.5%	
TIME BASE	10 ns to 1 s	115 ppm to 356 ppm	
BAND WIDTH	10 MHz to 1GHz	3.9% to 4.9%	

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory	Calibration & Standards Laboratory (Electro Thermal), Controllerate Of Quality Assurance (Electronics), JC Nagar P.O., Bangalore, Karnataka		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	07.11.2016
Certificate Number	C-0838	Valid Until	06.11.2018
Last Amended on	-	Page	3 of 4

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>MEASURE</u>			
1. DC VOLTAGE^{\$}	190 mV to 1.9 V	0.0039% to 0.0015%	Using DMM, Keithley 2002 by Direct Method
	1.9 V to 19 V	0.0015%	
	19 V to 1000 V	0.0015% to 0.0028%	
2. DC CURRENT^{\$}	190 μ A to 1.9 mA	0.045% to 0.043%	Using DMM, Keithley 2002 by Direct Method
	1.9 mA to 190 mA	0.043% to 0.046%	
	190 mA to 1.9 A	0.046% to 0.09%	
3. DC CURRENT[#]	2A to 100A	0.016% to 0.0054%	Using DMM Keithley 2002, with Standard Resistor L&N by V/R Method
4. AC VOLTAGE^{\$}	50 Hz to 1 kHz		Using DMM, Keithley 2002 by Direct Method
	190 mV to 19 V	0.1%	
	19 V to 190 V	0.1%	
	190 V to 750 V	0.1 to 0.13%	
	1 kHz to 50 kHz		
	190 mV to 19 V	0.08% to 0.1%	
5. AC CURRENT^{\$}	5 kHz to 50 kHz		Using DMM, Keithley 2002 by Direct Method
	19 V to 100 V	0.1% to 0.12%	
	40 Hz to 100 Hz		
	190 μ A to 1.9 mA	0.42% to 0.37%	
	1.9 mA to 190 mA	0.37 %	
	190 mA to 1.9 A	0.37 % to 0.43%	
	100 Hz to 1kHz		
	1.9 mA to 190 mA	0.19%	

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory	Calibration & Standards Laboratory (Electro Thermal), Controllerate Of Quality Assurance (Electronics), JC Nagar P.O., Bangalore, Karnataka		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	07.11.2016
Certificate Number	C-0838	Valid Until	06.11.2018
Last Amended on	-	Page	4 of 4

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
	50 Hz		
	2 A to 10 A	1.6% to 1.1%	
	10 A to 20 A	0.13%	
6. RESISTANCE^{\$}	1 Ω to 10 Ω	0.041% to 0.0066%	Using DMM, Keithley 2002 by Direct Method
	10 Ω to 100 Ω	0.0066% to 0.0051%	
	100 Ω to 1k Ω	0.0051% to 0.0019%	
	1 k Ω to 10 k Ω	0.002%	
	10 k Ω to 100 k Ω	0.002% to 0.005%	
	100 k Ω to 1M Ω	0.005% to 0.01%	
7. OSCILLOSCOPE^{\$}			Using Oscilloscope Calibrator Fluke 9500 by Direct Method
VERTICAL CHANNEL			
DC VOLTAGE	5mV to 190mV (1M Ω)	3.2 % to 2.03%	
TIME BASE / PULSE	2 ns to 10 ms	143 ppm to 446 ppm	
WIDTH			

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%.

^{\$}Only in Permanent Laboratory

*Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager