

Laboratory Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration Issue Date 27.01.2014

Certificate Number C-1011 Valid Until 26.01.2016

Last Amended on - Page 1 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>MEASURE</u></b>			
<b>1. DC VOLTAGE<sup>§</sup></b>	10 $\mu$ V to 1 mV	0.80% to 60 ppm	Using DC Ref Std , Ref Divider, Kelvin Varley (KV) Divider and Null Detector By Null Method
	1 mV to 10 mV	60ppm to 10 ppm	
	10 mV to 100 mV	10ppm to 4 ppm	
	100 mV to 1000 V	4 ppm	
<b>2. AC VOLTAGE<sup>§</sup></b>	<b>10 Hz to 40 Hz</b>		Using Micropot,DC Ref Std, KV Divider & Null meter
	1 mV to 135 mV	0.1% to 510 ppm	
	350 mV to 1 V	510 ppm to 61 ppm	
	<b>40 Hz to 20 kHz</b>		Using Micropot / Thermal Voltage Convertor (TVC) ,DC Ref Std, KV Divider & Nullmeter By Substitution Method
	1 mV to 135 mV	0.23% to 700 ppm	
	350 mV to 0.5 V	700 ppm to 20 ppm	
	<b>20 kHz to 30 kHz</b>		700 ppm to 20 ppm
	1 mV to 135 mV	0.23% to 700 ppm	
	350 mV to 0.5 V	700 ppm to 20 ppm	
	<b>30 kHz to 100 kHz</b>		20 ppm to 180 ppm
	1 mV to 135 mV	0.23% to 700 ppm	
	350 mV to 0.5 V	700 ppm to 20 ppm	
<b>100 kHz to 1 MHz</b>		20 ppm to 250 ppm	
1 mV to 135 mV	0.23% to 700 ppm		
350 mV to 0.5 V	700 ppm to 20 ppm		
	0.5 V to 20 V	20 ppm to 110 ppm	

**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014

**Certificate Number** C-1011 **Valid Until** 26.01.2016

**Last Amended on** - **Page** 2 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
3. DC CURRENT <sup>\$</sup>	10 $\mu$ A to 1 A 1 A to 20 A	30 ppm 30 ppm to 35 ppm	Using Std Resistor, DC Ref Std , Reference Divider, Kelvin Varley Divider and Null Detector By Direct Method
4. AC CURRENT <sup>\$</sup>	20A to 100 A	35 ppm to 500 ppm	Using Current Shunt, 8.5 DMM By Direct Method
	<b>1kHz</b> 10 $\mu$ A to 5 mA	120 ppm	Using Standard Resistor, MFC, 8.5 digit DMM By Substitution Method
	<b>40 Hz to 5 kHz</b> 5 mA to 1 A 1 A to 20 A	130 ppm to 160 ppm 160 ppm to 300 ppm	Using Primary AC Shunt Set, MFC, Ref Std, Null Detector, KV Divider By Substitution Method
	<b>5 kHz to 10 kHz</b> 5 mA to 1 A 1 A to 10 A	160 ppm 300 ppm to 320 ppm	Using Primary AC Shunt Set, MFC, Ref Std, Null Detector, KV Divider By Substitution Method
	<b>50Hz</b> 10 A to 100A	0.40%	Using Current Shunt, 8.5 DMM By Direct Method

**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014

**Certificate Number** C-1011 **Valid Until** 26.01.2016

**Last Amended on** - **Page** 3 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>5. DC RESISTANCE<sup>\$</sup></b>	100 $\mu\Omega$ to less than 1 m $\Omega$	550 ppm to 400 ppm	Using MFC , 8.5 DMM By Direct Method
	1 m $\Omega$ to 1 M $\Omega$	55 ppm to 15 ppm	Using Std Resistors, Kelvin Ratio Bridge By Comparison Method
	1 M $\Omega$ to 10 M $\Omega$ 10 M $\Omega$ to 100 M $\Omega$	15 ppm to 100 ppm 100 ppm to 700 ppm	Using Std Resistors, 8.5 DMM, Resistance Transfer Std By Comparison Method
	100 M $\Omega$ to 1 G $\Omega$ 1 G $\Omega$ to 100 G $\Omega$	700 ppm to 1.0% 1.0 % to 1.25%	Using 8.5 DMM, TeraOhm Meter, Discrete resistors By Comparison Method
<b>6. AC RESISTANCE<sup>\$</sup></b>	<b>1kHz</b> 1 $\Omega$ to 100 $\Omega$	0.10%	Using LCR Meter By Direct Method
	<b>1 kHz to 100 kHz</b> 100 $\Omega$ to 1 k $\Omega$	0.20 % to 1.0%	
	<b>1 kHz</b> 1k $\Omega$ , 10 k $\Omega$	120 ppm	Using LCR Meter & Std Resistor By Direct Method
<b>7. CAPACITANCE<sup>\$</sup></b>	<b>1kHz</b> 100pF	80 ppm	Using Std Capacitors and LCR Meter By Comparison Method
	1nF	275 ppm	
	10nF	280 ppm	
	100nF	280 ppm	
	1 $\mu$ F	280 ppm	

**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014

**Certificate Number** C-1011 **Valid Until** 26.01.2016

**Last Amended on** - **Page** 4 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks	
	<b>1 kHz to 100 kHz</b> 10pF to 10uF	0.20 % to 0.50 %	Using LCR Meter By Direct Method	
<b>8. INDUCTANCE<sup>\$</sup></b>	<b>1kHz</b> 10 $\mu$ H to 1mH 1mH to 10H	0.50% to 0.08% 0.08% to 0.8%	Using LCR Meter By Direct Method	
	100 $\mu$ H 1 mH 10 mH 100 mH 1 H 2 H	0.25% 400 ppm 350 ppm 275 ppm 410 ppm 400ppm	Using LCR Meter By Comparison Method	
<b>9. DC POWER<sup>\$</sup></b>	<b>1V to 500V, 0.1A to 20A</b> <b>500V to 1kV, 20A to 30A</b> <b>500 V to 1kV, 30A to 1kA</b>	0.1 W to 10 kW	0.02% to 0.12%	Using 8.5 Digit DMM By Direct Method
	10 kW to 30 kW 30 kW to 1 MW	0.12% 0.12% to 2.00%	Using 8.5 Digit DMM & Current Shunt/ Clamp Meter By Direct Method	
<b>10. AC POWER<sup>\$</sup></b> <b>Single Phase</b>	<b>50Hz, 0.25 PF to UPF</b> 0.1A/40V to 20A/300 V 4 W to 6 kW	0.07% to 0.25%	Using Watt Convertor , 8.5 DMM By Direct Method	
<b>11. PHASE ANGLE<sup>\$</sup></b>	+/- 180 deg	0.06 deg	Using Phase Meter & Phase Standard By Comparison Method	

**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014

**Certificate Number** C-1011 **Valid Until** 26.01.2016

**Last Amended on** - **Page** 5 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>12. POWER FACTOR<sup>\$</sup></b>	0.2 to 1 PF	0.012 PF	Using Power Meter By Direct Method
<b>13. FREQUENCY<sup>\$</sup></b>	0.1 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 26.5 GHz 26.5 GHz to 40 GHz	0.001 Hz 0.001 Hz to 0.1 Hz 0.1 Hz to 26.7 Hz 26.7 Hz to 40.15 Hz	Using Agilent 53131A, Agilent 5352B, Agilent 8510 C By Direct / Comparison Method
<b>14. RF POWER<sup>#</sup></b>	<b>10MHz to 50 MHz</b> 1 $\mu$ W to 20mW (-30dbm to 13dbm)  <b>50 MHz to 18 GHz</b> 100pW to 1 $\mu$ W (-70dbm to -30dbm)  1 $\mu$ W to 20mW (-30dbm to 13 dbm)  <b>18 GHz to 26.5 GHz</b> 10pW to 100mW (-70dbm to 13dbm)	0.21dbm to 0.26dbm  0.47dbm to 0.25dbm  0.26dbm to 0.29dbm  0.48dbm	Using Agilent 8481A, Agilent 8485D, Agilent 8481D, Agilent 8485D and 8485A By Direct / Comparison Method
<b>15. MODULATION<sup>\$</sup></b> <b>Amplitude Modulation</b> <b>Carrier Frequency</b> <b>10MHz to 1.3GHz</b> <b>50Hz to 100kHz</b>	Modulation Depth 5% to 99 %	3.8% to 5%	Using HP 8901A By Direct Method

**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014

**Certificate Number** C-1011 **Valid Until** 26.01.2016

**Last Amended on** - **Page** 6 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>Frequency Modulation Carrier Frequency 10MHz to 1.3GHz 50Hz to 100kHz</b>	Freq.Deviation 10kHz to 400kHz	1.8% to 6.8%	Using HP 8901A By Direct Method
<b>Phase Modulation Carrier Frequency 10MHz to 1.3GHz 200Hz to 10kHz</b>	Phase Deviation 10 rad to 80 rad	5.8%	Using HP 8901A By Direct Method
<b>16. TEMPERATURE SIMULATION<sup>#</sup> (Indicator/ Recorder/ Controller)</b>			
<b>RTD</b>	-200 °C to 800°C	0.007°C to 0.02°C	Using HP 3458A By Direct Method
<b>Thermocouple K Type</b>	3 °C to 1340 °C	0.009°C to 0.03°C	
<b>J Type</b>	2 °C to 750 °C	0.007°C to 0.03°C	
<b>R, S Type</b>	18 °C to 1700 °C	0.068°C to 0.12°C	
<b>T Type</b>	3 °C to 400 °C	0.04°C to 0.016°C	
<b>N Type</b>	5 °C to 1300 °C	0.02°C to 0.03°C	
<b>E Type</b>	2 °C to 800 °C	0.03°C to 0.002°C	
<b>B Type</b>	156 °C to 1800 °C	0.6°C to 0.5°C	

**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014

**Certificate Number** C-1011 **Valid Until** 26.01.2016

**Last Amended on** - **Page** 7 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>17. DC VOLTAGE*</b>	1 mV to 100 mV	0.04% to 0.001%	Using DMM 3458A By Direct Method
	0.1V to 10 V	0.001%	
	10V to 1000 V	0.001%	
	1kV to 40 kV	3.6%	Using 80k40 & 179 By Direct Method
<b>18. AC VOLTAGE*</b>	<b>20 Hz to 1 kHz</b>		Using DMM 3458A By Direct Method
	1 mV to 100 mV	0.4% to 0.002%	
	0.1 V to 10 V	0.02%	
	10 V to 700 V	0.02% to 0.5%	
	<b>1 kHz to 100 kHz</b>		
	0.1 V to 10 V	0.1%	
	10 V to 100 V	0.1% to 0.15%	
	<b>100 kHz to 1MHz</b>		Using 80k40 & 179 By Direct Method
	0.1 V to 10 V	1.2%	
	<b>50Hz</b>	6.5%	
1kV to 28 kV			
<b>19. DC CURRENT*</b>	10 $\mu$ A to 100 mA	2.5% to 0.005%	Using DMM 3458A By Direct Method
	100 mA to 1 A	0.005% to 0.02%	
	1A to 100 A	0.1%	Using Current Shunt, 8.5 DMM By Direct Method

**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014

**Certificate Number** C-1011 **Valid Until** 26.01.2016

**Last Amended on** - **Page** 8 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
20. AC CURRENT*	<b>10 Hz to 45 Hz</b> 10 $\mu$ A to 1 A	3.5% to 0.5%	Using DMM 3458A By Direct Method
	<b>45 Hz to 1 kHz</b> 10 $\mu$ A to 100 $\mu$ A 0.1 mA to 100 mA 0.1 A to 1 A	3.5% to 0.25% 0.25% 0.25%	
	<b>1 kHz to 10 kHz</b> 10 mA to 100 mA 0.1 A to 1 A	0.1% 0.1% to 0.4%	
21. DC RESISTANCE*	<b>50 Hz</b> 1A to 100A	0.5%	Using Current Shunt, DMM 3458A By Direct Method
22. FREQUENCY*	1 $\Omega$ to 1 G $\Omega$	0.002% to 0.6%	Using DMM 3458A By Direct Method
	1Hz to 15kHz 15kHz to 26.5 GHz	0.001 Hz to 1.74 Hz 1.74 Hz to 27 Hz	Using DMM 3458A Counter 5348A By Direct / Comparison Method
23. TIME INTERVAL*	1s to 10800 s	0.012 s to 1.4 s	Using TIMER CT6S-2P By Direct Method



**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014

**Certificate Number** C-1011 **Valid Until** 26.01.2016

**Last Amended on** - **Page** 9 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
----------------------------------	-------------------	--	---------

**SOURCE**

<b>24. DC VOLTAGE<sup>§</sup></b>	10 $\mu$ V to 1 mV	0.58% to 58 ppm	Using DC Ref Std , Ref Divider, Kelvin Varley (KV) Divider and Null Detector By Direct Method
	1 mV to 10 mV	58 ppm to 8 ppm	
	10 mV to 100 mV	8 ppm to 4 ppm	
	100 mV to 1000 V	4 ppm	
<b>25. AC VOLTAGE<sup>§</sup></b>	<b>10 Hz to 40 Hz</b>	0.1% to 510 ppm 510 ppm to 65 ppm	Using MFC, Micropot, KV Divider, DC Ref Std, & Null Detector By Substitution Method
	1 mV to 135 mV		
	350 mV to 1 V		
	<b>40 Hz to 20 kHz</b>	0.23% to 700 ppm 700 ppm to 40 ppm 40 ppm to 200 ppm	Using MFC, Micropot / TVC ,DC Ref Std, KV Divider & Null Detector By Substitution Method
	1 mV to 135 mV		
	350 mV to 0.5 V		
	<b>20 kHz to 30 kHz</b>	0.23% to 700 ppm 700 ppm to 40 ppm 40 ppm to 200 ppm	Using MFC, Micropot / TVC ,DC Ref Std, KV Divider & Null Detector By Substitution Method
	1 mV to 135 mV		
	350 mV to 0.5 V		
	<b>30 kHz to 100 kHz</b>	0.23% to 700 ppm 700 ppm to 40 ppm 40 ppm to 260 ppm	Using MFC, Micropot / TVC ,DC Ref Std, KV Divider & Null Detector By Substitution Method
	1 mV to 135 mV		
	350 mV to 0.5 V		
<b>100 kHz to 1 MHz</b>	0.23% to 700 ppm 700 ppm to 40 ppm 40 ppm to 0.08 %	Using MFC, Micropot / TVC ,DC Ref Std, KV Divider & Null Detector By Substitution Method	
1 mV to 135 mV			
350 mV to 0.5 V			
	0.5 V to 20 V		

**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014

**Certificate Number** C-1011 **Valid Until** 26.01.2016

**Last Amended on** - **Page** 10 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>26. DC CURRENT</b> <sup>\$</sup>	10 $\mu$ A to 1 A 1 A to 20 A	250 ppm to 50 ppm 50 ppm to 210 ppm	Using MFC, Std Resistor By Direct Method
	20A to 100 A	210 ppm to 0.1%	Using Current Shunt, DC Power Supply, 8.5 DMM By Direct Method
	100A to 1000 A	0.40%	Using MFC, Current Coil By Direct Method
<b>27. AC CURRENT</b> <sup>\$</sup>	<b>1kHz</b> 10 $\mu$ A to 5 mA	135 ppm	Using Standard Resistor, MFC, 8.5 digit DMM By Substitution Method
	<b>40 Hz to 5 kHz</b> 5 mA to 1 A 1 A to 20 A	160 ppm to 220 ppm 220 ppm to 625 ppm	Using Primary AC Shunt Set, MFC, Ref Std, Null Detector, KV Divider By Substitution Method
	<b>5 kHz to 10 kHz</b> 5 mA to 1 A 1 A to 10 A	900 ppm to 200 ppm 200 ppm to 350 ppm	Using Primary AC Shunt Set, MFC, Ref Std, Null Detector, KV Divider By Substitution Method
	<b>50Hz</b> 20 A to 1000A	0.50%	Using MFC, Current Coil By Substitution Method

**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014

**Certificate Number** C-1011 **Valid Until** 26.01.2016

**Last Amended on** - **Page** 11 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>28. DC RESISTANCE<sup>\$</sup></b>	100 $\mu\Omega$ to 1 m $\Omega$	500 ppm to 60 ppm	Using DMM By Direct Method
	1 m $\Omega$ to 1 M $\Omega$	60 ppm to 15 ppm	Using Std Resistors, Kelvin Ratio Bridge By Direct Method
	1 M $\Omega$ to 10 M $\Omega$ 10 M $\Omega$ to 100 M $\Omega$	15 ppm to 20 ppm 20 ppm to 225 ppm	Using Std Resistors, 8.5 DMM, Resistance Transfer Std By Direct Method
	100 M $\Omega$ to 1 G $\Omega$ 1 G $\Omega$ to 100 G $\Omega$	225 ppm to 0.2% 0.20%	Using 8.5 DMM, TeraOhm Meter, Discrete resistors By Direct Method
<b>29. DISCRETE VALUES<sup>\$</sup></b>	0.0001 $\Omega$	75 ppm	Using Std Resistor By Direct Method
	0.001 $\Omega$	55 ppm	
	0.01 $\Omega$	40 ppm	
	0.1 $\Omega$	30 ppm	
	1 $\Omega$	25 ppm	
	10 $\Omega$	15 ppm	
	100 $\Omega$	15 ppm	
	1 k $\Omega$	15 ppm	
	10 k $\Omega$	15 ppm	
	100 k $\Omega$	15 ppm	
	1 M $\Omega$	15 ppm	
	10 M $\Omega$	20 ppm	
	1 G $\Omega$	0.20 %	
	10 G $\Omega$	0.30 %	
100 G $\Omega$	0.20 %		

**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014

**Certificate Number** C-1011 **Valid Until** 26.01.2016

**Last Amended on** - **Page** 12 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>30. AC RESISTANCE<sup>\$</sup></b>	<b>1 kHz</b> 1k $\Omega$ , 10 k $\Omega$	150 ppm	Using Std Resistor By Direct Method
	<b>1 kHz</b> 100 $\Omega$ to 1 k $\Omega$	0.1% to 150 ppm	Using LCR Meter & AC Resistance Box By Comparison Method
	<b>100 kHz</b> 100 $\Omega$ to 1 k $\Omega$	0.25% to 0.85%	Using LCR Meter & AC Resistance Box By Comparison Method
<b>31. CAPACITANCE<sup>\$</sup> DISCRETE VALUES</b>	<b>1kHz</b> 100pF	80 ppm	Using Std Capacitors By Direct Method
	1nF	275 ppm	
	10nF	280 ppm	
	100nF	280 ppm	
	1 $\mu$ F	280 ppm	
<b>1 kHz to 100 kHz</b> 10pF to 10uF	0.2% to 0.45%	Using LCR Meter & Capacitance Box By Comparison Method	
<b>32. INDUCTANCE<sup>\$</sup></b>	<b>1kHz</b> 100 $\mu$ H to 1mH	0.5 % to 0.08%	Using LCR Meter, Inductance Box By Comparison Method
	1mH to 10H	0.08% to 0.8%	
	100 $\mu$ H	0.25%	Using Std Inductors By Direct Method
	1 mH	400 ppm	
	10 mH	350 ppm	
	100 mH	275 ppm	
	1 H	410 ppm	
2 H	400 pp		

**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014

**Certificate Number** C-1011 **Valid Until** 26.01.2016

**Last Amended on** - **Page** 13 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>33. TEMPERATURE SIMULATION<sup>#</sup> (Controllers / Indicators/ Recorder)</b>			
<b>RTD</b>	-100°C to 800°C	0.02°C to 0.12°C	Using Transmile 3010 By Direct Method
<b>Thermocouple K Type</b>	-140°C to 1340°C	0.02°C to 0.16°C	
<b>J Type</b>	-180°C to 750°C	0.03°C to 0.1°C	
<b>R, S Type</b>	100°C to 1700°C	0.03°C to 0.2°C	
<b>T Type</b>	-200°C to 400°C	0.05°C	
<b>N Type</b>	-200°C to 1300°C	0.032°C	
<b>E Type</b>	0°C to 800°C	0.04°C to 0.09°C	
<b>B Type</b>	600°C to 1800°C	0.62°C to 0.5°C	
<b>34. DC POWER<sup>#</sup></b>	<b>1V to 1kV, 0.3A to 20A</b> 0.3 W to 20 kW	0.03% to 0.1%	Using MFC By Direct Method
	<b>1V to 1kV, 20A to 1kA</b> 20 kW to 1 MW	0.1% to 0.33%	
<b>35. AC POWER<sup>#</sup> Single Phase</b>	<b>50Hz, UPF</b> <b>0.1 A 10 V</b> <b>20A 1000 V</b> 1 W to 20 kW	0.15% to 0.1%	Using MFC By Direct Method

**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014

**Certificate Number** C-1011 **Valid Until** 26.01.2016

**Last Amended on** - **Page** 14 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
36. AC ENERGY <sup>#</sup> Single Phase	50Hz,UPF 10V to 1000V 0.1A to 20A 1Wh to 20kWh	0.2% to 0.15%	Using MFC & Universal Counter By Direct Method
37. PHASE ANGLE <sup>\$</sup>	50 Hz +/- 180 deg	0.05 deg	Using Phase Standard By Direct Method
38. POWER FACTOR <sup>#</sup>	0.2 to 1 PF	0.013 PF	Using MFC By Comparison Method
39. RF ATTENUATION <sup>\$</sup>	100MHz to 18 GHz 1 dB to 60dB	0.43 dB to 2.70dB	Using Power meter 437B, Power Sensor 8481A, 8485D, Attenuator 8494B,8495B By Direct/ Comparison / Indirect Method
40. VSWR <sup>#</sup>	10 MHz to 18 GHz 1.05 10 MHz to 26.5 GHz 1.1 to 2.00	0.07% to 0.1% 0.072% to 0.273%	Using Maury Microwave 8033A & Maury Microwave 2611A By Direct/ Comparison / Indirect Method

**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014

**Certificate Number** C-1011 **Valid Until** 26.01.2016

**Last Amended on** - **Page** 15 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>41. OSCILLOSCOPE<sup>#</sup></b>			
<b>Amplitude</b>	0.002 to 120 V 2ns to 5ns	0.30% 0.03% to 0.15%	Using MFC Time Marker Generator By Direct Method
<b>Time Marker</b>	10Hz to 100kHz	4.00%	
<b>Bandwidth</b>	100kHz to 1GHz	4.00%	
<b>42. DC VOLTAGE<sup>*</sup></b>	1 mV to 100 mV 0.1V to 1000 V	0.25% to 0.005% 0.005% to 0.003%	Using MFC 3010 By Direct method
<b>43. AC VOLTAGE<sup>*</sup></b>	<b>40 Hz to 10 kHz</b> 20 mV to 1000 V	0.3% to 0.1%	Using MFC 3010 By Direct method
	<b>10 kHz to 40 kHz</b> 20 mV to 100 V	0.4% to 0.1%	
	<b>40 kHz to 100 kHz</b> 20 mV to 20 V	1.2% to 0.1%	
	<b>100 kHz to 500 kHz</b> 20 mV to 2 V	1.2% to 0.5%	
<b>44. DC CURRENT<sup>*</sup></b>	100 $\mu$ A to 1 A 1 A to 20 A 20 A to 1000 A	0.03% to 0.02% 0.02% to 0.04% 0.04% to 0.5%	Using MFC 3010 By Direct method
<b>45. AC CURRENT<sup>*</sup></b>	<b>45Hz to 1 kHz</b> 25 $\mu$ A to 20 A	2.5% to 0.5%	Using MFC 3010 By Direct method
	<b>1 kHz to 10 kHz</b> 0.3 mA to 200 mA	0.8%	
	<b>50 Hz</b> 20 A to 1000A	0.5%	Using Current Coil

**Laboratory** Northlab India Pvt. Ltd.-Hosur (Plant-4), Plot No. 21, Phase-II, Hosur Valley Site, Viswanathapuram, Hosur, Tamil Nadu  
**Accreditation Standard** ISO/IEC 17025:2005  
**Discipline** Electro-Technical Calibration **Issue Date** 27.01.2014  
**Certificate Number** C-1011 **Valid Until** 26.01.2016  
**Last Amended on** - **Page** 16 of 16

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
46. DC RESISTANCE*	1 $\Omega$	0.5%	By Direct method
	10 $\Omega$	0.06%	Using MFC 3010
	100 $\Omega$	0.008%	By Direct method
	1 k $\Omega$	0.003%	
	10 k $\Omega$	0.002%	
	100 k $\Omega$	0.003%	
	1 M $\Omega$	0.004%	
	10 M $\Omega$	0.0012%	
	100 M $\Omega$	0.3%	
	1 G $\Omega$	1.5%	
47. CAPACITANCE*	1 nF to 10 $\mu$ F	0.3% to 0.9%	Using MFC 3010 By Direct method
48. INDUCTANCE*	100 mH to 10 H	3%	Using Inductance Box By Direct method

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%

\$Only in Permanent Laboratory

\*Only for Site Calibration

# The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.