

Laboratory	Perfect Calibration Centre Pvt. Ltd., No. 40, 71st Cross, 5th Block, Rajaji Nagar, Bangalore, Karnataka		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Optical Calibration	Issue Date	21.10.2014
Certificate Number	C-1149	Valid Until	20.10.2016
Last Amended on	-	Page	1 of 1

	Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
1.	LUX METER^{\$}	100 Lux to 25000 Lux	3.1 %	Using Calibrated Master Lux meter By Comparison Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

^{\$} Only in Permanent Laboratory

Neeraj Verma
Convenor

Avijit Das
Program Manager