

**Laboratory** HEICO Engineering Services Private Limited, B-59/4, Naraina  
 Industrial Area, Phase - II, New Delhi  
**Accreditation Standard** ISO/IEC 17025: 2005  
**Certificate Number** CC-2406 **Page** 1 of 1  
**Validity** 06.10.2017 to 05.10.2019 **Last Amended on** --

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>MECHANICAL CALIBRATION</u></b>				
<b>I.</b>	<b>UTM, TENSION CREEP AND TORSION TESTING MACHINE</b>			
1.	Force* UTM CTM Under Compression Mode (Class-1)	5 kN to 2000 kN	0.62%	Using Calibrated Proving Ring/ Load Cells as per IS 1828 (Part-I)

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%  
 \*Only for Site Calibration

---

**Battal Singh**  
 Convenor

---

**Avijit Das**  
 Program Director