

Laboratory **Ordnance Factory Katni Laboratory, Katni Premises, Katni, Madhya Pradesh**

Accreditation Standard **ISO/IEC 17025: 2005**

Discipline **Non Destructive Testing** **Issue Date** **18.10.2016**

Certificate Number **T-0084** **Valid Until** **23.05.2018**

Last Amended on **-** **Page** **1 of 1**

S.No.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
I.	METALS AND ALLOYS			
1.	Radiographic Test	Detection of Flaw with X-Ray	IS 2595: 1978 (RA 2008)	20 to 320 kV Tube Voltage, 1 to 5 mA Tube-Current, penetration 5 mm to 78 mm Fe or equivalent

-X-X-X-X-X-X-X-X-X-X-X-X-