

Laboratory Central Testing Laboratory, Eveready Industries India Ltd., P4,
Transport Depot Road, Taratola, Kolkata, West Bengal

Accreditation Standard ISO/IEC 17025: 2005

Discipline Electrical Testing **Issue Date** 26.03.2015

Certificate Number T-0340 **Valid Until** 25.03.2017

Last Amended on - **Page** 1 of 1

S.No.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
I.	BATTERIES			
1.	Multipurpose Dry Batteries R20 R14 R6 R03	Dimensions and terminals Verification of markings Life test -Initial -Delayed -Delayed life test under dry heat condition Leakage test	IS 8144 : 1997 IS 6303 : 1984	0.01 mm to 150 mm Qualitative 15 s to 24 hours 1 Ω to 10 kΩ 0.01 Ω to 0.1 Ω 0.001 V DC to 12 VDC Qualitative
2.	Heavy Duty Dry Batteries R20 R14 R6 R03	Dimension and terminals Verification of markings Life test -Initial -Delayed -Delayed life test under dry heat conditions Resistance to leakage of electrolyte	IS 9128:1999	0.01 mm to 150 mm Qualitative 15 s to 24 hours 1 Ω to 10 kΩ 0.01 Ω to 0.1 Ω 0.001 VDC to 12 VDC Qualitative
3.	Primary Batteries	Dimension and terminals Verification of markings Discharge performance Leakage Test	IEC 60086-1: 2007 IEC 60086-2: 2007	0.01 mm to 150 mm Qualitative 15 s to 24 hours 1 Ω to 10 kΩ Ω to 0.1 Ω 0.001 VDC to 12 VDC Qualitative

-X-X-X-X-X-X-X-X-X-X-X-X-

Mallika Gope
Convenor

N. Venkateswaran
Program Manager