

Laboratory	Haryana State Pollution Control Board, Laboratory (HQ), Plot No. C-11, Sector-6, Panchkula, Haryana		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Chemical Testing	Issue Date	18.09.2014
Certificate Number	T-1155	Valid Until	17.09.2016
Last Amended on	-	Page	1 of 2

S.No.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
I. WATER				
1.	Waste Water	pH	APHA, 4500-H+B, 4-92 (22 nd Edition) 2012	1 to 12.0
		Total Suspended Solid	APHA, 2540-D, 2-66 (22 nd Edition) 2012	5 mg/L to 5000 mg/L
		Total Dissolved Solid	APHA, 2540-C, 2-65 (22 nd Edition) 2012	10 mg/L to 10000 mg/L
		Total Solid	APHA, 2540-B, 2-64 (22 nd Edition) 2012	5 mg/L to 7000 mg/L
		Dissolved Oxygen	APHA, 4500-OC, 4-139 (22 nd Edition) 2012	1 mg/L to 14.0 mg/L
		BOD (3 Days)	IS 3025	5 mg/L to 50000 mg/L
		COD	APHA, 5220-B, 5-17 to 5-18 (22 nd Edition) 2012	5 mg/L to 150000 mg/L
		Oil & Grease	APHA, 5520-B, 5-40, (22 nd Edition) 2012	2 mg/L to 500 mg/L

Laboratory	Haryana State Pollution Control Board, Laboratory (HQ), Plot No. C-11, Sector-6, Panchkula, Haryana		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Chemical Testing	Issue Date	18.09.2014
Certificate Number	T-1155	Valid Until	17.09.2016
Last Amended on	-	Page	2 of 2

S.No.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
2.	Drinking Water	pH	APHA, 4500-H+B, 4-92 (22 nd Edition) 2012	1.0 to 12.0
		Total Suspended Solid	APHA, 2540-D, 2-66 (22 nd Edition) 2012	5 mg/L to 5000 mg/L
		Total Dissolved Solid	APHA, 2540-C, 2-65 (22 nd Edition) 2012	10 mg/L to 10000 mg/L
		Total Solid	APHA, 2540-B, 2-64 (22 nd Edition) 2012	5 mg/L to 7000 mg/L
		Dissolved Oxygen	APHA, 4500-OC, 4-139 (22 nd Edition) 2012	1 mg/L to 14 mg/L
		BOD (3 Days)	IS 3025	5 mg/L to 50000 mg/L
		COD	APHA, 5220-B, 5-17 To 5-18 (22 nd Edition) 2012	5 mg/L to 150000 mg/L
		Oil & Grease	APHA, 5520-B, 5-40 (22 nd Edition) 2012	2 mg/L to 500 mg/L

-X-X-X-X-X-X-X-X-X-X-X-X-

**Anuja Anand
Convenor**

**N. Venkateswaran
Program Manager**