

Laboratory **Quality Assurance Laboratory, Plant-II, Bharat Aluminium Company Limited, Korba, Chhattisgarh**
Accreditation Standard **ISO/IEC 17025: 2005**
Discipline **Electrical Testing** **Issue Date** **07.08.2015**
Certificate Number **T-1969** **Valid Until** **06.08.2017**
Last Amended on **-** **Page** **1 of 1**

S. No.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
I.	CABLES AND ACCESSORIES			
1.	Aluminium & Aluminium Alloy Wire /Rod	Resistance	IS 3635 : 1966 (RA 2005) Clause 3 Routine Method	0.250 Milli Ω to 0.500 Milli Ω
		Conductivity at 20 ° C	IS 5484 : 1997 (RA 2001) Clause 9	48 % IACS to 68 % IACS

~~-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-~~

Rini Narayan
Convenor

N. Venkateswaran
Program Manager