

Laboratory Crompton Greaves Consumer Electricals Ltd., Photometry Laboratory,
Lighting Division, Vill. - Kural, Ta-Padra, Dist: Vadodara, Gujarat

Accreditation Standard ISO/IEC 17025: 2005

Discipline Photometry Testing **Issue Date** 22.09.2016

Certificate Number T-2112 **Valid Until** 21.09.2018

Last Amended on - **Page** 1 of 1

S.No.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
I.	LAMPS, LUMINARIES & ACCESSORIES			
1.	Incandescent Lamp for General Lighting Purposes	Luminous Flux	IS 418: 2004	1 Lumen to 50,000 Lumen
2.	Fluorescent Tubular Lamp	a) Luminous Flux b) Colour Coordinates	IS 2418: 1977	a) 1 Lumen to 50,000 Lumen b) X: 0.001 to 0.999 Y: 0.001 to 0.999
3.	Compact Fluorescent Lamp	a) Luminous Flux b) Colour Coordinates	IS 15111: 2002	a) 1 Lumen to 50,000 Lumen b) X: 0.001 to 0.999 Y: 0.001 to 0.999
4.	Sodium Vapour Lamp	a) Luminous Flux b) Colour Coordinates	IS 9974 (Part 2): 1981	1 Lumen to 50,000 Lumen
5.	Self Ballasted LED Lamps	a) Luminous Flux b) Colour Coordinates	IS 16102: 2012	a) 1 Lumen to 50,000 Lumen b) X: 0.001 to 0.999 Y: 0.001 to 0.999

~~-X-X-X-X-X-X-X-X-X-X-X-X-~~

Neeraj Verma
Convenor

N. Venkateswaran
Program Manager