

Laboratory	Food Safety and Analytical Quality Control Laboratory, Christy Friedgram Industry, A2 & A3, SIDCO Industrial Estate, Andipalayam, Tiruchengode, District Namakkal, Tamil Nadu		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Biological Testing	Issue Date	30.06.2016
Certificate Number	T-2904	Valid Until	29.06.2018
Last Amended on	-	Page	1 of 1

S. No.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
I.	FOOD AND AGRICULTURAL PRODUCTS			
1.	Food for Infant Nutrition, Amylase Rich Energy Food -I & II, Energy Food, Complementary Weaning Food Containing Amylase Activity, Processed Cereal Based Complementary Food, Protein Rich food Supplements for Infant and Preschool Children Fruits and Vegetable Products, Cereals, Pulses and Its Products, Salt, Spices and Condiments	Total Bacterial Count at 30 °C for 72 hours	IS 5402: 2012 (2 nd Revision)	≥10 CFU/g
		Coliform Count at 30°C or 37 °C for 24 hours	IS 5401 (Part 1): 2012 (3 rd Revision)	≥ 10 CFU/g
		Yeast &Mould at 25 °C for 3,4 or 5 days	IS 5403: 1999 (RA 2009) (1 st Revision)	≥ 10 CFU/g
II.	WATER			
1.	Potable Water/ Drinking Water	Standard Plate Count at 37 °C for 24 h	IS 1622: 1981 (RA 2009) (1 st Revision)	≥ 1 CFU/ml

~~-X-X-X-X-X-X-X-X-X-X-X-X-~~