

Laboratory Regional Meter Testing Laboratory, MPMKVVL Co. Ltd, Nishta Parisar, Govindpura, Bhopal, Madhya Pradesh

Accreditation Standard ISO/IEC 17025: 2005

Discipline Electrical Testing **Issue Date** 24.05.2014

Certificate Number T-2964 **Valid Until** 23.05.2016

Last Amended on - **Page** 1 of 1

S.No.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
1.	ELECTRICAL INDICATING & RECORDING INSTRUMENTS			
1.	AC STATIC WATT HOUR METERS CLASS 1 & 2, AC STATIC WATT HOUR METERS AND VAR HOUR METERS CLASS 0.2 s, 0.5 & 1.0 s	Limits of Error Test of Starting Current Test of No Load Test of Repeatability of Error Test of meter Constant (Dial) Test	IS: 13779: 1999 RA 2002 IS: 14697: 1999 RA 2004 IEC 62052-11 IEC 62053-21 IEC 62053-22 CBIP 304: 2008	Voltage: 63.5 V to 240 V Current:5 mA to 120 A Frequency:50 Hz PF 0.5 lag, UPF, 0.866 lead

-X-X-X-X-X-X-X-X-X-X-X-X-