

Laboratory	Coal Testing Laboratory, MSPGCL, Koradi T.P.S.Koradi, District. Nagpur, Maharashtra		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Chemical Testing	Issue Date	18.09.2014
Certificate Number	T-3117	Valid Until	17.09.2016
Last Amended on	-	Page	1 of 1

S.No.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
--------------	-----------------------------------	--------------------------------	--	---

I. COAL, COKE & OTHER SOLID FUEL

1. Coal	1. Proximate Analysis	IS:1350 (Part-1)1984 RA 2007	
	a) Moisture		1 to 20%
	b) Ash		10 to 70%
	c) Volatile Matter		1 to 40%
	2. GCV	IS:1350 (Part-2)1970 RA 2005	1000 to 8500 cal/gm

-X-X-X-X-X-X-X-X-X-X-X-X-