

Laboratory Material Testing Laboratory of Senor Metals Pvt. Ltd., Plot -353,
GIDC-II, Jamnagar, Gujarat

Accreditation Standard ISO/IEC 17025: 2005

Discipline Chemical Testing **Issue Date** 08.12.2014

Certificate Number T-3232 **Valid Until** 07.12.2016

Last Amended on - **Page** 1 of 1

S.No.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
I.	METALS & ALLOYS			
1.	Copper & Copper Alloys	Sn Pb Zn Fe Ni Al Si Mn P Bi Mg Cr Cd As Sb Zr	BS EN15079-2007 Chemical Analysis by Optical Emission Spectrometer	0.01 % to 8 % 0.001 % to 15.00 % 0.01 % to 45.00 % 0.001 % to 6.10 % 0.005 % to 31.50 % 0.01 % to 15.00 % 0.01 % to 3.50 % 0.01 % to 5.50 % 0.01 % to 0.95 % 0.01 % to 1.00 % 0.01 % to 0.18 % 0.01 % to 2.00 % 0.001 % to 1.20 % 0.01 % to 0.20 % 0.01 % to 0.50 % 0.01 % to 0.07 %

~~-X-X-X-X-X-X-X-X-X-X-X-X-~~

Pooja Singh
Convenor

N. Venkateswaran
Program Manager