

**Laboratory** Test Laboratory, Goldwyn Limited, 15, Noida Special Economic Zone, Noida, Uttar Pradesh

**Accreditation Standard** ISO/IEC 17025: 2005

**Discipline** Photometry Testing **Issue Date** 24.04.2015

**Certificate Number** T-3386 **Valid Until** 23.04.2017

**Last Amended on** - **Page** 1 of 1

S.No.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
<b>I.</b>	<b>LIGHT SOURCES &amp; LUMINAIRES</b>			
1.	<b>LED Lamp, LED Luminaires &amp; Accessories</b>	LED & LED Luminaire Lumen measurement	LM-79-08 Cl. 9.0 IS 16106:2012 Cl. 11.2	5 lm to 3000 lm
		CCT measurement	LM-79-08 IS 16106: 2012	2700 K to 8000 K
		Chromaticity		
		X coordinate	LM-79-08 Cl. 12.0	0 to 1
		Y coordinate	IS 16106:2012 Cl. 14	0 to 1

-X-X-X-X-X-X-X-X-X-X-X-X-