

Laboratory Womika Coal & Minerals Technologies Pvt. Limited, 108, Pandurang
Gawande Layout, Nagpur, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number TC-5557

Page 1 of 1

Validity 16.03.2018 to 15.03.2020

Last Amended on --

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
-----	----------------------------	-------------------------	---	--

CHEMICAL TESTING

I.	SOLID FUELS			
1.	Coal	Moisture	IS 1350 (Part 1)	1 % to 12 %
		Ash	IS 1350 (Part 1)	2 % to 60 %
		Volatile Matter	IS 1350 (Part 1)	10 % to 40 %
		Fixed Carbon	IS 1350 (Part 1)	15.0 % to 45.0 %
		Gross Calorific Value	IS 1350 (Part 2)	2000 kcal/kg to 8000 kcal/kg
		Sulphur	IS 1350 (Part 3)	0.05 % to 10.00 %