

Laboratory Coal Testing Laboratory, MSPGCL, Koradi T.P.S. Koradi,
District Nagpur, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number TC-5716 (*in lieu of T-3117*)

Page 1 of 1

Validity 18.09.2016 to 17.09.2018

Last Amended on 30.10.2017

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
-----	----------------------------	-------------------------	---	--

CHEMICAL TESTING

I.	SOLID FUELS			
1.	Coal	Proximate Analysis		
		Moisture	IS 1350 (Part 1): 1984 (RA 2013)	1% to 20%
		Ash	IS 1350 (Part 1): 1984 (RA 2013)	10% to 70%
		Volatile Matter	IS 1350 (Part 1): 1984 (RA 2013)	1% to 40%
		Determination of GCV of Coal	IS 1350 (Part 2): 1970 (RA 2015)	2000 Kcal/kg to 7500 Kcal/kg

Ruchi Guntuku
Convenor

N. Venkateswaran
Program Director