

Laboratory Area Coal Testing Laboratory, EDC Siding Pench Area, WCL, Post-Parasia, Dist- Chhindwara, Madhya Pradesh
Accreditation Standard ISO/IEC 17025: 2005
Certificate Number TC-5973 **Page 1 of 1**
Validity 11.07.2017 to 10.07.2019 **Last Amended on --**

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
-----	----------------------------	-------------------------	---	--

CHEMICAL TESTING

I.	SOLID FUELS			
1.	Coal	Moisture	IS 1350	1 % to 26%
		Ash	IS 1350	5 % to 74 %
		Volatile Matter	IS 1350	10 % to 40%
		Gross Calorific Value	ASTM-D5865-13	2000 Cal/g to 8000 Cal/g

Gaurav Saini
Convenor

N. Venkateswaran
Program Director