

Laboratory Rabies Diagnostic Laboratory, Block II, III Floor, Department of Animal Biotechnology, Madras Veterinary College, TANUVAS, Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number TC-7839

Page 1 of 1

Validity 11.09.2018 to 10.09.2020

Last Amended on --

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
-----	----------------------------	-------------------------	---	--

BIOLOGICAL TESTING

I.	VETERINARY TESTING			
1.	Brain impression smears/ corneal impression smears in animals	Rabies antigen detection by Direct Fluorescent Antibody Technique (DFAT)	OIE Terrestrial Manual, 2018, Vol. 1, Chapter 2.1.17, Page No. 4	Qualitative (Positive or Negative for rabies antigen)
2.	Serum samples from animals	Anti-rabies antibody detection by Fluorescent antibody neutralization assay (FAVN)	OIE Terrestrial Manual, 2018, Vol. 1, Chapter 2.1.17, Page No. 14	Qualitative ≥0.5 IU/ml - Positive for rabies antibody <0.5 IU/ml - Negative for rabies antibody