

Laboratory Yathva Energy Solutions Private Limited, Plot No. 29, 10th Cross,
Bhairava Nagar, Anantapuramu, Andhra Pradesh

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number TC-7868

Page 1 of 2

Validity 26.09.2018 to 25.09.2020

Last Amended on --

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
-----	----------------------------	-------------------------	---	--

ELECTRICAL TESTING

AT LABORATORY				
I.	ELETRICAL INDICATING & RECORDING INSTRUMENTS			
1.	AC Static Watthour meters Class 1, 2, 1.0 _s , 0.5 _s & 0.2 _s (Single phase & 3 phase)	Test of Limits of Error	IS 13779 IS 14697	Voltage: 5 V to 520 V Current: 1 mA to 120 A Frequency: 47 Hz to 53 Hz Power: 520 V x 120 A Active/Reactive Energy: 1Wh to 28KWh

Laboratory Yathva Energy Solutions Private Limited, Plot No. 29, 10th Cross,
Bhairava Nagar, Anantapuramu, Andhra Pradesh

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number TC-7868

Page 2 of 2

Validity 26.09.2018 to 25.09.2020

Last Amended on --

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
-----	----------------------------	-------------------------	---	--

ELECTRICAL TESTING

AT SITE				
I.	ELETRICAL INDICATING & RECORDING INSTRUMENTS			
1.	AC Static Watthour meters Class 1, 2, 1.0 _s , 0.5 _s & 0.2 _s (Single phase & 3 phase)	Test of Limits of Error	IS 13779 IS 14697	Voltage: 5 V to 520 V Current: 1mA to 120 A Frequency: 47 Hz to 53 Hz Power: 520 V x 120 A Active/Reactive Energy: 1Wh to 28KWh