

<b>Laboratory</b>	<b>61850 Innovation Centre Integration Laboratory, ABB Ltd., HSB 115, Special Type, III<sup>rd</sup> Cross, Peenya Industrial Estate (1st Stage), Bangalore, Karnataka</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro-Technical Calibration</b>	<b>Issue Date</b>	<b>03.02.2015</b>
<b>Certificate Number</b>	<b>C-0635</b>	<b>Valid Until</b>	<b>02.02.2017</b>
<b>Last Amended on</b>	<b>-</b>	<b>Page</b>	<b>1 of 3</b>

	<b>Quantity Measured/ Instrument</b>	<b>Range / Frequency</b>	<b>*Calibration Measurement Capability (<math>\pm</math>)</b>	<b>Remarks</b>
<b>I. SOURCE</b>				
<b>1</b>	<b>DC VOLTAGE</b>	10 mV to 320 mV 320 mV to 3 V 3 V to 1000 V	0.1% to 0.01% 0.01% to 0.02% 0.02% to 0.03%	Using Wavetek 9100 by Direct Method
<b>2</b>	<b>AC VOLTAGE</b>	<b>50 Hz to 1 kHz</b> 10 mV to 32 mV 32 mV to 320 mV 320 mV to 100 V 100 V to 1000 V	4.5% to 0.4% 0.4% to 0.06% 0.06% 0.073	Using Wavetek 9100 by Direct Method
<b>3</b>	<b>DC CURRENT</b>	100 $\mu$ A to 320 $\mu$ A 0.32 mA to 0.32 A 0.32 A to 3.2 A 3.2 A to 20 A	0.03% to 0.02% 0.02% to 0.03% 0.03% to 0.08% 0.08% to 0.10%	Using Wavetek 9100 by Direct Method
<b>4</b>	<b>AC CURRENT</b>	<b>50 Hz to 1 kHz</b> 100 $\mu$ A to 300 $\mu$ A 0.3 mA to 320 mA 0.32 A to 3.2 A 3.2 A to 20 A	1.2% to 0.2% 0.2% to 0.1% 0.1% to 0.2% 0.2% to 0.3%	Using Wavetek 9100 by Direct Method
<b>5</b>	<b>RESISTANCE</b>	1 $\Omega$ to 40 $\Omega$ 40 $\Omega$ to 400 $\Omega$ 400 $\Omega$ to 400 k $\Omega$ 0.4 M $\Omega$ to 100 M $\Omega$  100 M $\Omega$ to 900 M $\Omega$	0.2% to 0.07% 0.07% to 0.03% 0.03% to 0.04% 0.04% to 0.2%  2.5%	Using Wavetek 9100 by Direct Method  Using Decade Mega Ohm Box 8400 HV
<b>6</b>	<b>FREQUENCY</b>	10 Hz to 1 MHz	0.0032%	Using Wavetek 9100 by Direct Method

**Bibin Philip**  
Convenor

**Avijit Das**  
Program Manager

<b>Laboratory</b>	<b>61850 Innovation Centre Integration Laboratory, ABB Ltd., HSB 115, Special Type, III<sup>rd</sup> Cross, Peenya Industrial Estate (1st Stage), Bangalore, Karnataka</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro-Technical Calibration</b>	<b>Issue Date</b>	<b>03.02.2015</b>
<b>Certificate Number</b>	<b>C-0635</b>	<b>Valid Until</b>	<b>02.02.2017</b>
<b>Last Amended on</b>	<b>-</b>	<b>Page</b>	<b>2 of 3</b>

	<b>Quantity Measured/ Instrument</b>	<b>Range / Frequency</b>	<b>*Calibration Measurement Capability (<math>\pm</math>)</b>	<b>Remarks</b>
<b>MEASURE</b>				
<b>7</b>	<b>DC VOLTAGE</b>	1 mV to 10 mV 10 mV to 1000 V	0.5% to 0.05% 0.05% to 0.01%	Using 6 ½ DMM Fluke 8646A by Direct Method
<b>8</b>	<b>AC VOLTAGE</b>	<b>10 Hz to 20 kHz</b> 100 mV to 100 V	0.2% to 0.12%	Using 6 ½ DMM Fluke 8646A by Direct Method
		<b>45 Hz to 1 kHz</b> 100 V to 1000 V	0.12%	
		<b>50 Hz</b> 1 kV to 5 kV	7.8% to 6.7%	
<b>9</b>	<b>DC CURRENT</b>	100 $\mu$ A to 100 mA 100 mA to 10 A	0.1% 0.1% to 0.3%	Using 6 ½ DMM Fluke 8646A by Direct Method
<b>10</b>	<b>AC CURRENT</b>	<b>10 Hz to 1 kHz</b> 100 $\mu$ A to 400 mA	0.22% to 0.3%	Using 6 ½ DMM Fluke 8646A by Direct Method
		<b>45 Hz to 1 kHz</b> 400 mA to 10 A	0.3%	
		<b>50 Hz</b> 10 A to 30 A	0.09%	Using Yokogawa WT 3000 by Direct Method
<b>11</b>	<b>RESISTANCE</b>	1 $\Omega$ to 10 $\Omega$ 10 $\Omega$ to 10 M $\Omega$ 10 M $\Omega$ to 100 M $\Omega$	0.5% to 0.05% 0.05% 0.05% to 0.95%	Using 6 ½ DMM Fluke 8646A by Direct Method
<b>12</b>	<b>FREQUENCY</b>	10 Hz to 1 MHz	0.04%	Using 6 ½ DMM Fluke 8646A by Direct Method

**Bibin Philip**  
Convenor

**Avijit Das**  
Program Manager

**Laboratory** 61850 Innovation Centre Integration Laboratory, ABB Ltd., HSB 115, Special Type, III<sup>rd</sup> Cross, Peenya Industrial Estate (1st Stage), Bangalore, Karnataka

**Accreditation Standard** ISO/IEC 17025: 2005

**Discipline** Electro-Technical Calibration **Issue Date** 03.02.2015

**Certificate Number** C-0635 **Valid Until** 02.02.2017

**Last Amended on** - **Page** 3 of 3

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
13 PHASE ANGLE	50 Hz 0° to 360°	0.12 °	Using Yokogawa WT 3000 by Direct Method

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%