

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** Page **1 of 18**

Validity **25.04.2018 to 24.04.2020** Last Amended on **18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>ELECTRO TECHNICAL CALIBRATION</u>				
I.	MEASURE			
1.	DC Voltage [#]	1 mV to 100 mV 100 mV to 10 V 10V to 1000 V	0.92% to 0.011% 0.011% to 0.007% 0.007 % to 0.009%	Using 6 ½ DMM Fluke 8846 A By Direct Method
2.	AC Voltage [#]	50 Hz 3 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	1.64 % to 0.12 % 0.12 % to 0.11 % 0.11 % to 0.10 % 0.10% 0.11%	Using 6 ½ DMM Fluke 8846 A By Direct Method
3.	DC Current [#]	200 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 10 A	0.066 % to 0.07 % 0 07 % to 0.08% 0.065% 0.065% to 0.10% 0.10% to 0.2 %	Using 6 ½ DMM Fluke 8846 A By Direct Method
4.	AC Current [#]	50 Hz 33 μ A to 1 mA 1 mA to 3 A 3 A to 10 A	0.68 % to 0.57 % 0.57 % to 0.26% 0.26 % to 0.28 %	Using 6 ½ DMM Fluke 8846 A By Direct Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **2 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
5.	DC Resistance [#]	10μΩ to 100 μΩ 2 Ω to 10 Ω 10 Ω to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ	2.19 % to 0.81 % 0.61% to 0.13 % 0.13 % to 0.014 % 0.014 % to 0.079 % 0.079 % to 0.95 %	Using 6 ½ DMM Fluke 8846 A & Calibrator By Direct Method Using 6 ½ DMM Fluke 8846 A By Direct Method
6.	Time [#]	10 Sec to 30 min 30 min to 23 Hrs	0.5 Sec to 0.69 Sec 0.69 Sec to 3.54 sec	Using Digital Stop Watch By Comparison Method
7.	Temperature Simulation ^{\$} RTD Mode J,K,T Type Thermocouple R,S Type Thermocouple	(-)200°C to 800°C (-)200°C to 1370°C 0°C to 1760°C	0.04°C to 0.09 °C 0.20 °C to 0.50 °C 0.35 °C to 0.50 °C	Using Fluke 7526A Precision process Calibrator By Direct Method
8.	Frequency [#]	10 Hz to 1kHz	0.57% to 0.12%	Using 6 ½ DMM Fluke 8846 A By Direct Method
9.	AC High Voltage [*]	2 kV to 15 k V	3.6 % to 2.4 %	Using HV Probe with DMM By Direct Method
10.	DC High Voltage [*]	2 kV to 30 kV	3.34 % to 2.73 %	Using HV Probe with DMM By Direct Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **3 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
11.	Temperature Simulation* PT-100 Thermocouple (J,K,T,R,S Type)	(-)200°C to 800°C (-)200°C to 1750°C	0.16 °C to 0.46°C 0.38 °C to 1.17°C	Using Beamax calibrator By Direct Method
12.	Capacitance [§]	100µF to 1 nF	0.42 % to 6.67%	Using LCR Meter By Direct Method
13.	Inductance [§]	100µH to 1H	0.41 %	Using LCR Meter By Direct Method
II. SOURCE				
1.	DC Voltage [§]	1 mV to 330 mV 330 mV to 30 V 30 V to 1000 V	1.28% to 0.041% 0.041% to 0.014% 0.014 % to 0.023 %	Using Fluke 5080A Calibrator By Direct Method
2.	DC Voltage*	1 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 to 1000 V	1.36 % to 0.13 % 0.13% to 1.28 % 1.28% 1.28% 1.28%	Using Multifunction Calibrator with Fluke 8846 A 6 ½ DMM By Indirect Method
3.	AC Voltage [§]	50 Hz 10 mV to 33 mV 33 mV to 330 mV 330 mV to 1000 V	1.193 % to 0.63% 0.63 % to 0.42 % 0.42% to 0.20%	Using Fluke 5080A Calibrator By Direct Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **4 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
4.	AC Voltage*	3 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	1.63 % to 0.4 % 0.4 % to 0.37 % 0.37% 0.37 % to 0.40 % 0.40 % to 1.27 %	Using Multifunction Calibrator with Fluke 8846 A 6 ½ DMM By Indirect Method
5.	DC Current ^s	3.3 mA to 330 mA 330 mA to 3 A 3 A to 20 A 20 A to 1000 A	0.097 % to 0.072 % 0.072 % to 0.24 % 0.24 % to 0.62 % 1.07%	Using Fluke 5080A Calibrator By Direct Method Using Fluke 5080A Calibrator and Current Coil By Direct Method
6.	DC Current*	1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 10 A	1.27% 1.27% to 1.28 % 1.28 % 1.28 % to 0.23 %	Using Multifunction Calibrator with Fluke 8846 A 6 ½ DMM By Indirect Method
7.	AC Current ^s	50 Hz 3.3 mA to 330 mA 330 mA to 3 A 3 A to 20 A 20 A to 1000 A	0.31 % to 0.22 % 0.22 % to 0.27 % 0.27 % to 0.89 % 0.89 % to 0.99 %	Using Fluke 5080A Calibrator By Direct Method Using Fluke 5080A Calibrator and Current Coil By Direct Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **5 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
8.	AC Current*	50 Hz 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 10 A	0.65 % 0.65 % to 0.39 % 0.39% 0.39 % to 0.45 %	Using Multifunction Calibrator with Fluke 8846 A 6 ½ DMM By Indirect Method
9.	DC Resistance (4 Wire Method) [§]	10µΩ 100µΩ 1 mΩ 10 mΩ 100 mΩ.	0.51µΩ 1.5µΩ, 0.066 mΩ, 0.067 mΩ 0.075 mΩ	Using Micro ohm(Ω) meter calibrator By Direct Method
		DC Resistance (2 Wire Method) [§]	1 mΩ to 1 Ω 1 Ω to 9 kΩ	5.9% to 0.21% 0.21%
	10 kΩ to 100kΩ		0.06 % to 2.38%	Using Decade Box ByDirect Method
	100 kΩ to 1000 MΩ		2.38 % to 2.42 %	Using Vaiseshika Deacade Box By Direct Method
10.	Resistance *	2 Ω to 4000 Ω	0.34 % to 1.8 %	Using Beamax Calibrator Direct Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **6 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
11	Temperature Indicator / Controller [§] PT-100 Thermocouple J,K,T Type R,S Type	(-)200°C to 630°C (-)200°C to 1370°C 0°C to 1750°C	0.08 °C 0.18 °C to 0.50°C 1.17 °C to 0.93 °C	Using Fluke 7526A Calibrator Direct Method
12	Temperature Indicator / Controller [*] Pt-100 Thermocouple (J, K, R, S, T Type)	(-)200°C to 800°C (-)200°C to 1750°C	0.21 °C to 0.49°C 0.36 °C to 1.2 °C	Using Beamax Calibrator Direct Method
13.	Ac Power [#]	At 50 Hz PF -0.5 to 0.5 120 W to 1.2 kW (240 V, 1A to 5 A)	1.14%	Using Load Manager By Comparison Method
14.	Frequency [#]	40 Hz to 1000 Hz	0.93% to 0.26%	Using Multifunction Calibrator Fluke 725 By Direct Method
15.	Capacitance [§]	1nF to 1µF 1µF to 100 µF	3.5%	Using Decade Capacitance Box By Direct Method
16.	Inductance [§]	100µH to 1H	2.89 % to 2.8 %	Using Decade Inductance Box By Direct Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **7 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>MECHANICAL CALIBRATION</u>				
I. DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)				
1.	Calipers ^s (Vernier / Digital / Dial) L.C : 0.01 mm ^φ	0 to 600 mm	15.4 μ m	Using Caliper Checker & Slip Gauge
2.	Height Gauge ^s (Vernier / Digital / Dial) L.C : 0.01 mm ^φ	0 to 600 mm	17 μ m	Using Caliper Checker & Slip Gauge
3.	External Micrometer ^s L.C : 0.001 mm ^φ	0 to 150 mm	2.3 μ m	Using Micrometer Check Set , Gauge Block Set
	L.C : 0.01 mm	150 mm to 600 mm	8.8 μ m	Using Micrometer Check Set , Long Gauge Block Set
4.	Plunger Dial ^b L.C : 0.001 mm ^φ L.C : 0.01 mm	0 to 10 mm	1 μ m	Using Single Axis Measuring Machine
		0 to 50 mm	6 μ m	
5.	Lever Dial ^s L.C: 0.001 mm ^φ L.C : 0.01 mm	0 to 0.14 mm	1.3 μ m	Using Single Axis Measuring Machine
		0 to 1.6 mm	6 μ m	

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **8 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
6.	Bore Gauge ^s (For Transmission Error) L.C : 0.001 mm ^φ	Up to 1 mm	1.3 μ m	Using Single Axis Measuring Machine
7.	Plain Plug Gauge ^s	Up to 225 mm	1.9 μ m	Using Single Axis Measuring Machine
8.	Plain Ring Gauge ^s	3mm to 200 mm	2.1 μ m	Using Single Axis Measuring M/c , Machine , Master Ring Gauge
9.	Thread Plug Gauge ^s	3 mm to 225 mm	2.1 μ m	Using Single Axis Measuring, Machine, Thread Measuring Wire
10.	Thread Ring Gauge ^s (Effective Diameter)	3mm to 100 mm	1.9 μ m	Using Single Axis Measuring Machine , Master Ring Gauge
11.	Snap Gauge ^s	Up to 150 mm	1.4 μ m	Using Single Axis Measuring Machine
12.	Dial Snap Gauge ^s	Up to 150 mm	6 μ m	Using Gauge Block
13.	Dial Thickness Gauge ^s L.C : 0.01 mm ^φ L.C : 0.001 mm	Up to 20 mm 0 to 1 mm	5.8 μ m 1 μ m	Using Gauge Block Set

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **9 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
14.	Feeler Gauge [§]	Up to 3 mm	2.3 μ m	Using Digital Micrometer
15.	Measuring Pin / Thread Measuring Wire [§]	0.17 to 20 mm	2.3 μ m	Using Single Axis Measuring Machine
16.	Thickness Foils [§]	Up to 2 mm	2.3 μ m	Using Digital Micrometer
17.	Measuring Scale [§]	Up to 1000 mm	290 μ m	Using Tape & Scale Calibrator
18.	Measuring Tape [§]	Up to 50 mtr	290 \sqrt{L} L is in mtr	Using Tape & Scale Calibrator
19.	Bevel Protractor [§] (L.C : 1 Min) ^ϕ	0 – 90° - 0 - 90°	1 Min	Using Angle Gauge Block By Comparison method
20.	Degree Protractor [§] (L.C : 1 Degree) ^ϕ	0 - 90° - 0	35 Min	Using Angle Gauge Block By Comparison Method
21.	Combination Set [§] (L.C : 1 Degree) ^ϕ	0 - 90°-0	35 Min	Using Angle Gauge Block By Comparison method
22.	Coating Thickness Gauge [§]	Up to 2 mm	6.2 μ m	Using Thickness Foils By Comparison Method
23.	Radius Gauge [§]	Up to R 40 mm	11.8 μ m	Using Vision Inspection System By Comparison Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **10 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
24.	Thread Pitch Gauge ^s (Pitch & Angle)	0.4 to 6 mm 60 ° & 55 °	6.8 μ m 4.1 Min arc	Using Vision Inspection System by Comparison method
25.	Test Sieves ^s	0.032 to 50 mm 50 mm to 125 mm	6.4 μ m 28 μ m	Using Vision Inspection System Digital Caliper by Comparison Method
26.	Pie Tape ^s	1 Mtr to 50 mtr	290 \sqrt{L} L is in Mtr	Using Tape & Scale Calibrator By Comparison Method
27.	Micrometer Setting Standard ^s	25 mm to 300 mm 300 mm to 575 mm	2.8 μ m 5.7 μ m	Using Single Axis Measuring M/c , Machine Comparator Stand , Digital Dial , Long Gauge Block By Comparison Method
28.	Tachometer ^s (Contact)	10 to 5000 Rpm	5.2% to 0.15 % rdg	Using Digital Tachometer By Comparison method
29.	Tachometer ^s (Non Contact)	10 to 13000 Rpm	0.21% to 0.36 % rdg	Using Digital Tachometer By Comparison method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **11 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
II. DIMENSION (PRECISION INSTRUMENTS)				
1.	Profile Projector / Optical Microscope / Tool Makers Microscope / Vision Measuring M/C * (L.C : 0.0001 mm) ^Φ	Linear 0 - 200 mm	6.3 μm	Glass Scale, Glass Scale / Digital Caliper, Angle Gauge By Comparison Method (JIS B 7184)
		Magnification 10 X to 100 X	0.6%	
		Angular 0 - 360°	4.7 Min	
III. PRESSURE INDICATING DEVICES				
1.	(Pneumatic Pressure Gauge) Digital & Dial Pressure Gauge, Pressure Transmitter, Pressure Switch #	0 to 10 Mbar	0.02 mbar	Using Multifunction Calibrator, Using Digital Pressure Gauge & Pressure Comparator, Using Digital Pressure Gauge & Pressure Comparator By Comparison Method using DKD R-6-1
		0 to 50 Mbar 0 to 30 Bar	0.63 mbar 0.01 bar	
2.	(Hydraulic Pressure Gauge) Digital & Dial Pressure Gauge, Pressure Transmitter, Pressure Switch \$	0 to 200 Bar	0.1 bar	Using Digital Pressure Gauge & Pressure Comparator, Using Dead Weight Tester By Comparison Method as per DKD R-6-1
		0 to 60 Bar	0.06 % rdg	
		60 Bar to 1000 Bar	0.06 % rdg	

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **12 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
3.	Vacuum Dial / Digital Gauges, [#]	(-) 0.9 to 0 Bar	0.003 bar	Using Digital Vacuum Gauge & Comparator By Comparison Method as per DKD R-6-2
4.	Barometer ^{\$}	940 to 1030 Mbar	1.7 mbar	Using Digital Pressure Gauge & Pressure Comparator By Comparison Method
5.	(Hydraulic Pressure Gauge) Digital & Dial Pressure Gauge, Pressure Transmitter, Pressure Switch [*]	0 to 300 Bar	0.19 bar	Using Digital Pressure Gauge & Pressure Comparator By Comparison Method as per DKD R-6-1
IV.	ACCOUSTICS			
1.	Sound Level Meter ^{\$}	94 & 114 Db	0.8 dB	Using Sound level Calibrator By Direct Method
2.	Amplitude/ Vibration Meter ^{\$} Velocity Displacement Acceleration	 2 mm/s to 20 mm/s 10 μ m to 145 μ m 2.7 m/s ² to 45 m/s ²	 3.85% 5.3 μ m 4.7 %	Using Vibration meter By Comparison Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **13 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
V.	WEIGHTS			
1	Mass-Weights ^s	1 mg	0.003 mg	Using E1 class weights and balance of readability 0.001mg as per OIML R-111
		2 mg	0.003 mg	
		5 mg	0.003 mg	
		10 mg	0.003 mg	
		20 mg	0.003 mg	
		50 mg	0.002 mg	
		100 mg	0.002 mg	
		200 mg	0.002 mg	
		500 mg	0.002 mg	
		1 G	0.004 mg	
2 G	0.004 mg			
5 G	0.006 mg			
10 G	0.02 mg	Using E1 class weights and balance of readability 0.01mg as per OIML R-111		
20 G	0.02 mg			
50 G	0.03 mg			
100 G	0.03 mg			
200 G	0.05 mg			
500 G	1 mg	Using F1 class weights and balance of readability 1mg as per OIML R-111		
1 kg	3 mg			
2 kg	10 mg	Using F1 class weights and balance of readability 10 mg as per OIML R-111		
5 kg	30 mg			
10 kg	100 mg	Using F1 class weights and balance of readability 100 mg as per OIML R-111		
20 kg	300 mg			

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **14 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
		50 kg	2 g	Using M1 class weights and balance of readability 1 g as per OIML R-111
VI.	VOLUME			
1.	Micro-Pipette ^s (L.C : 0.01 mg) ^Φ	1µl to 10 ml @ 27°C 10µl to 100 ml @ 27°C 100µl to 1000 ml @ 27°C	0.10 µl 0.11 µl 0.60 µl	Using Digital Micro balance with readability 0.001 mg Gravimetric method based on ISO 8655
2.	Glassware ^s Like Pipettes, Burettes, Measuring Cylinder, Volumetric Flask Etc	1ml to 10ml @27°C >10ml to 200ml @27°C >200ml to 1000ml @ 27°C 1000ml to 5ltr @27°C >5 Ltr to 20ltr @27°C	6.1 µl 6.6 µl 0.4ml 0.4 ml 6 ml	Using Weighing balance of 200g capacity and 0.01mg readability and distilled water By Gravimetric method based on IS/ISO 4787
VII.	WEIGHING SCALE AND BALANCE			
1.	Mass-Electronic Weighing Balances With Readability* L.C. : 0.001 mg ^Φ	Maximum capacity up to 5g	0.06 mg	Using E1 class weights By Calibration of electronic weighing balance of Class I and coarser As per OIML R-76-1
	L.C. : 0.01 mg ^Φ	Maximum capacity up to 100g	0.43 mg	

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** Page **15 of 18**

Validity **25.04.2018 to 24.04.2020** Last Amended on **18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	L.C : 0.1 mg ^ϕ	Maximum capacity up to 200g	0.50 mg	Using E1 & E2 class weights By Calibration of electronic weighing balance of Class I and coarser As per OIML R-76-1
	L.C : 1 mg ^ϕ	Maximum capacity up to 1000g	7 mg	Using E2 & F1 class weights by Calibration of electronic weighing balance of Class II and coarser As per OIML R-76-1
	L.C : 10 mg ^ϕ	Maximum capacity up to 10 kg	40 mg	
	L.C : 100 mg ^ϕ	Maximum capacity up to 25 kg	250 mg	
	L.C : 1 g ^ϕ L.C : 10g ^ϕ	Maximum capacity up to 100 kg	3 g 24 g	Using F1 and M1 class weights Calibration of electronic weighing balance of Class III and coarser As per OIML R-76-1

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **16 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
<u>THERMAL CALIBRATION</u>				
1.	RTD/ Thermocouple With/ Without Controller / Indicator / Analogue Temp.Gauge [§]	(-)80 °C to 50 °C 50 °C to 250 °C 250 °C to 600 °C	0.26°C 0.42°C 0.48 °C	Using DAQS with SPRT Sensor By Comparison Method
2.	Thermocouple With /Without Controller / Indicator / Dry Bath [§]	600 to 1200 °C	2.88 °C	Using DAQS with "S" Type thermocouple Sensor By Comparison Method
3	RTD/ Thermocouple With/ Without Controller / Indicator / Dry Bath,Liquid Bath Analogue Temp.Gauge [*]	(-)80 °C to 50 °C 50 °C to 250 °C 250 °C to 600 °C	0.28 °C 0.72 °C 0.72 °C	Using 6.5 digit Multimeter with RTD Sensor By Comparison Method
4	Thermocouple With /Without Controller / Indicator / Dry Bath [*]	600 °C to 1200°C	3.47 °C	Using Beamax calibrator with "S" Type thermocouple Sensor By Comparison Method
5.	Glass Thermometer [§]	(-)30 to 200 °C	1.24 °C	Using Temperature indicator with Sensor By Comparison Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta, Vapi, Dist: Valsad, Gujarat**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2657 (In lieu of C-0425,C-0426, C-0427)** **Page** **17 of 18**

Validity **25.04.2018 to 24.04.2020** **Last Amended on 18.05.2018**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
6.	Liquid Bath, Dry Block [#]	(-)30 to 400 °C	1.48 °C	Using DAQS with SSPRT Sensor By Comparison Method
7.	Thermo Hygrometer [#]	10°C to 50 °C @ 46 %RH 20 to 30 % RH @40 °C 30 to 90 %RH @25 °C	0.40 °C 2.7% RH 1.87 %RH	Using Hygropalm ,Humidity Generator By Comparison Method
8.	Data Logger (Temperature) ^{\$}	-30°C to 50 °C	0.62 °C	Using PT100 with Indicator & Low Temp By Comparison Method.
9.	IR Thermometer ^{\$}	50 °C to 300 °C	0.68 °C	Using Temp indicator with Pt-100 Sensor & Black Body Source By Comparison Method
10.	Deep Freezer, Environmental Chamber [*] (Single Point)	(-)30 to 10 °C	1.17 °C	Using Indicator with Sensor By Comparison Method
11.	Oven ,Furnace, Deep Freezer Environmental Chamber [*] (Multi Point)	(-)20 °C to 100 °C	5.50 °C	Using indicator with sensor By Comparison Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Accurate Labs, 311, 312, Govinda Complex, G.I.D.C Char Rasta,
Vapi, Dist: Valsad, Gujarat

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2657 (In lieu of C-0425,C-0426, C-0427) Page 18 of 18

Validity 25.04.2018 to 24.04.2020 Last Amended on 18.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>FLUID FLOW CALIBRATION</u>				
1.	Velocity (Air) [§]	0.6 m/s to 20 m/s	7.5 %	Using Hot wire anemometer By Comparison using hot wire anemometer & Wind Tunnel

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

[§] Only in Permanent Laboratory

[¶] Only for Site Calibration

[#] The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

^ϕ Laboratory can also calibrate instruments/devices of coarser resolution / least count within the accredited range using same reference standard/ master equipment under the scope of accreditation.

Dheeraj Chawla
Convenor

Avijit Das
Program Director