

Laboratory Adarsha Calibration Services, 204/4, Ullal, Sir.M. Vishveswaraiya Layout, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 29.09.2014

Certificate Number C-1135 **Valid Until** 28.09.2016

Last Amended on 18.01.2016 **Page** 1 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
I. SOURCE			
1. DC Voltage ^{\$}	1 mV to 329 mV 329 mV to 1V 1 V to 1000 V	0.0036 mV to 0.027 mV 0.027 mV to 0.00006 V 0.00006 V to 0.067 V	Using Multi-function calibrator FLUKE 5500E by Direct Method
2. AC Voltage ^{\$}	50 Hz to 1 kHz 30mV to 300 mV 300 mV to 3 V 3 V to 1000 V	0.08 mV to 0.21 mV 0.21 mV to 0.0013 V 0.0013 V to 0.69 V	Using Multi-function calibrator FLUKE 5500E by Direct Method
3. DC Current ^{\$}	10 μ A to 3.29 mA 3.29 mA to 329 mA 329 mA to 10 A 10 A to 20 A 20 A to 550 A	0.06 μ A to 0.001 mA 0.001 mA to 0.093 mA 0.093 mA to 0.009 A 0.009 A to 0.4 A 0.4 A to 5.2 A	Using Multi-function calibrator FLUKE 5500E & Current Coil by Direct Method
4. AC Current ^{\$}	50Hz 100 μ A to 329 μ A 329 μ A to 329 mA 329 mA to 10 A 10 A to 20 A 20 A to 550 A	0.64 μ A to 0.83 μ A 0.83 μ A to 0.86 mA 0.86 mA to 0.33 A 0.33 A to 0.7 A 0.7 A to 5.8 A	Using Multi-function calibrator FLUKE 5500E & Current Coil by Direct Method
5. Resistance ^{\$} (2 Wire)	1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 290 M Ω	0.011 Ω to 0.019 Ω 0.019 Ω to 0.03 Ω 0.03 Ω to 0.0002 k Ω 0.0002 k Ω to 0.002 k Ω 0.002 k Ω to 0.02 k Ω 0.02 k Ω to 0.0002 M Ω 0.0002 M Ω to 0.008 M Ω 0.008 M Ω to 0.58 M Ω 0.58 M Ω to 2.1 M Ω	Using High Precision Decade Resistance box,(7400) by Direct Method

Neeraj Verma
Convenor

Avijit Das
Program Manager

Laboratory	Adarsha Calibration Services, 204/4, Ullal, Sir.M. Vishveswaraiya Layout, Bangalore, Karnataka		
Accreditation Standard	ISO/IEC 17025:2005		
Discipline	Electro-Technical Calibration	Issue Date	29.09.2014
Certificate Number	C-1135	Valid Until	28.09.2016
Last Amended on	18.01.2016	Page	2 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
	1 Ω to 10 Ω	0.003 Ω to 0.006 Ω	Using High Precision Decade Resistance box,(7400) by Direct Method
	10 Ω to 100 Ω	0.006 Ω to 0.06 Ω	
	100 Ω to 1k Ω	0.06 Ω to 0.0006 k Ω	
	1k Ω to 10k Ω	0.0006 k Ω to 0.006 k Ω	
	10k Ω to 0.1M Ω	0.006 k Ω to 0.00006 M Ω	
	0.1 M Ω to 1 M Ω	0.006 M Ω to 0.023 M Ω	Using High Stability Decade Mega ohm box.(8400) by Direct Method
	1 M Ω to 10 M Ω	0.023 M Ω to 0.244 M Ω	
	10 M Ω to 100 M Ω	0.244 M Ω to 2.603 M Ω	
	100 M Ω to 1G Ω	2.603 M Ω to 0.024 G Ω	
6. Frequency ^{\$}	45 Hz to 1.0M Hz	0.0012Hz to 0.006 MHz	Using Multi-function calibrator FLUKE 5500E by Direct Method
7. Capacitance ^{\$}	@ 1 kHz		Using Multi-function calibrator FLUKE 5500E by Direct Method
	0.35 nF to 1nF	0.015 nF to 0.017 nF	
	1 nF to 10 nF	0.017 nF to 0.07 nF	
	10 nF to 100 nF	0.07 nF to 0.4 nF	
	100 nF to 300 nF	0.4 nF to 1.2 nF	
	300 nF to 1 μ F	1.2 nF to 0.008 μ F	
	1 μ F to 10 μ F	0.008 μ F to 0.05 μ F	
	10 μ F to 100 μ F	0.05 μ F to 0.7 μ F	
8. TEMPERATURE SIMULATION ^{\$}			Using Multi-function Calibrator fluke5500E by Direct Method
K Type	-200 $^{\circ}$ C to 1350 $^{\circ}$ C	0.5 $^{\circ}$ C	
J Type	-200 $^{\circ}$ C to 1200 $^{\circ}$ C	0.32 $^{\circ}$ C	
B Type	600 $^{\circ}$ C to 1800 $^{\circ}$ C	1 $^{\circ}$ C	
E Type	-200 $^{\circ}$ C to 1000 $^{\circ}$ C	0.6 $^{\circ}$ C to 0.25 $^{\circ}$ C	
N Type	-200 $^{\circ}$ C to 1300 $^{\circ}$ C	0.5 $^{\circ}$ C	
R Type	0 $^{\circ}$ C to 1750 $^{\circ}$ C	1 $^{\circ}$ C	
S Type	0 $^{\circ}$ C to 1750 $^{\circ}$ C	1 $^{\circ}$ C	
T Type	-200 $^{\circ}$ C to 400 $^{\circ}$ C	0.7 $^{\circ}$ C to 0.2 $^{\circ}$ C	

Neeraj Verma
Convenor

Avijit Das
Program Manager

Laboratory Adarsha Calibration Services, 204/4, Ullal, Sir.M. Vishveswaraiya Layout, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 29.09.2014

Certificate Number C-1135 **Valid Until** 28.09.2016

Last Amended on 18.01.2016 **Page** 3 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
9. DC Voltage *	10 mv to 30 V	0.06 mV to 0.07 V	Using Handy Calibrator CA-150 by Direct method
10. DC Current *	2 mA to 20 mA	0.01 mA to 0.1 mA	Using Handy Calibrator CA-150 by Direct method
11. Resistance *	50 Ω to 450 Ω 450 Ω to 4.5 k Ω 4.5 k Ω to 45 k Ω	0.15 Ω to 0.2 Ω 0.2 Ω to 0.007 k Ω 0.007 k Ω to 0.12 k Ω	Using Handy Calibrator CA-150 by Direct method
12. TEMPERATURE SIMULATION *			
K Type	-200 $^{\circ}$ C to 1350 $^{\circ}$ C	1 $^{\circ}$ C	Using Handy Calibrator CA-150 by Direct method
J Type	-200 $^{\circ}$ C to 1200 $^{\circ}$ C	1 $^{\circ}$ C	
B Type	600 $^{\circ}$ C to 1800 $^{\circ}$ C	2 $^{\circ}$ C	
E Type	-200 $^{\circ}$ C to 1000 $^{\circ}$ C	1 $^{\circ}$ C	
N Type	-200 $^{\circ}$ C to 1300 $^{\circ}$ C	1.2 $^{\circ}$ C	
R Type	100 $^{\circ}$ C to 1500 $^{\circ}$ C	2.3 $^{\circ}$ C	
S Type	100 $^{\circ}$ C to 1500 $^{\circ}$ C	2.3 $^{\circ}$ C	
T Type	-200 $^{\circ}$ C to 400 $^{\circ}$ C	1 $^{\circ}$ C	
RTD(PT385)	-200 $^{\circ}$ C to 800 $^{\circ}$ C	0.6 $^{\circ}$ C	Using Handy Calibrator CA-150 by Direct method
II. MEASURE			
1. DC Voltage \$	1 mV to 100 mV 100 mV to 10 V 10 V to 1000 V	0.0041 mV to 0.009 mV 0.009 mV to 0.00035 V 0.00035 V to 0.06 V	Using 6½ Digit DMM FLUKE 8846A by Direct Method
2. AC Voltage v	50 Hz 100 V to 1000 V	0.16 V to 1.1 V	Using 6½ Digit DMM FLUKE 8846A by Direct Method

Neeraj Verma
Convenor

Avijit Das
Program Manager

Laboratory Adarsha Calibration Services, 204/4, Ullal, Sir.M. Vishveswaraiya Layout, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 29.09.2014

Certificate Number C-1135 **Valid Until** 28.09.2016

Last Amended on 18.01.2016 **Page** 4 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
3. DC Current ^{\$}	10 μ A to 100 μ A 100 μ A to 100 mA 100 mA to 10 A	0.036 μ A to 0.088 μ A 0.088 μ A to 0.064 mA 0.064 mA to 0.025 A	Using 6½ Digit DMM FLUKE 8846A by Direct Method
4. AC Current ^{\$}	50 Hz 30 μ A to 100 μ A 100 μ A to 100 mA 100 mA to 10 A	0.083 μ A to 0.25 μ A 0.25 μ A to 0.166 mA 0.166 mA to 0.025A	Using 6½ Digit DMM FLUKE 8846A by Direct Method
5. Resistance ^{\$}	1 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω	0.0036 Ω to 0.016 Ω 0.016 Ω to 0.00013 k Ω 0.00013 k Ω to 0.0013 k Ω 0.0013 k Ω to 0.0129 k Ω 0.0129 k Ω to 0.0001 M Ω 0.0001 M Ω to 0.0048 M Ω 0.0048 M Ω to 0.94 M Ω	Using 6½ Digit DMM FLUKE 8846A by Direct Method
6. Frequency ^{\$}	50 Hz to 1 MHz	0.009 Hz to 0.0002 MHz	Using 6½ Digit DMM FLUKE 8846A by Direct Method
7. Capacitance ^{\$}	@ 1 kHz 1 nF to 100 nF @ 100 Hz 100 nF to 100 μ F	0.005 nF to 0.41nF 0.41 nF to 0.3 μ F	Using Auto Compute LCR-Q Meter by Direct Method
8. Inductance ^{\$}	@ 1 kHz 100 μ H to 1 mH 1 mH to 10 mH 10 mH to 100 mH 100 mH to 1 H 1 H to 10 H	0.6 μ H to 0.004 mH 0.004 mH to 0.04 mH 0.04 mH to 0.4 mH 0.4 mH to 0.004 H 0.004 H to 0.053 H	Using Auto Compute LCR-Q Meter by Direct Method

Neeraj Verma
Convenor

Avijit Das
Program Manager

Laboratory Adarsha Calibration Services, 204/4, Ullal, Sir.M. Vishveswaraiya Layout, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 29.09.2014

Certificate Number C-1135 **Valid Until** 28.09.2016

Last Amended on 18.01.2016 **Page** 5 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
9. TEMPERATURE SIMULATION \$			
K Type	-200 °C to 1350 °C	0.5 °C	Using Multi-function Calibrator fluke 5500E by Direct Method
J Type	-200 °C to 1200 °C	0.32 °C	
B Type	600 °C to 1800 °C	1 °C	
E Type	-200 °C to 1000 °C	0.6 °C to 0.25 °C	
N Type	-200 °C to 1300 °C	0.5 °C	
R Type	0 °C to 1750 °C	1 °C	
S Type	0 °C to 1750 °C	1 °C	
T Type	-200 °C to 400 °C	0.7 °C to 0.2 °C	
10. Time \$	5 sec to 9000 sec	0.31 % to 0.093 %	Using Digital Time Totaliser by Direct Method
11. DC Voltage *	50 mV to 30 V	0.1 mV to 0.06 V	Using Handy Calibrator CA-150 Handy Calibrator by Direct Method
12. DC Current *	2 mA to 90 mA	0.06 mA to 0.122 mA	Handy Calibrator CA-150 by Direct Method
13. Resistance *	50 Ω to 450 Ω 450 Ω to 4.5 k Ω 4.5 k Ω to 45 k Ω	0.14 Ω to 0.38 Ω 0.38 Ω to 0.06 k Ω 0.06 k Ω to 0.09 k Ω	Using Handy Calibrator CA-150 by Direct Method
14. TEMPERATURE SIMULATION *			
K Type	-190 °C to 1350 °C	2.5 °C	Using Handy Calibrator CA-150 by Direct Method
J Type	-190 °C to 1200 °C	2.5 °C	
B Type	600 °C to 1800 °C	3.5 °C	
E Type	-200 °C to 1000 °C	2.5 °C	
N Type	-200 °C to 1300 °C	2.5 °C	
R Type	100 °C to 1500 °C	3.4 °C	
S Type	100 °C to 1500 °C	3.4 °C	
T Type	-200 °C to 400 °C	2.5 °C	
RTD (PT385)	-200 °C to 800 °C	1.2 °C	

Neeraj Verma
Convenor

Avijit Das
Program Manager

Laboratory Adarsha Calibration Services, 204/4, Ullal, Sir.M. Vishveswaraiya Layout, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration Issue Date 29.09.2014

Certificate Number C-1135 Valid Until 28.09.2016

Last Amended on 18.01.2016 Page 6 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
15. DC High Voltage *	1 kV to 5 kV	2.5 %	Using HV Probe by Direct Method
16. AC High Voltage *	50 Hz 1 kV to 5 kV	6 %	Using HV Probe by Direct Method
17. Time *	5 sec to 9000 sec	0.31 % to 0.093 %	Using Digital Time Totaliser by Comparison Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

\$Only in Permanent Laboratory

*Only for Site Calibration

Neeraj Verma
Convenor

Avijit Das
Program Manager