Laboratory	**Adept Advisors India Pvt. Ltd., 205 B/3, Cosmos Commercial Complex, New Shahupuri, Station Road, Kolhapur, Maharashtra**		
Accreditation Standard	**ISO/IEC 17025:2005**		
Discipline	**Mechanical Calibration**	**Issue Date**	**20.03.2016**
Certificate Number	**C-0812**	**Valid Until**	**19.03.2018**
Last Amended on	**-**	**Page**	**1 of 1**

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (±)	Remarks
I. PRESSURE AND VACUUM			
1. PRESSURE-HYDRAULIC♣ (Dial & Digital Pressure Gauge)	0 to 30 bar 30 bar to 700 bar	0.14 bar 0.7 bar	Using Digital Pressure Gauge with Hydraulic Pump by Comparison Method based on DKD-R-6-1
2. NEGATIVE PRESSURE-VACUUM♣ (Dial & Digital Pressure Gauge)	(-) 0.8 bar to 0 bar	0.07 bar	Using Digital Vacuum Gauge, With Vacuum Pump by Comparison Method based on DKD-R-6-1

*** Measurement Capability is expressed as an uncertainty (±) at a confidence probability of 95%**
♣Only for Site Calibration

Neeraj Verma
Convenor

Avijit Das
Program Manager