

Laboratory Alcalab Private Limited, 3rd Floor, Ashiana Trade Centre, Adityapur, Jamshedpur, Jharkhand

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 24.09.2015

Certificate Number C-0289 **Valid Until** 23.09.2017

Last Amended on - **Page** 1 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
<u>SOURCE</u>			
1. DC VOLTAGE [§]	0.1 mV to 100 mV	3.52 % to 0.01 %	Using DMM 8508A with Calibrator 5500A by Direct/Comparison Method
	100 mV to 1 V	0.01 % to 0.009 %	
	1 V to 100 V	0.009 %	
	100 V to 1000 V	0.009 % to 0.0066 %	
2. DC CURRENT [§]	100 μ A to 10 mA	0.65 % to 0.075 %	Using DMM 8508A with Calibrator 5500A by Direct/Comparison Method
	10 mA to 100 mA	0.0758 %	
	100 mA to 10A	0.075 % to 0.09 %	
3. DC HIGH CURRENT [§]	10 A to 550 A	0.9 %	Using DMM 8508A with Calibrator 5500A + 50 Turn Coil by Direct /Comparison Method
4. RESISTANCE [§]	10 m Ω to 1 Ω	0.15 % to 0.08 %	Using DMM 8508A with Calibrator 5500A & Decade Megohm Boxes (Model-8400 & 8400 HV) & Decade Resistance Box (Model-7400) by Direct /Comparison Method
	1 Ω to 10 Ω	0.08 %	
	10 Ω to 10 k Ω	0.08 %	
	10 k Ω to 10 M Ω	0.08 %	
	10 M Ω to 100 M Ω	0.08 %	
1 G Ω to 10 G Ω	6.2 %		
5. HIGH RESISTANCE [§] (for Insulation Tester)	0.5 M Ω to 10 G Ω (50V to 5000V)	2.75 % to 6.3 %	Using Decade Megohm Boxes (Model-8400 & 8400 HV) by Direct / Comparison Method
6. RESISTANCE [§] (for Earth Tester)	0.5 Ω to 200 Ω	2.3 % to 0.5 %	Using Decade Resistance Box (Model 7400) by Direct /Comparison Method

Naveen Jangra
Convenor

Avijit Das
Program Manager

Laboratory	Alcalab Private Limited, 3rd Floor, Ashiana Trade Centre, Adityapur, Jamshedpur, Jharkhand		
Accreditation Standard	ISO/IEC 17025:2005		
Discipline	Electro-Technical Calibration	Issue Date	24.09.2015
Certificate Number	C-0289	Valid Until	23.09.2017
Last Amended on	-	Page	2 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
7. AC VOLTAGE^{\$}	50 Hz to 10 kHz 1mV to 10 mV	0.5 %	Using DMM 8508A with Calibrator 5500A By Direct Method
	10 mV to 10 V 10 V to 100 V	0.5 % to 0.02 % 0.02 %	
8. AC CURRENT^{\$}	50 Hz to 10 kHz 100 V to 1000 V	0.02 %	Using DMM 8508A with Calibrator 5500A By Direct Method
	50 Hz to 10 kHz 100 μ A to 10 mA 10 mA to 100 mA 100 mA to 10 A	0.45 % to 0.15 % 0.15 % 0.15 % to 0.16 %	
9. AC HIGH CURRENT^{\$}	10 A to 550A	0.6 % to 0.9 %	Using DMM 8508A with Calibrator 5500A + 50 Turn Coil By Direct Method
10. POWER / ENERGY^{\$} (1$\Phi$ & 3Φ)	50 Hz 110 V to 250 V 1 A to 5 A 0.5 to 1.00 P.F. (Lead & Lag)	2.5 % to 0.3 %	Using Energy Meter (Conzerv) with Calibrator by Comparison Method
11. CAPACITANCE^{\$}	1 kHz 50 nF to to 1000 μ F	1.5 %	Using Calibrator 5500A By Direct Method
12. FREQUENCY^{\$}	10 Hz to 1 MHz	0.08 %	Using Calibrator 5500A By Direct Method
13. INDUCTANCE^{\$}	1 kHz 10 μ H to 1000 mH	2.9 % to 4.8 %	Using Decade Inductance Box (Zeal-ZSDIB) By Direct Method

Naveen Jangra
Convenor

Avijit Das
Program Manager

Laboratory Alcalab Private Limited, 3rd Floor, Ashiana Trade Centre, Adityapur, Jamshedpur, Jharkhand

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 24.09.2015

Certificate Number C-0289 **Valid Until** 23.09.2017

Last Amended on - **Page** 3 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
14. POWER FACTOR^s	45 Hz to 55 Hz (-)1 to 1	0.25 %	Using Energy Meter (Conzerv) with Calibrator by Comparison Method
15. TEMPERATURE SIMULATION[#]			
RTD TYPE	(-)200 °C to 600 °C	0.07 °C to 0.42 °C	Using Calibrator 5500A + ITS 90 (mV & Ω method) By Direct Method
Thermocouples			
J Type	0 °C to 750 °C	0.3 °C	
K Type	(-) 200 °C to 1250 °C	0.3 °C	
R Type	0 °C to 1600 °C	0.65 °C	
S Type	0 °C to 1600 °C	0.65 °C	
16. DC VOLTAGE*	2 mV to 100 mV 100 mV to 1 V 1 V to 1000 V	0.92 % 0.92 % 0.92 % to 0.06 %	Using Calibrator 5500A By Direct/ Comparison Method
17. DC CURRENT*	100 μ A to 1 mA 1 mA to 100 mA 100 mA to 1 A 1 A to 10A	0.65 % 0.65 % 0.65 % 0.2 %	Using Calibrator 5500A By Direct/ Comparison Method
18. RESISTANCE*	1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 50 M Ω	1.12 % to 0.15 % 0.15 % 0.15 % 0.15 % 0.15 % to 0.6 %	Using Calibrator 5500A By Direct/ Comparison Method
19. HIGH RESISTANCE* (for Insulation Tester)	0.2 M Ω to 10 G Ω (50V to 5000V)	2.75 % to 6.3 %	Using Decade Megohm Boxes By Direct/ Comparison Method

Naveen Jangra
Convenor

Avijit Das
Program Manager

Laboratory Alcalab Private Limited, 3rd Floor, Ashiana Trade Centre, Adityapur, Jamshedpur, Jharkhand
Accreditation Standard ISO/IEC 17025:2005
Discipline Electro-Technical Calibration **Issue Date** 24.09.2015
Certificate Number C-0289 **Valid Until** 23.09.2017
Last Amended on - **Page** 4 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
20. RESISTANCE* (for Earth Tester)	0.2 Ω to 200 Ω	3.5 % to 0.5 %	Using Decade Resistance Box (Model 7400) By Direct/ Comparison Method
21. AC VOLTAGE*	50 Hz to 10 kHz 30 mV to 100 mV 100 mV to 10 V 10 V to 100 V	0.5 % to 0.68 % 0.68 % 0.68 %	Using Calibrator 5500A By Direct/ Comparison Method
	50 Hz to 1 kHz 100 V to 1000 V	0.15 % to 0.21 %	
22. AC CURRENT*	50 Hz 10 mA to 100 mA 100 mA to 10 A	0.35 % 0.35 % to 0.17 %	Using Calibrator 5500A By Direct/ Comparison Method
23. FREQUENCY*	50 Hz to 1 MHz	0.02 %	Using Calibrator 5500A By Direct/ Comparison Method
24. POWER / ENERGY* (1 Φ & 3 Φ)	50 Hz 110 V to 250 V 1 A to 5 A 0.5 to 1.00 P.F. (Lead & Lag)	3.0 % to 0.3 %	Using Energy Meter (Conzerv) By Direct/ Comparison Method
25. STOP WATCH / TIMER MECHANICAL / DIGITAL*			Using Master Stop Watch by Comparison Method
	L. C. : 0.01minute L. C. : 1 minute	0 to 60 m 0.5 second 35 second	

Naveen Jangra
 Convenor

Avijit Das
 Program Manager

Laboratory Alcalab Private Limited, 3rd Floor, Ashiana Trade Centre, Adityapur, Jamshedpur, Jharkhand
Accreditation Standard ISO/IEC 17025:2005
Discipline Electro-Technical Calibration **Issue Date** 24.09.2015
Certificate Number C-0289 **Valid Until** 23.09.2017
Last Amended on - **Page** 5 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
<u>MEASURE</u>			
1. DC VOLTAGE ^{\$}	0.1 mV to 100 mV	0.25 % to 0.005 %	Using DMM 8508 A By Direct Method
	100 mV to 1 V	0.005 % to 0.002 %	
	1 V to 1000 V	0.002 %	
2. DC CURRENT ^{\$}	100 μ A to 1 mA	0.005 % to 0.008 %	Using DMM 8508 A By Direct Method
	1 mA to 100 mA	0.008 % to 0.005 %	
	100 mA to 1 A	0.005 % to 0.025 %	
	1 A to 10 A	0.025 % to 0.052 %	
3. FREQUENCY ^{\$}	5 Hz to 1 MHz	0.15 % to 0.06 %	Using DMM 8508 A by Comparison Method
4. RESISTANCE ^{\$}	10 m Ω to 1 Ω	0.31 %	Using DMM 8508 A
	1 Ω to 10 Ω	0.31 %	
	10 Ω to 100 Ω	0.31 %	
	100 Ω to 100 M Ω	0.31 %	
	100 M Ω to 1 G Ω	0.31 % to 0.35 %	
5. AC VOLTAGE ^{\$}	50 Hz to 10 kHz	0.69 % to 0.09 % 0.09 % 0.013 % to 0.09 %	Using DMM 8508 A By Direct Method
	1 mV to 10 mV		
	10 mV to 10 V		
6. AC CURRENT ^{\$}	50 Hz to 10 kHz	0.15 % to 0.07 % 0.07 % to 0.14 %	Using DMM 8508 A By Direct Method
	100 μ A to 100 mA		
	100 mA to 10 A		

Naveen Jangra
 Convenor

Avijit Das
 Program Manager

Laboratory Alcalab Private Limited, 3rd Floor, Ashiana Trade Centre, Adityapur, Jamshedpur, Jharkhand

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 24.09.2015

Certificate Number C-0289 **Valid Until** 23.09.2017

Last Amended on - **Page** 6 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
7. DC HIGH VOLTAGE ^{\$}	1 KV to 5 KV	3.7 %	Using High Voltage Probe (Model - 80K-40) with 4½ DMM By Direct Method
8. AC HIGH VOLTAGE ^{\$}	1 KV to 20 KV	4.1 %	Using High Voltage Probe (Model-80K-40) with 4½ DMM By Direct Method
9. POWER / ENERGY ^{\$} (1Φ & 3Φ)	50 Hz 110 V to 250 V 1 A to 5 A 0.5 to 1.00 P.F. (Lead & Lag)	2.5 % to 0.3 %	Using Energy Meter (Conzerv) with Calibrator By Comparison Method
10. POWER FACTOR ^{\$}	45 Hz to 55 Hz (-)1 to 1	0.25 %	Using Energy Meter (Conzerv) with Calibrator by Comparison Method
11. TEMPERATURE SIMULATION ^{\$}			
RTD TYPE	(-)200 °C to 600 °C	0.07 °C to 0.42 °C	Using Fluke 5500A + ITS 90
Thermocouples			
J Type	0 °C to 750 °C	0.3 °C	
K Type	(-)200 °C to 1250 °C	0.3 °C	
R Type	0 °C to 1600 °C	0.65 °C	
S Type	0 °C to 1600 °C	0.65 °C	
12. STOP WATCH / TIMER ^{\$} (MECHANICAL / DIGITAL)			
L.C. 0.01 minute	0 to 60 Minutes	0.5 sec	Using Master Stop Watch
L.C. 1 minute		35 sec	By Comparison Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

^{\$}Only in Permanent Laboratory

^{*}Only for Site Calibration

[#]The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Naveen Jangra
Convenor

Avijit Das
Program Manager