

**Laboratory**                      **Alfatek Services, Sapthagiri, Tripadapuram Hill, Kulathoor P.O. Trivandrum, Kerala**

**Accreditation Standard**   **ISO/IEC 17025:2005**

**Discipline**                      **Mechanical Calibration**                      **Issue Date**    **25.10.2016**

**Certificate Number**        **C-0361**    **Valid Until**    **24.10.2018**

**Last Amended on**         **-**    **Page**                **1 of 4**

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
<b>I. WEIGHING SCALE AND BALANCE</b>			
<b>1. MASS CALIBRATION OF ELECTRONIC BALANCE OF DENSIMETER &amp; DIGITAL WEIGHING MACHINE* ( Resolution= 0.0001 g)</b>	0 to 200 g	0.5 mg	Using E2 class standard weights (1 mg to 200 g) as per OIML - R 76
<b>II. DIMENSION (Basic Measuring Instrument, Gauge ETC.)</b>			
<b>1. CALIBRATION OF RAPID PLASTIMETER FOR TIME INDICATOR<sup>#</sup></b>	15 s	0.74 s	Using Digital Stop watch as per ASTM D 3194-04
<b>2. PRI AGEING CHAMBER AGEING TIME<sup>#</sup></b>	30 min	0.74 s	Using Digital Stop watch as per ASTM D 3194-04
<b>3. CALIBRATION OF RAPID PLASTIMETER LVDT<sup>#</sup> (Length)</b>	0.25 mm to 1 mm	1.9 µm	Using Gauge blocks 'O ' grade as per ASTM D 3194-04

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Manager

**Laboratory**                      **Alfatek Services, Sapthagiri, Tripadapuram Hill, Kulathoor P.O. Trivandrum, Kerala**

**Accreditation Standard**   **ISO/IEC 17025:2005**

**Discipline**                      **Mechanical Calibration**                      **Issue Date**    **25.10.2016**

**Certificate Number**        **C-0361**    **Valid Until**    **24.10.2018**

**Last Amended on**            -    **Page**                **2 of 4**

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
----------------------------------	-------------------	--	---------

### III. TORQUE GENERATING DEVICES

1. TORQUE CALIBRATION OF MOONEY VISCOMETER BY KNOWN TORQUE*	100 Mooney Units (8.3 Nm)	0.084 Mooney units	Using certified weights, Dimensional Instruments as per ASTM D1646 clause 9
2. CALIBRATION OF OSCILLATING DISC RHEOMETER BY KNOWN TORQUE*	Specific Torques of 15.29 dN-m (13.53 lbs-in) @ 1° arc 45.62dN-m (40.37 lbs-in) @ 3° arc using Torque Std Sl.No. 12S593R  Specific Torques of 20.75 dN-m(18.37 lbs-in) @ 1° arc 62.02 dN-m (54.88 lbs-in) @ 3° arc using Torque Std. Sl.No. 140601TQ	0.24 dNm	Using Certified ODR Torque Standard as per ASTM D2084- clause 10
3. CALIBRATION OF MOVING DIE RHEOMETER BY KNOWN TORQUE*	Specific torque of 23.84dN-m (21.10lbs-in) @ 0.5° arc of oscillation, 47.66 dN-m ( 42.18lbs-in)@ 1° arc of oscillation using Torque Std. Sl.No. EKT 1284R  Specific torque of 20.89 dN-m (18.49 lbs-in) @ 0.5° arc of oscillation, 41.72 dN-m (36.92 lbs-in) @ 1° arc of oscillation using TS Sl.No. 130906	0.24 dNm	Using Certified MDR Torque Standard as per ASTM D5289-07a clause - 6.5

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Manager

**Laboratory**                      **Alfatek Services, Sapthagiri, Tripadapuram Hill, Kulathoor P.O. Trivandrum, Kerala**

**Accreditation Standard**   **ISO/IEC 17025:2005**

**Discipline**                      **Mechanical Calibration**                      **Issue Date**    **25.10.2016**

**Certificate Number**        **C-0361**                      **Valid Until**    **24.10.2018**

**Last Amended on**        **-**                      **Page**                **3 of 4**

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
----------------------------------	-------------------	--	---------

**IV. HARDNESS TESTING MACHINES**

1.	<b>CALIBRATION OF SHORE-A DUROMETER FOR SPRING FORCE<sup>§</sup></b>	0 to 100 Shore A	0.93 Shore A	Using Durocalibrator as per ASTM D 2240-05 / ISO 18898-2006
	<b>CALIBRATION OF SHORE-A DUROMETER FOR INDENTOR EXTENSION<sup>#</sup></b>	0 to 100 ShoreA	0.78 Shore A	Using Slip Gauges as per ASTM D 2240-05 / ISO 18898-2006
2.	<b>CALIBRATION OF IRHD MICRO HARDNESS TESTER (Method-M) INCREMENTAL INDENTATION DEPTH<sup>*</sup></b>	30 IRHD to 100 IRHD	0.18 IRHD	Using Dimensional Instruments based on ISO 18898/ISO 48
3.	<b>CALIBRATION OF IRHD MACRO HARDNESS TESTER (Method-N) INCREMENTAL INDENTATION DEPTH<sup>*</sup></b>	30 IRHD to 100 IRHD	0.32 IRHD	Using Dimensional Instruments based on ISO 18898/ISO 48

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Manager

**Laboratory**                      **Alfatek Services, Sapthagiri, Tripadapuram Hill, Kulathoor P.O. Trivandrum, Kerala**  
**Accreditation Standard**   **ISO/IEC 17025:2005**  
**Discipline**                      **Mechanical Calibration**                      **Issue Date**    **25.10.2016**  
**Certificate Number**        **C-0361**    **Valid Until**    **24.10.2018**  
**Last Amended on**         **-**    **Page**                **4 of 4**

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
<b>V. UTM, TENSION CREEP AND TORSION TESTING MACHINE</b>			
1. CALIBRATION OF FORCE OF UTM IN COMPRESSION MODE*	2.5 kN to 100 kN	0.28 %	Using Certified Load cell with Indicator as per ISO 7500/ IS 1828-2005 (Part 1)
2. CALIBRATION OF FORCE OF UTM IN TENSION MODE*	250 N to 9000 N	0.28 %	Using Certified Load cell with Indicator as per ISO 7500/ IS 1828-2005 (Part 1)

\* Measurement Capability is expressed as an uncertainty (±) at a confidence probability of 95%

\$Only in Permanent Laboratory

\*Only for Site Calibration

# The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

**Mohit Kaushik**  
**Convenor**

**Avijit Das**  
**Program Manager**