

Laboratory APT Kalibration, No. B-409, 4th Floor, B- Block, Pushp Ahuja Estate, Abids Road, Hyderabad, Telangana

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2426 **Page** 2 of 4

Validity 30.10.2017 to 29.10.2019 **Last Amended on** 01.01.2019

| Sl. | Quantity Measured / Instrument | Range/Frequency | *Calibration Measurement Capability (\pm) | Remarks |
|------|--|--------------------------------------|---|---|
| III. | DIMENSION (BASIC MEASURING INSTRUMENTS, GAUGE ETC.) | | | |
| 1. | Templates, Inspection JIG and Fixture/Cube Moulds [§] (Height, Thickness) | Upto 300 mm | 17.6 μ m | Using Profile Projector, Digital Vernier Caliper, Micrometer By Comparison Method. |
| 2. | Test Sieve [§] | 45 μ m to 4 mm 4 mm to 125 mm | 10 μ m 58 μ m | Using Profile Projector, Digital Vernier Caliper By Comparison Method as per IS 460 (Part I, II, III) |
| 3. | Thread Pitch Gauges [§] (Pitch, Flank Angle) | Upto 10 mm Flank Angle@60 deg | 7.7 μ m 5 min of arc | Using Profile Projector By Comparison Method as per IS 460 (Part I, II, III) |
| 4. | Radius gauges [§] (concave and convex profiles) | 0.6 mm to 25 mm | 7.7 μ m | Using Profile Projector By Comparison Method as per IS 5273 |
| 5. | Bevel Protractors, Combination Sets [§] (Error of indication) | Upto 360 ° | 8.3 min of arc | Using Profile Projector By Comparison Method as per IS 4239 |
| 6. | Vernier Caliper [§] L.C. 0.01 mm | Upto 600 mm | 14 μ m | Using Caliper Checker By Comparison Method as per IS 3651- I, II, III |
| 7. | Height Gauge [§] L.C. 0.01 mm | Upto 500 mm | 10 μ m | Using Caliper Checker By Comparison Method as per IS 2921 |

Laboratory APT Kalibration, No. B-409, 4th Floor, B- Block, Pushp Ahuja Estate,
Abids Road, Hyderabad, Telangana

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2426

Page

3 of 4

Validity 30.10.2017 to 29.10.2019

Last Amended on 01.01.2019

| Sl. | Quantity Measured / Instrument | Range/Frequency | *Calibration Measurement Capability (\pm) | Remarks |
|-----|---|---|---|--|
| 8. | Angle Graticule ^{\$} | Upto 360° | 8.3 min of arc | Using Profile Projector By Comparison Method as per JIS B 7541 |
| II. | WEIGHING SCALE AND BALANCE | | | |
| 1. | Electronic Weighing Balance* d \geq 0.01 g d \geq 1.00 g d \geq 10.0 g | 1g to 4500g 20g to 30kg 200g to 100kg | 0.07g 3.0 g 20.0 g | Using F1 Class standard weights 1g to 20 kg As per OIMLR-76 Calibration of weighing balance of Class-II & coarser. |

Laboratory APT Kalibration, No. B-409, 4th Floor, B- Block, Pushp Ahuja Estate, Abids Road, Hyderabad, Telangana

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2426 Page 4 of 4

Validity 30.10.2017 to 29.10.2019 Last Amended on 01.01.2019

| Sl. | Quantity Measured / Instrument | Range/Frequency | *Calibration Measurement Capability (\pm) | Remarks |
|-----------------------------------|---|--------------------|---|---|
| <u>THERMAL CALIBRATION</u> | | | | |
| I. | TEMPERATURE | | | |
| 1. | Temperature Indicator of Environmental Chamber, Furnace, Oven \$ (Single Point) | (-)35 °C to 250 °C | 0.87 °C | Using 4 Wire RTD Sensor with Temperature Indicator By Comparison Method |

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

\$Only in Permanent Laboratory

*Only for Site Calibration