

**Laboratory** Autocal Solutions Pvt. Ltd., Plot No. BH – 74, Gala No. 1 & 2, Jai Talja Bhavani Indl. Premises, Telco Road, MIDC – Bhosari, Pune, Maharashtra

**Accreditation Standard** ISO/IEC 17025: 2005

**Discipline** Electro-Technical Calibration **Issue Date** 11.09.2015

**Certificate Number** C-1268 **Valid Until** 10.09.2017

**Last Amended on** - **Page** 1 of 8

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
<b><u>SOURCE</u></b>			
<b>1. DC VOLTAGE<sup>#</sup></b>	0.1mV to 1 mV	3.58% to 0.36%	Using Multifunction Calibrator, Fluke-5500E by Direct Method
	1mV to 10 mV	0.36% to 0.042%	
	10 mV to 100 mV	0.042% to 0.010%	
	100 mV to 1 V	0.010 % to 0.0017 %	
	1 V to 100 V	0.0017% to 0.0024%	Using Multifunction Calibrator, Fluke-5520A by Direct Method
	100 V to 1000 V	0.0024% to 0.0025%	
<b>2. DC CURRENT<sup>#</sup></b>	10 µA to 100 mA	0.26% to 0.016%	Using Multifunction Calibrator, Fluke-5520A by Direct Method
	100 mA to 1 A	0.016% to 0.029%	
	1A to 10 A	0.029% to 0.064%	
	10 A to 20 A	0.064% to 0.12%	
	20 A to 1000 A	0.73% to 0.3%	Using Multifunction Calibrator, Fluke-5520A + 50Turn Current Coil, by Direct Method
<b>3. AC VOLTAGE<sup>#</sup></b>	<b>45 Hz to 10 kHz</b>		Using Multifunction Calibrator, Fluke-5500E by Direct Method
	2 mV to 10 mV	2.53% to 0.41%	
	10 mV to 100 mV	0.41% to 0.082%	
	100 mV to 1 V	0.082% to 0.03%	
	<b>45 Hz to 1 kHz</b>		Using Multifunction Calibrator, Fluke-5520A by Direct Method
	1 V to 10 V	0.03% to 0.026%	
	10 V to 1000 V	0.026% to 0.037%	
	<b>10 kHz to 100 kHz</b>		
	1 V to 100 V	0.1% to 0.29%	

Ram Ashray  
Convenor

Avijit Das  
Program Manager

<b>Laboratory</b>	<b>Autocal Solutions Pvt. Ltd., Plot No. BH – 74, Gala No. 1 &amp; 2, Jai Talja Bhavani Indl. Premises, Telco Road, MIDC – Bhosari, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro-Technical Calibration</b>	<b>Issue Date</b>	<b>11.09.2015</b>
<b>Certificate Number</b>	<b>C-1268</b>	<b>Valid Until</b>	<b>10.09.2017</b>
<b>Last Amended on</b>	<b>-</b>	<b>Page</b>	<b>2 of 8</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
4. AC CURRENT <sup>#</sup>	45 Hz to 1 kHz 30 µA to 3 A 3 A to 10 A 10 A to 20 A	0.62% to 0.08% 0.08% to 0.14% 0.14% to 0.20%	Using Multifunction Calibrator, Fluke-5520A by Direct Method
	50 Hz 20 A to 1000 A	0.67% to 0.30%	Using Multifunction Calibrator, Fluke-5520A + 50Turn Current Coil, by Direct Method
5. DC RESISTANCE <sup>#</sup>	100 mΩ to 100 kΩ	1.31% to 0.0036%	Using Multifunction Calibrator, Fluke-5520A by Direct Method
	100 kΩ to 1MΩ	0.0036% to 0.0043%	
	1MΩ to 10 MΩ	0.0043% to 0.016%	
	10 MΩ to 100 MΩ	0.016% to 0.063%	
	100 MΩ to 1GΩ	0.063% to 1.8%	
6. HIGH RESISTANCE <sup>#</sup>	0.1 Ω to 100 MΩ	2.31%	Using Standard Megohm Decade Box & Discrete High Resistance by Direct Method
	50 MΩ	2.32%	
	200 MΩ	2.32%	
	300 MΩ	2.32%	
	400 MΩ	2.32%	
	1 GΩ to 100 GΩ	2.33%	
7. CAPACITANCE <sup>#</sup>	1kHz 1 nF to 10 µF	1.73 % to 0.41 %	Using Multifunction Calibrator, Fluke-5520A by Direct Method
	10 µF to 100 µF	0.41 % to 0.64 %	
	100 µF to 1 mF	0.64 % to 0.84 %	
8. INDUCTANCE <sup>#</sup>	1kHz 0.1 mH	1.67%	Using Standard Discrete Inductance by Direct Method
	1 mH	1.16%	
	10 mH	1.16%	
	100 mH	1.16%	
	1000 mH	1.16%	

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Manager

<b>Laboratory</b>	<b>Autocal Solutions Pvt. Ltd., Plot No. BH – 74, Gala No. 1 &amp; 2, Jai Talja Bhavani Indl. Premises, Telco Road, MIDC – Bhosari, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro-Technical Calibration</b>	<b>Issue Date</b>	<b>11.09.2015</b>
<b>Certificate Number</b>	<b>C-1268</b>	<b>Valid Until</b>	<b>10.09.2017</b>
<b>Last Amended on</b>	<b>-</b>	<b>Page</b>	<b>3 of 8</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>9. TEMPERATURE INDICATOR /CONTROLLER BY SIMULATION METHOD<sup>#</sup> RTD (PT-100)</b>	-200 °C to 800 °C	0.05 °C	Using Multifunction Calibrator, Fluke-5520A Conversion by using ITS-90 Scale for $\Omega$ to °C and mV to °C
<b>Thermocouple (K, J, N, E, T, R, S, B, C, L, U)</b>	-200 °C to 1370 °C	0.12 °C	
<b>10. POWER FACTOR<sup>#</sup></b>	<b>50Hz</b> 0.1 PF to UPF Lead/ Lag	0.008 PF	Using Multifunction Calibrator, Fluke-5520A by Direct Method
<b>11. AC POWER<sup>#</sup> (1 phase)</b>	<b>50Hz</b>		
<b>@ UPF</b>	120 V to 240 V 0.01 A to 20 A 1.2 W to 4.8 kW	0.65% to 0.18%	Using Multifunction Calibrator, Fluke-5520A + 50Turn Current Coil, by Direct Method
<b>@ 0.8PF (Lag/ Lead)</b>	120 V to 240 V 0.1 A to 20 A 9.6 W to 3.8 kW	0.99% to 0.25%	
<b>@ 0.5PF (Lag/ Lead)</b>	120 V to 240 V 0.1 A to 20 A 6 W to 2.4 kW	1.20% to 0.39%	
<b>@ 0.2PF (Lag/ Lead)</b>	120 V to 240 V 0.1 A to 20 A 2.4 W to 960 W	3.09% to 1.16%	

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Manager

<b>Laboratory</b>	<b>Autocal Solutions Pvt. Ltd., Plot No. BH – 74, Gala No. 1 &amp; 2, Jai Talja Bhavani Indl. Premises, Telco Road, MIDC – Bhosari, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro-Technical Calibration</b>	<b>Issue Date</b>	<b>11.09.2015</b>
<b>Certificate Number</b>	<b>C-1268</b>	<b>Valid Until</b>	<b>10.09.2017</b>
<b>Last Amended on</b>	<b>-</b>	<b>Page</b>	<b>4 of 8</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>12. FREQUENCY<sup>#</sup></b>	10 Hz to 250 MHz	0.58% to 0.023%	Using Multifunction Calibrator, Fluke-5520A by Direct Method
<b>13. PERIOD<sup>#</sup></b>	0.1 s to 4 ns	0.577% to 0.023%	Using Multifunction Calibrator, Fluke-5520A by Direct Method
<b>14. OSCILLOSCOPE<sup>#</sup></b>			
<b>DC AMPLITUDE</b>	<b>50<math>\Omega</math> output</b> 5 mV to 2.2 V	1.91% to 0.46%	Using Multifunction Calibrator, Fluke-5520A by Direct Method
	<b>1M<math>\Omega</math> output</b> 5 mV to 33 V 33 V to 100 V	0.53% to 0.23% 1.75% to 0.78%	
<b>AC AMPLITUDE</b>	<b>1M<math>\Omega</math> output</b> <b>@ 1kHz (V p-p)</b> 5 mV p-p to 50 V p-p 50 V p-p to 100 V p-p	1.81% to 0.2% 0.2 % to 0.74%	
<b>TIME BASE</b>	<b>50<math>\Omega</math> output</b> 5 mV p-p to 2 V p-p	1.81% to 0.60%	
<b>BANDWIDTH</b>	2 ns to 10 ms 10 ms to 5 s  50 kHz to 250 MHz	0.32% to 0.64% 0.64% to 0.60%  3.47% to 6%	

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Manager

**Laboratory** Autocal Solutions Pvt. Ltd., Plot No. BH – 74, Gala No. 1 & 2, Jai Talja Bhavani Indl. Premises, Telco Road, MIDC – Bhosari, Pune, Maharashtra

**Accreditation Standard** ISO/IEC 17025: 2005

**Discipline** Electro-Technical Calibration **Issue Date** 11.09.2015

**Certificate Number** C-1268 **Valid Until** 10.09.2017

**Last Amended on** - **Page** 5 of 8

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
<b><u>MEASURE</u></b>			
<b>1. DC VOLTAGE<sup>#</sup></b>	100 µV to 1000 V	0.62% to 0.0012%	Using 81/2 Agilent 3458A DMM by Direct Method
	1 kV to 10 kV	5.09% to 2.39%	Using 80k-40 HV Probe with DMM by Direct Method
	10 kV to 90 kV	3.55% to 3.67%	Using HV Divider SEV & kV Meter by Direct Method
<b>2. DC CURRENT<sup>#</sup></b>	10 µA to 10 mA 10 mA to 1 A	0.021% to 0.009% 0.009% to 0.014%	Using 81/2 Agilent 3458A DMM by Direct Method
	1 A to 20 A	0.36% to 0.35%	Using 81/2 Agilent 3458A DMM with Agilent shunt by Direct / Comparison Method
	<b>45 Hz to 1 kHz</b> 1 mV to 100 mV 100 mV to 700 V	0.91% to 0.03% 0.03% to 0.06%	Using 81/2 Agilent 3458A DMM by Direct Method
<b>3. AC VOLTAGE<sup>#</sup></b>	<b>45 Hz to 1 kHz</b> 700 V to 1000 V	0.12% to 0.11%	Using 61/2 Fluke 8846A DMM by Direct Method
	<b>1 kHz to 10 kHz</b> 1 mV to 10 V 10 V to 700 V	0.24% to 0.035% 0.035% to 0.080%	Using 81/2 Agilent 3458A DMM by Direct Method
	<b>50 Hz</b> 1 kV to 10 kV	7.21% to 6.29%	Using 80k-40 HV Probe with DMM by Direct Method

Ram Ashray  
Convenor

Avijit Das  
Program Manager

<b>Laboratory</b>	<b>Autocal Solutions Pvt. Ltd., Plot No. BH – 74, Gala No. 1 &amp; 2, Jai Talja Bhavani Indl. Premises, Telco Road, MIDC – Bhosari, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro-Technical Calibration</b>	<b>Issue Date</b>	<b>11.09.2015</b>
<b>Certificate Number</b>	<b>C-1268</b>	<b>Valid Until</b>	<b>10.09.2017</b>
<b>Last Amended on</b>	<b>-</b>	<b>Page</b>	<b>6 of 8</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
	10 kV to 80 kV	4.27% to 3.64%	Using HV Divider SEV & kV Meter by Direct Method
<b>4. AC CURRENT<sup>#</sup></b>	<b>45 Hz to 1 kHz</b> 30 µA to 100 mA 100 mA to 1A	0.16% to 0.10% 0.10% to 0.14%	Using 81/2 Agilent 3458A DMM by Direct Method
	1 A to 20 A	0.41% to 0.36%	Using 81/2 Agilent 3458A DMM with Agilent shunt By Direct / Comparison Method
<b>5. DC RESISTANCE<sup>#</sup></b>	10 mΩ to 10 Ω 10 Ω to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 1 GΩ	0.82% to 0.0028% 0.0028% to 0.007% 0.007% to 0.065% 0.065% to 0.58%	Using 81/2 Agilent 3458A DMM by Direct Method
<b>6. AC RESISTANCE<sup>#</sup></b>	<b>1kHz</b> 1 Ω to 100 kΩ	0.14% to 0.13%	Using RLC Meter PM6304 By Direct Method
<b>7. FREQUENCY<sup>#</sup></b>	1 Hz to 250 MHz	0.024% to 0.0058%	Using Frequency Counter, Agilent 53220A Direct / Comparison Method
<b>8. PERIOD<sup>#</sup></b>	5 ns to 2 s	0.06% to 0.03%	Using Frequency Counter, Agilent 53220A Direct / Comparison Method
<b>9. CAPACITANCE<sup>#</sup></b>	<b>1kHz</b> 10 pF to 10 µF 10 µF to 100 µF	1.16% to 0.23 % 0.23%	Using RLC Meter PM6304 By Direct Method

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Manager

<b>Laboratory</b>	<b>Autocal Solutions Pvt. Ltd., Plot No. BH – 74, Gala No. 1 &amp; 2, Jai Talja Bhavani Indl. Premises, Telco Road, MIDC – Bhosari, Pune, Maharashtra</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro-Technical Calibration</b>	<b>Issue Date</b>	<b>11.09.2015</b>
<b>Certificate Number</b>	<b>C-1268</b>	<b>Valid Until</b>	<b>10.09.2017</b>
<b>Last Amended on</b>	<b>-</b>	<b>Page</b>	<b>7 of 8</b>

<b>Quantity Measured/ Instrument</b>	<b>Range / Frequency</b>	<b>*Calibration Measurement Capability (±)</b>	<b>Remarks</b>
<b>10. INDUCTANCE<sup>#</sup></b>	<b>1kHz</b> 100 µH to 1 H	1.29% to 0.12%	Using RLC Meter PM6304 by Direct Method
<b>11. TIME<sup>#</sup></b>	6 s to 30 min 30 min to 24 Hrs	0.1s to 1.2 s 1.2 s to 4.0 s	Using Digital Timer by Direct Method
<b>12. PROCESS CALIBRATOR BY SIMULATION METHOD<sup>#</sup></b>			
<b>RTD (PT-100)</b>	-200 °C to 800 °C	0.008 °C to 0.024 °C	Using Multifunction Calibrator, Fluke-5520A
<b>Thermocouple (K, J, N, E, T, R, S, B, C, L, U)</b>	-200 °C to 1370 °C	0.035°C to 0.032°C	Conversion by using ITS-90 Scale for Ω to °C and mV to °C
<b>13. CURRENT TRANSFORMER RATIO ERROR / PHASE ERROR<sup>§</sup></b>	5 A to 3200 A / 5 A 2000 A to 6000 A / 5 A	0.013 % 0.084 %	Using ELTEL Class 0.005 CT and ELTEL AITTS – 98 Bridge
<b>14. CURRENT TRANSFORMER RATIO ERROR/ PHASE ERROR<sup>*</sup></b>	5 A to 2000 A / 5 A 2000 A to 6000 A / 5 A	0.066 % 0.084 %	Using ELTEL Class 0.05 CT and ELTEL AITTS – 98 Bridge
<b>15. POTENTIAL TRANSFORMER RATIO ERROR/ PHASE ERROR<sup>#</sup></b>	6.6 – 11 – 22 – 33 kV/ 110 V	0.10 %	Using Standard PT and ELTEL AITTS – 98 Bridge
<b>16. CT BURDENS<sup>#</sup></b>	1.25 VA to 30 VA	1.75 %	Using ELTEL AITTS – 98 Bridge

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Manager

**Laboratory** Autocal Solutions Pvt. Ltd., Plot No. BH – 74, Gala No. 1 & 2, Jai Talja Bhavani Indl. Premises, Telco Road, MIDC – Bhosari, Pune, Maharashtra

**Accreditation Standard** ISO/IEC 17025: 2005

**Discipline** Electro-Technical Calibration **Issue Date** 11.09.2015

**Certificate Number** C-1268 **Valid Until** 10.09.2017

**Last Amended on** - **Page** 8 of 8

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
17. PT BURDENS <sup>#</sup>	2.5 VA to 50 VA	2.0 %	Using ELTEL AITTS – 98 Bridge

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%

\$Only in Permanent Laboratory

\*Only for Site Calibration

# The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Ram Ashray  
Convenor

Avijit Das  
Program Manager