

Laboratory Beltronics Calibration Laboratory, No. 59, Venkateshwara Layout, 6th Main, Appurao Road, Chamaraajpet, Bengaluru, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2830 (In lieu of C-0271, C-0849, C-1265) **Page** 1 of 10

Validity 28.08.2018 to 27.08.2020 **Last Amended on** 30.08.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>ELECTRO TECHNICAL CALIBRATION</u>				
I.	SOURCE			
1.	DC Voltage [#]	1mV to 1V 1V to 1000V	0.36% to 0.01% 0.01%	Using MFC Fluke 5500 E by Direct Method
2.	AC Voltage [#]	50Hz 1mV to 1 V 1V to 10V 10V to 1000V	2.5% to 0.044% 0.044% to 0.055% 0.055% to 0.068%	Using MFC Fluke 5500 E by Direct Method
3.	DC Current [#]	10 μ A to 100mA 100mA to 1A 1A to 10A 10A to 500A	0.6% to 0.02% 0.02% to 0.04% 0.04% to 0.075% 0.075% to 0.8%	Using MFC Fluke 5500 E with Current Coil by Direct Method
4.	AC Current [#]	50Hz 3.3mA to 100mA 100mA to 1A 1A to 10A 10A to 500A	0.18% to 0.14% 0.14% to 0.18% 0.18% to 0.40% 0.40% to 0.95%	Using MFC Fluke 5500 E with Current Coil by Direct Method
5.	DC Resistance [#]	1 Ω to 10k Ω 10k Ω to 100k Ω 100k Ω to 1M Ω 1M Ω to 10M Ω 10M Ω to 100M Ω 100k Ω to 10M Ω 10M Ω to 10G Ω	0.95% to 0.018% 0.018% to 0.02% 0.02% to 0.025% 0.025% to 0.076% 0.076% to 0.6% 1.15% to 1.16% 1.16% to 2.31%	Using MFC Fluke 5500 E by Direct Method Using HV Decade Mega Ohm Box by Direct Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Beltronics Calibration Laboratory, No. 59, Venkateshwara Layout, 6th Main, Appurao Road, Chamarajapet, Bengaluru, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2830 (In lieu of C-0271, C-0849, C-1265) **Page** 2 of 10

Validity 28.08.2018 to 27.08.2020 **Last Amended on** 30.08.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
6.	Capacitance [#]	1nF to 10nF 10nF to 1 μ F 1 μ F to 10 μ F 10 μ F to 100 μ F 100 μ F to 300 μ F	1.73% to 0.7% 0.7% to 0.41% 0.41% to 0.53% 0.53% to 0.7% 0.7% to 0.95%	Using MFC Fluke 5500 E by Direct Method
7.	Frequency [#]	10Hz to 1MHz 10MHz to 20GHz	0.06% to 0.0074% 0.00025% to 0.00035%	Using MFC Fluke 5500 E & Signal Generator HP 83712A by Direct Method
8.	AC Power [#]	50 Hz (120 V to 240 V & 0.1 A to 10 A) 1.2 W to 2400 W @ UPF 2.4 W to 480 W @ 0.2 Lag 6 W to 1200 W @ 0.5 Lag 9.6 W to 1920 W @ 0.8 Lead	0.18% 1.5% 0.52% 0.22%	Using MPC Fluke 5500 A by Direct Method
9.	Power Factor [#]	0.2 (Lead/Lag) PF to 1 PF	0.004PF	Using MPC Fluke 5500 A by Direct Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Beltronics Calibration Laboratory, No. 59, Venkateshwara Layout, 6th Main, Appurao Road, Chamarajapet, Bengaluru, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2830 (In lieu of C-0271, C-0849, C-1265) **Page** 3 of 10

Validity 28.08.2018 to 27.08.2020 **Last Amended on** 30.08.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
10.	Oscilloscope [#] Amplitude Time Bandwidth	5mV to 33V @DC 5mVp-p to 50Vp-p @AC 1 kHz 5ns to 5s 50 kHz to 200 MHz	2.95% to 0.3% 1.44% to 0.3% 0.012% to 0.58% 8.62%	Using MPC Fluke 5500 A by Direct Method
11.	Temperature Simulation [#] PT 100 K Type J Type S Type R Type T Type N Type	(-)200°C to 600°C (-)200°C to 1300°C (-)200°C to 1200°C 50°C to 1700°C 50°C to 1700°C (-)200°C to 400°C (-)150°C to 1000°C	0.16°C 0.48°C 0.33°C 0.57°C 0.67°C 0.75°C 0.34°C	
II.	MEASURE			
1.	DC Voltage [#]	1mV to 1V 1V to 1000V	0.41% to 0.007% 0.007% to 0.0085%	Using DMM Fluke 8846 A by Direct Method
2.	AC Voltage [#]	1mV to 1000 V	4.65% to 0.10%	Using DMM Fluke 8846 A by Direct Method
3.	DC Current [#]	10 μ A to 100 mA 100mA to 1A 1A to 10A	0.38% to 0.064% 0.064% to 0.082% 0.082% to 0.2%	Using DMM Fluke 8846 A by Direct Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Beltronics Calibration Laboratory, No. 59, Venkateshwara Layout, 6th Main, Appurao Road, Chamaraajpet, Bengaluru, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2830 (In lieu of C-0271, C-0849, C-1265) **Page** 4 of 10

Validity 28.08.2018 to 27.08.2020 **Last Amended on** 30.08.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
4.	AC Current [#]	1mA to 1A 1 A to 10 A	0.18% to 0.18% 0.18 % to 0.25%	Using DMM Fluke 8846 A by Direct Method
5.	Resistance [#]	1m Ω to 10 Ω 10 Ω to 1 M Ω 1 M Ω to 10 M Ω 10M Ω to 100 M Ω 100 M Ω to 1 G Ω	0.085% to 0.05% 0.05% to 0.014% 0.014% to 0.048% 0.048 % to 0.94% 0.94 % to 2.32%	Using D.Micro Ohm Meter & DMM Fluke 8846 A by Direct Method
6.	Frequency ^{\$}	10 Hz to 20 GHz	0.00032 to 0.00025%	Using MicroWave Frequency Counter HP 5350 B by Direct Method
7.	Temperature Simulation [#] K Type J Type S Type R Type	(-)200°C to 1300°C (-)200°C to 1200°C 50°C to 1700°C 50°C to 1700°C	0.47° C 0.32°C 0.55°C 0.66°C	Using MFC Fluke 5500 E by Direct Method
8.	High Voltage [*]	DC Voltage 2kV to 10kV AC Voltage 50Hz 1kV to 10Kv	3.45% 5.83%	Using HV Probe Fluke 80K-40 by Direct Method
9.	Inductance [#]	1 kHz 100 μ H to 10 H	0.54%	Using LCR Meter Aplab 4912 by Direct Method
10.	Time [#]	1 sec to 24 Hrs	0.007 sec to 49.9 sec	Using Beltronics 503 Time Totaliser by Direct Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Beltronics Calibration Laboratory, No. 59, Venkateshwara Layout, 6th Main, Appurao Road, Chamarajapet, Bengaluru, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2830(In lieu of C-0271, C-0849, C-1265) **Page** 5 of 10

Validity 28.08.2018 to 27.08.2020 **Last Amended on** 30.08.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
11.	RF Power ^s	10 MHz to 4 GHz (-)5 dBm to 10 dBm	0.32dBm	Using RF Power meter HP 437B with Power sensor HP 8482A

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Beltronics Calibration Laboratory, No. 59, Venkateshwara Layout, 6th Main, Appurao Road, Chamarajapet, Bengaluru, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2830 (In lieu of C-0271, C-0849, C-1265) **Page** 6 of 10

Validity 28.08.2018 to 27.08.2020 **Last Amended on** 30.08.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>MECHANICAL CALIBRATION</u>				
I.	PRESSURE INDICATING DEVICES			
1.	Pneumatic Digital & Analog Pressure Gauge [#]	0 to 20 bar	0.063 bar	Using Digital Pressure Gauge & Pressure Comparator by Comparison Method
2.	Vacuum Digital & Analog Pressure Gauge [#]	0 to (-)0.85 bar	0.034 bar	Using Digital Pressure Gauge & Pressure Comparator by Comparison Method
3.	Hydraulic Digital & Analog Pressure Gauge [#]	0 to 400 bar	0.82 bar	Using Digital Pressure Gauge & Pressure Comparator by Comparison Method
4.	Hydraulic Digital & Analog Pressure Gauge [#]	0 to 700 bar	1.242 bar	Using Digital Pressure Gauge & Pressure Comparator by Comparison Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Beltronics Calibration Laboratory, No. 59, Venkateshwara Layout, 6th Main, Appurao Road, Chamarajapet, Bengaluru, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2830 (In lieu of C-0271, C-0849, C-1265) **Page** 7 of 10

Validity 28.08.2018 to 27.08.2020 **Last Amended on** 30.08.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
II.	ACCELERATION AND SPEED			
1.	Speed rpm, Tachometer Contact Type [#]	100 rpm to 1000 rpm 1000 rpm to 4500 rpm	2.0 rpm 2.4 rpm	Using Contact Tachometer With Tacho Calibrator (rpm generator) by Direct/ Comparison Method
	Non-Contact Type [#]	100 rpm to 10000 rpm 10000 rpm to 49500 rpm	1.4 rpm 13.33 rpm	Using Contact Tachometer With Tacho Calibrator (rpm generator) by Direct/ Comparison Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Beltronics Calibration Laboratory, No. 59, Venkateshwara Layout, 6th Main, Appurao Road, Chamarajapet, Bengaluru, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2830 (In lieu of C-0271, C-0849, C-1265) **Page** 8 of 10

Validity 28.08.2018 to 27.08.2020 **Last Amended on** 30.08.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>THERMAL CALIBRATION</u>				
I.	TEMPERATURE			
1.	All type of Sensor(RTD/TC) with or without Temperature Indicator Digital Thermometer, Temperature Gauge ^{\$}	(-)40°C to 50°C 50°C to 600°C	0.19°C 0.28°C	Using SSPRT with Indicator, Digital Thermometer (Yudian) & Liquid/Dry baths by Comparison Method
2.	All type of Sensor(RTD/TC) with or without Temperature Indicator Digital Thermometer, Temperature Gauge ^{\$}	600°C to 1200°C	1.62°C	Using "S" Type Thermocouple with Indicator, Digital Thermometer (Yudian) & Dry baths by Comparison Method
3.	Glass Thermometer ^{\$}	(-)10°C to 50°C 50°C to 180°C	0.24°C 0.63°C	Using SSPRT with Indicator, Digital Thermometer (Yudian) & Liquid baths by Comparison Method
4.	Temperature Indicator of Oil Bath, Low/High Temperature bath & Dry well Bath ^{\$}	40°C to 50°C 50°C to 600°C 600°C to 1200°C	0.41°C 1.71°C 2.63°C	Using SSPRT & S Type Thermocouple with Indicator by Comparison Method
5.	Specify Heat & Humidity Hygrometer ^{\$}	10% to 95% @25 \pm 4°C	1.7%	Using Temperature & Humidity with Indicator & RH calibrator by Comparison Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Beltronics Calibration Laboratory, No. 59, Venkateshwara Layout, 6th Main, Appurao Road, Chamaraajpet, Bengaluru, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2830 (In lieu of C-0271, C-0849, C-1265) **Page** 9 of 10

Validity 28.08.2018 to 27.08.2020 **Last Amended on** 30.08.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
6.	IR Thermometer [*]	50°C to 500°C	2.66°C	Using IR Thermometer & Black body furnace by Comparison Method
7.	All type of Sensor(RTD/TC) with or without Temperature Indicator Digital Thermometer, Temperature gauge [*]	25°C to 50°C 50°C to 600°C 600°C to 1200°C	0.37°C 0.41°C 1.69°C	Using SSPRT with Indicator, Digital Thermometer (Yudian) & Liquid/Dry baths by Comparison Method
8.	Temperature Indicator of Oil Bath, oven, Furnace, Autoclave (for all non-medical devices) Low/High Temperature bath & Dry well Bath [*]	(-)40°C to 50°C 50°C to 600°C 600°C to 1200°C	0.62°C 2.04°C 2.86°C	Using SSPRT & S Type Thermocouple with Indicator by Comparison Method
9.	Oven, Furnace, Autoclave (for all non-medical devices) 5 Positions Volume:- Up to 3ft [*]	(-)40°C to 350°C	1.48°C	Using(05nos) RTD sensors with paperless data Recorder by Comparison Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Beltronics Calibration Laboratory, No. 59, Venkateshwara Layout, 6th Main, Appurao Road, Chamarajapet, Bengaluru, Karnataka
Accreditation Standard ISO/IEC 17025: 2005
Certificate Number CC-2830 (In lieu of C-0271, C-0849, C-1265) **Page** 10 of 10
Validity 28.08.2018 to 27.08.2020 **Last Amended on** 30.08.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
II.	SPECIFIC HEAT AND HUMIDITY			
1.	Indicator of Humidity chamber*	10% to 95% @25 \pm 4°C	4.69%	Using Temperature & Humidity with Indicator by Comparison Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

\$ Only in Permanent Laboratory

* Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Dheeraj Chawla
Convenor

Avijit Das
Program Director