

Laboratory	Best Standards Technologies Pvt. Ltd., Door No. 49, Moorthy Nagar, 3rd Street, Chettiar Agaram, Porur, Chennai, Tamil Nadu		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	20.10.2014
Certificate Number	C-0853	Valid Until	19.10.2016
Last Amended on	-	Page	1 of 6

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>MEASURE</u>			
1. DC VOLTAGE[§]	1 mV to 100 mV 100 mV to 1 V 1 V to 1000 V	0.42% to 0.010% 0.010% to 0.006% 0.007%	Using 6½ DMM Agilent 34401A by Direct Method
DC VOLTAGE[#]	1 kV to 2 kV 2 kV to 10 kV 10 kV to 25 kV	2.85% to 2.79% 2.79% to 2.85% 2.85% to 3.42%	Using HV Probe & DMM PD 28 by Direct Method.
DC VOLTAGE[*]	200 V to 900 V	0.29% to 0.090%	Using Standard 6½ DMM by Direct method.
2. AC VOLTAGE[§]	1 kHz to 20 kHz 10 mV to 10 V 10 kHz to 20 kHz 10 V to 750 V	0.54% to 0.12% 0.11%	Using 6½ DMM Agilent 34401A by Direct Method
AC VOLTAGE[#]	50 Hz 1 kV to 25 kV	6.48% to 6.39 %	Using HV Probe & DMM PD 28 by Direct Method.
AC VOLTAGE[*]	50 Hz 40 mV to 20 V 20 V to 700 V	0.93% to 0.76% 0.76% to 0.52%	Using 6½ DMM Agilent 34401A by Direct Method
3. DC CURRENT[§]	10 mA to 100 mA 100 mA to 1 A 1 A to 20 A	0.085% to 0.064% 0.064% to 0.13% 0.13% to 0.35%	Using 6½ DMM Agilent 34401A & current shunt Agilent 34330A by Direct method.

Shally Sharma
Convenor

Avijit Das
Program Manager

Laboratory Best Standards Technologies Pvt. Ltd., Door No. 49, Moorthy Nagar, 3rd Street, Chettiar Agaram, Porur, Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Discipline Electro-Technical Calibration **Issue Date** 20.10.2014

Certificate Number C-0853 **Valid Until** 19.10.2016

Last Amended on - **Page** 2 of 6

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
4. AC CURRENT^{\$}	1 kHz 100 mA to 1 A 1 A to 10 A 10 A to 20 A	0.58% to 0.18% 0.18% to 0.64% 0.64% to 0.47%	Using 6½ DMM Agilent 34401A & current shunt Agilent 34330A by Direct method.
5. RESISTANCE^{\$}	10 Ω to 100 Ω 100 Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω	0.059% to 0.020% 0.020% to 0.013% 0.013% 0.013% 0.013% to 0.047% 0.047% to 0.94%	Using 6½ DMM Agilent 34401A by Direct Method
6. FREQUENCY[#]	3 Hz to 100 Hz 100 Hz to 10 kHz 0.01 MHz to 550 MHz	0.12% to 0.012% 0.012% 0.08% to 0.0016%	By using standard 6½ DMM & Frequency counter Goodwill GFC 8055G by Direct Method
7. TIMER[#]	10 s to 7200 s	4.3% to 0.70%	Using Standard Timer Quantum by Direct method
8. TEMPERATURE SIMULATION[#] (Indicator , Controller And Recorder)			
RTD	-100°C to 800°C	0.30°C	Using MFC UNAMAT TRX By Direct method.
THERMOCOUPLES			
'J' TYPE	-180°C to 750°C	0.51°C	Using MFC Unomat
'T' TYPE	-250°C to 400°C	2.1°C to 0.62°C	TRX
'K' TYPE	-140°C to 1340°C	7.0°C to 0.99°C	By Direct Method.
'R' TYPE	50°C to 1700°C	3.6°C to 2.5°C	
'S' TYPE	100°C to 1700°C	3.7°C to 2.7°C	
'E' TYPE	0°C to 800°C	2.3°C to 0.42°C	
'N' TYPE	-200°C to 1300°C	0.77°C	
'B' TYPE	600°C to 1810°C	3.1°C	

Shally Sharma
Convenor

Avijit Das
Program Manager

Laboratory	Best Standards Technologies Pvt. Ltd., Door No. 49, Moorthy Nagar, 3rd Street, Chettiar Agaram, Porur, Chennai, Tamil Nadu		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	20.10.2014
Certificate Number	C-0853	Valid Until	19.10.2016
Last Amended on	-	Page	3 of 6

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
9. DC CURRENT*	200 mA to 1 A 1 A to 10 A 10 A to 20 A	0.090% to 0.13% 0.13% to 0.35% 0.35%	Using Standard 6½ DMM with current shunt by Direct method.
10.. AC CURRENT *	50 Hz 200 mA to 1000 mA 1000 mA to 10 A 10 A to 20 A	0.70% to 1.4% 1.4% to 2.7% 2.7% to 1.6%	Using Standard 6½ DMM with current shunt by Direct method.
11. RESISTANCE*	20 Ω to 200 Ω 200 Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 10 M Ω	0.060% 0.060% to 0.030% 0.030% 0.030% to 0.27%	Using Standard 6½ DMM by Direct method.
<u>SOURCE</u>			
1. DC VOLTAGE#	10 mV to 300 mV 300 mV to 1 V 1 V to 1000 V	0.060% to 0.0090% 0.0090% to 0.014% 0.014% to 0.012%	Using Multifunction calibrator Wavetek 9000 by Direct method.
2. AC VOLTAGE#	10 Hz to 3 kHz 10 mV to 100 mV 100 mV to 300 mv 50 Hz to 3 kHz 300 mV to 100 V 100 V to 300 V 50 Hz to 1 kHz 300 V to 500 V 500 V to 750 V	4.5% to 0.11% 0.11% to 0.055% 0.054% to 0.080% 0.080% to 0.082% 0.082% to 0.072% 0.072% to 0.068%	Using Multifunction calibrator Wavetek 9000 by Direct method.

Laboratory	Best Standards Technologies Pvt. Ltd., Door No. 49, Moorthy Nagar, 3rd Street, Chettiar Agaram, Porur, Chennai, Tamil Nadu		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	20.10.2014
Certificate Number	C-0853	Valid Until	19.10.2016
Last Amended on	-	Page	4 of 6

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks	
AC VOLTAGE[§]	10 Hz to 30 kHz 100 mV to 3 V 3 V to 100 V	0.13% to 0.090% 0.090% to 0.12%	Using Multifunction calibrator Wavetek 9000 by Direct method.	
	50 kHz to 3 kHz 100 mV to 3 V 3 V to 100 V	0.54% to 0.33% 0.33% to 0.53%		
3. DC CURRENT[‡]	100 μ A to 300 μ A	0.042% to 0.030%	Using Multifunction calibrator Wavetek 9000 by Direct method.	
	300 μ A to 3 mA	0.030% to 0.027%		
	3 mA to 30 mA	0.029%		
	30 mA to 300 mA	0.028%		
	300 mA to 3 A	0.028% to 0.081%		
	3 A to 10 A	0.083%		
10 A to 20 A	0.083% to 0.098%			
4. AC CURRENT[‡]	10 A to 100 A	1.4% to 1.1%	Using Standard MFC Wavetek 9000 with Current coil Zeal by Direct Method	
	100 A to 500 A	1.1% to 0.91%		
	500 A to 1000 A	0.91% to 0.81%		
	10 Hz to 3 kHz 100 μ A to 300 μ A	0.43% to 0.20%		Using Multifunction calibrator Wavetek 9000 by Direct method.
	300 μ A to 3 mA	0.20% to 0.090%		
	3 mA to 300 mA	0.090% to 0.11%		
300 mA to 3A	0.11% to 0.14%			
3 A to 20 A	0.14% to 0.28%			
3 kHz & 10 kHz 30 mA to 300 mA	0.14%			
300 mA to 3 A	0.14% to 0.39%			
3 A to 10 A	0.39% to 0.69%			
50 Hz 10 A to 100 A	1.9% to 1.3%	Using Standard MFC with 100 TURNS Current coil by Direct Method		
100 A to 500 A	1.3% to 1.2%			

Shally Sharma
Convenor

Avijit Das
Program Manager

Laboratory	Best Standards Technologies Pvt. Ltd., Door No. 49, Moorthy Nagar, 3rd Street, Chettiar Agaram, Porur, Chennai, Tamil Nadu		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	20.10.2014
Certificate Number	C-0853	Valid Until	19.10.2016
Last Amended on	-	Page	5 of 6

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
	500 A to 1000 A	1.2% to 1.1%	
5. RESISTANCE[#]	1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω	1.2% to 0.14% 0.14% to 0.050% 0.050% to 0.030% 0.030% to 0.038% 0.038% to 0.098% 0.098% to 0.20% 0.20% to 0.12%	Using Multifunction calibrator Wavetek 9000 by Direct method.
6. CAPACITANCE^{\$}	1 kHz 100 pF to 1nF 1 nF to 10 nF 10 nF to 100 nF 1 μ F to 10 μ F	3.5% to 6.7% 6.7% to 3.5% 3.5% to 3.5% 7.4% to 3.5%	Using Standard Capacitance Box by Direct Method
7. INDUCTANCE^{\$}	1 kHz 100 μ H to 1 mH 1 mH to 10 mH 10 mH to 100 mH 100 mH to 1 H	3.5% to 6.9% 6.9% to 3.7% 3.7% to 3.7% 3.7% to 8.2%	Using Standard Inductance Box by Direct Method
8. FREQUENCY[#]	1 Hz to 10 MHz	0.0030 % to 0.00030 %	Using Multifunction calibrator Wavetek 9000 by Direct method.
9. TEMPERATURE SIMULATION[#] (indicator, Controller and Recorder) RTD	-100°C to 800°C	0.30°C	Using MFC UNMAT TRX(s.no:04823) By Direct method.
THERMOCOUPLES			
'J' TYPE	-180°C to 750°C	0.62°C	
'T' TYPE	-250°C to 400°C	2.1°C to 0.62°C	

Shally Sharma
Convenor

Avijit Das
Program Manager

Laboratory Best Standards Technologies Pvt. Ltd., Door No. 49, Moorthy Nagar, 3rd Street, Chettiar Agaram, Porur, Chennai, Tamil Nadu
Accreditation Standard ISO/IEC 17025: 2005
Discipline Electro-Technical Calibration **Issue Date** 20.10.2014
Certificate Number C-0853 **Valid Until** 19.10.2016
Last Amended on - **Page** 6 of 6

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
'K' TYPE	-140°C to 1340°C	7.03°C to 0.99°C	
'R' TYPE	50°C to 1700°C	3.61°C to 2.52°C	
'S' TYPE	100°C to 1700°C	3.7°C to 2.7°C	
'E' TYPE	0°C to 800°C	2.3°C to 0.42°C	
'N' TYPE	-200°C to 1300°C	0.77°C	
'B' TYPE	600°C to 1810°C	3.1°C	

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

\$ Only in Permanent Laboratory

* Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Shally Sharma
 Convenor

Avijit Das
 Program Manager