

Laboratory	Bharti Automation Private Limited, 48-49, Vasundhra Nagar, Bhiwadi, Distt. Alwar, Rajasthan		
Accreditation Standard	ISO/IEC 17025:2005		
Discipline	Electro-Technical Calibration	Issue Date	24.04.2015
Certificate Number	C-0680	Valid Until	23.04.2017
Last Amended on	-	Page	1 of 3

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
<u>MEASURE</u>			
1. DC VOLTAGE^{\$}	10 mV to 100 mV	0.9 to 0.05	Using 6 .5 DMM, FLUKE 8845
	100 mV to 10 V	0.05 to 0.06	
	10 V to 100 V	0.06 to 0.05	
	100 V to 1000 V	0.05 to 0.04	
	>1 kV to 10 kV	0.2 kV to 0.6 kV	
2. AC VOLTAGE^{\$}	50 Hz		Using 6 .5 DMM, FLUKE 8845
	100 mV to 100 V	0.15 to 0.12	
	100 V to 750 V	0.12	
3. AC HIGH VOLTAGE^{\$}	50 Hz		Using HV Probe with DMM
	>1kV to 20 KV	0.2 kV to 1.3 kV	
4. DC CURRENT^{\$}	2 mA to 20 mA	0.2 to 0.1	Using 6 .5 DMM, FLUKE 8845 A
	20 mA to 100 mA	0.1 to 0.12	
	100 mA to 1 A	0.12 to 0.2	
	1 A to 10 A	0.2 to 0.22	
5. AC CURRENT^{\$}	1 mA to 100 mA	1 to 0.6	Using 6 .5 DMM, FLUKE 8845 A
	100 mA to 1 A	0.6 to 0.4	
	1 A to 10 A	0.4 to 0.29	
6. FREQUENCY^{\$}	50 Hz to 300 kHz	0.12 to 0.3	Using 6 .5 DMM, FLUKE 8845 A
7. STOP WATCH^{\$}	20 s 10 15 min.	0.7 s	Using Digital Stop Watch
8. INDUCTANCE^{\$}	1 mH to 100 mH	3.1 to 2.2	Using LCR Meter Agilent U1731
	100 mH to 1 H	2.2 to 1.7	

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory **Bharti Automation Private Limited, 48-49, Vasundhra Nagar, Bhiwadi, Distt. Alwar, Rajasthan**

Accreditation Standard **ISO/IEC 17025:2005**

Discipline **Electro-Technical Calibration** **Issue Date** **24.04.2015**

Certificate Number **C-0680** **Valid Until** **23.04.2017**

Last Amended on **-** **Page** **2 of 3**

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (±)	Remarks
9. CAPACITANCE [§]	100 nF to 3.3 µF	1.1 to 1.4	Using LCR Meter Agilent U1731
10. RESISTANCE [§] (2 WIRE)	10 Ω to 100 Ω 100 Ω to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1MΩ to 100 MΩ	0.02 to 0.09 0.013 to 0.02 0.013 to 0.02 0.02 0.02 to 1	Using 6 .5 DMM, FLUKE 8845 A
11. AC POWER /ENERGY [#] 1 φ & 3 φ ACTIVE COSØ=0.50 LAG to 1	50Hz 240Wh to 2.4 kWh 110 V to 415 V 1A to 5A	0.13 %	Using Standard Energy meter 0.1 Class Yokogawa WT 330
<u>SOURCE</u>			
1. AC CURRENT [#]	1 A to 20 A 20 A to 2000 A	0.4 % to 0.3 % 0.3 % to 2.3 %	Precision High Current Source Model: bap1/hcs3
2. TEMPERATURE SIMULATION* Temperature Indicator / Temperature Controller / PID / Temperature Recorder & Thermocouple Indicator			
RTD (pt-100)	-200 °C to 600 °C	1.1 °C	
Thermocouple R Type	200 °C to 1700 °C	1.5 °C	Using Universal Calibrator of masibus
S Type	200 °C to 700 °C	2.1 °C	(Unicab3001M)
K Type	50 °C to 1300 °C	1.1 °C	

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory	Bharti Automation Private Limited, 48-49, Vasundhra Nagar, Bhiwadi, Distt. Alwar, Rajasthan		
Accreditation Standard	ISO/IEC 17025:2005		
Discipline	Electro-Technical Calibration	Issue Date	24.04.2015
Certificate Number	C-0680	Valid Until	23.04.2017
Last Amended on	-	Page	3 of 3

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (±)	Remarks
J Type	50 °C to 750 °C	1.1 °C	
T Type	-150 °C to 400 °C	1.1 °C	

* Measurement Capability is expressed as an uncertainty (±) at a confidence probability of 95%

§ Only in Permanent Laboratory

* Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Vishal Shukla
Convenor

Avijit Das
Program Manager