

**Laboratory**                      **Birla Tyres Calibration Laboratory, AT/PO- Chhanpur Via- Kuruda, Balasore, Orissa**  
**Accreditation Standard**   **ISO/IEC 17025:2005**  
**Discipline**                      **Mechanical Calibration**                      **Issue Date**      **21.07.2015**  
**Certificate Number**        **C-0103**                                              **Valid Until**      **31.05.2017**  
**Last Amended on**        **-**                                              **Page**              **1 of 1**

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability ( $\pm$ )	Remarks
<b>I. PRESSURE &amp; VACCUM</b>			
1. <b>PRESSURE GAUGE/ DIGITAL PRESSURE GAUGE<sup>\$</sup></b>	Upto 20 bar	0.024 bar	Using Druck DPI 610 Low Pressure Calibrator by Comparison Method
	Upto 400 bar	1.2 bar	Using Druck DPI 610 High Pressure Calibrator by Comparison Method

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%  
<sup>\$</sup>Only in Permanent Laboratory

---

**Naveen Jangra**  
**Convenor**

---

**Avijit Das**  
**Program Manager**