

Laboratory **Birla Tyres Calibration Laboratory, AT/PO- Chhanpur Via- Kuruda, Balasore, Orissa**
Accreditation Standard **ISO/IEC 17025:2005**
Discipline **Thermal Calibration** **Issue Date** **01.06.2015**
Certificate Number **C-0104** **Valid Until** **31.05.2017**
Last Amended on **-** **Page** **1 of 1**

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
I. TEMPERATURE			
1. RTD, GLASS TEMPERATURE INDICATOR WITH PROBE, TEMPERATURE DIAL GAUGE ^S	0 °C to 150 °C	0.2 °C	Using Standard RTD, Druck, MFC & Julabo Bath by Comparison Method
2. RTD TEMPERATURE INDICATOR WITH PROBE ^S	150 °C to 400 °C	0.38 °C	Using Standard RTD, Druck, MFC, Ametek Dry Block Temperature Calibrator by Comparison Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%
^SOnly in Permanent Laboratory

Naveen Jangra
Convenor

Avijit Das
Program Manager