

Laboratory **Cali-Labs Pvt. Ltd., HX-21, E-7, Arera Colony, Bhopal, Madhya Pradesh**
Accreditation Standard **ISO/IEC 17025:2005**
Discipline **Electro-Technical Calibration** **Issue Date** **30.06.2015**
Certificate Number **C-0106** **Valid Until** **29.06.2017**
Last Amended on - **Page** **1 of 3**

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
<u>SOURCE</u>			
1. DC VOLTAGE[#]	1 mV to 10 mV 10 mV to 1V 1 V to 1000 V	1.0 % to 0.1 % 0.1 % to 0.015 % 0.015 % to 0.01 %	Using Multifunction calibrator 5500 A Fluke by Direct Method
2. DC CURRENT[#]	10 µA to 100 µA 100 µA to 100 mA 100 mA to 10A	0.6 % to 0.07 % 0.07 % to 0.02 % 0.02 % to 0.09 %	Using Multifunction calibrator 5500 A Fluke by Direct Method
3. RESISTANCE[#]	1 mΩ to 10 Ω	1.35 % to 2.4 %	Using Shunts & Decade Box Maxwell, Model –DR-8 by Direct Method
	10 Ω to 100 kΩ 100 kΩ to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 300 MΩ	0.2 % to 0.02 % 0.02 % to 0.08 % 0.08 % to 0.6 % 0.6 % to 1.0 %	Using Multifunction calibrator 5500 A Fluke by Direct Method
	300 MΩ to 100 GΩ	3.2 %	Using Meg Ohm Box Maxwell, Model DB 231 for calibration of Meggers upto 5kV by Direct Method
4. AC VOLTAGE[#]	50 Hz to 1 kHz 10 mV to 10 V 10 V to 1000V	0.45 % to 0.06 % 0.06 % to 0.1 %	Using Multifunction calibrator 5500 A Fluke by Direct Method
5. AC CURRENT[#]	50 Hz to 1 kHz 100 µA to 1 A 1 A to 10 A	0.45 % to 0.16 % 0.16 % to 0.15 %	Using Multifunction calibrator 5500 A Fluke by Direct Method

Naveen Jangra
Convenor

Avijit Das
Program Manager

Laboratory	Cali-Labs Pvt. Ltd., HX-21, E-7, Arera Colony, Bhopal, Madhya Pradesh		
Accreditation Standard	ISO/IEC 17025:2005		
Discipline	Electro-Technical Calibration	Issue Date	30.06.2015
Certificate Number	C-0106	Valid Until	29.06.2017
Last Amended on	-	Page	2 of 3

	Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
6.	AC POWER[#]	50 Hz 0.5 W to 6 kW @ UPF to 0.2 PF	0.3 % to 2.0 %	Using Multifunction calibrator 5500 A Fluke by Direct Method
7.	FREQUENCY[#]	50 Hz to 100 kHz	0.015 % to 0.065 %	Using Multifunction calibrator 5500 A Fluke by Direct Method
8.	CAPACITANCE[#]	1 kHz 0.5 nF to 100 µF 100 µF to 300 µF	3.0 % to 0.7 % 0.7 % to 1.2 %	Multifunction calibrator 5500 A Fluke by Direct Method
9.	TEMPERATURE SIMULATION[#]			
	Thermocouples			
	S Type	50 °C to 1750 °C	0.75 °C	Using Multifunction Calibrator
	R Type	50 °C to 1750 °C	0.75 °C	5500 A Fluke
	K Type	(-) 200 °C to 1350 °C	0.75 °C	And for Site Radix Calibrator
	J Type	(-) 200 °C to 750 °C	0.75 °C	by Simulation Method
	B Type	50 °C to 1750 °C	0.75 °C	
	N Type	50 °C to 1750 °C	0.75 °C	
	T Type	(-) 200 °C to 400 °C	0.75 °C	
	E Type	(-) 200 °C to 1000 °C	0.75 °C	
	RTD	(-) 200 °C to 800 °C	0.35 °C	
	<u>MEASURE</u>			
1.	DC VOLTAGE^{\$}	1 mV to 200 mV 200 mV to 1000 V	1.00 % to 0.045 % 0.045 % to 0.04 %	Using 6½ DMM Fluke 8846A / Fluke 5500 by Direct/Comparison Method
2.	DC CURRENT^{\$}	10 µA to 10 A	0.7 % to 0.2 %	Using 6½ DMM Fluke 8846A by Direct Method
3.	RESISTANCE^{\$}	10 Ω to 200 kΩ 200 kΩ to 180 MΩ 180 MΩ to 1 GΩ	0.7 % to 0.2 % 0.2 % to 2.4 % 2.4 %	Using 6½ DMM Fluke 8846A by Direct Method

Naveen Jangra
Convenor

Avijit Das
Program Manager

Laboratory	Cali-Labs Pvt. Ltd., HX-21, E-7, Arera Colony, Bhopal, Madhya Pradesh		
Accreditation Standard	ISO/IEC 17025:2005		
Discipline	Electro-Technical Calibration	Issue Date	30.06.2015
Certificate Number	C-0106	Valid Until	29.06.2017
Last Amended on	-	Page	3 of 3

	Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
4.	AC VOLTAGE^{\$}	50 Hz 10 mV to 200 mV 0.2 V to 1000 V	1.5 % to 0.5 % 0.5 % to 0.11 %	Using 6 ½ DMM Fluke 8846A by Direct Method
5.	AC CURRENT^{\$}	50 Hz 100 μ A to 100 mA 100 mA to 10 A	2.0 % to 0.6 % 0.6 % to 0.7 %	Using 6 ½ DMM Fluke 8846A by Direct Method
6.	LOW RESISTANCE[#]	1 m Ω to 10 Ω	0.6 %	Using Micro Ohm Meter to Motwane, Model to L2065 by Direct Method
7.	FREQUENCY^{\$}	50 Hz to 1MHz	0.8 % to 0.1 %	Using 6½ DMM Fluke 8846A by Direct Method
8.	AC HIGH VOLTAGE^{\$}	50 Hz 1 kV to 5 kV AC	3.5 % to 2.4 %	Using Motwane HV Probe with DMM motwane by Comparison Method
	AC HIGH VOLTAGE^{*#}	50 Hz 1 kV to 20 kV AC	3.5 %	Using Motwane HV Probe with DMM motwane by Comparison Method
9.	DC HIGH VOLTAGE[#]	1 kV to 5 kV	2.0 %	Using Motwane HV Probe with DMM motwane by Comparison Method
10.	DIGITAL STOP WATCH[#]	10 s to 59 minute	1.5 s to 2 s	Stop Watch by Comparison Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

^{\$} Only in Permanent Laboratory

^{*#} Only for Site Calibration

[#] The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Naveen Jangra
Convenor

Avijit Das
Program Manager