Laboratory	Calibration Lab, Deccan Charters Pvt. Ltd., Jakkur Aerodrum, Bellary Road, Bengaluru, Karnataka		
Accreditation Standard	ISO/IEC 17025: 2005		
Certificate Number	CC-2468	Page 1 of 1	
Validity	06.12.2017 to 05.12.2019	Last Amended on -	

SI.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks		
MECHANICAL CALIBRATION						
I.	PRESSURE INDICATING DEVICES					
1.	Hydraulic - Dial and Digital Pressure Gauges ^{\$}	0.689 bar to 34.450 bar 13.789 bar to 689.475 bar	0.030% rdg 0.030 % rdg	Using GE Pressurements Deadweight Tester by Direct Method as per DKD R-6-1		

 * Measurement Capability is expressed as an uncertainty (±) at a confidence probability of 95% $^{\rm \$}$ Only in Permanent Laboratory