

Laboratory Calibration & Standards Laboratory, Controllerate of Quality Assurance (Electronics), JC Nagar Post, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration Issue Date 30.07.2014

Certificate Number C-0838 Valid Until 29.07.2016

Last Amended on - Page 1 of 5

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
----------------------------------	-------------------	--	---------

SOURCE^s

1. DC Voltage	1mV to 100mV	0.047% to 0.0014%	Using Fluke5720A By Direct Method	
	100 mV to 1 V	0.0014% to 0.0007%		
	1 V to 10 V	0.0007% to 0.0005%		
	10 V to 1000 V	0.0005% to 0.0008%		
2. DC Current	10 μ A to 100 μ A	0.051% to 0.012%	Using Fluke5720A By Direct Method	
	100 μ A to 1 mA	0.012% to 0.005%		
	1 mA to 10 mA	0.005% to 0.0045%		
	10 mA to 100 mA	0.0045% to 0.006%	Using Multi Function calibrator, Fluke 5720A , Trans conductance Amplifier, DMM / Keithley 2002, Std Resistor L & N 0.0001 Ω by Indirect Method	
	100 mA to 1 A	0.006% to 0.011%		
	1A to 2A	0.011% to 0.010%		
	2A to 100A	0.010% to 0.047%		
3. AC Voltage	20 Hz to 40 Hz	0.058% to 0.031%	Using Fluke5720A by Direct Method	
	20 mV to 20 V			
	40 Hz to 1 kHz	0.61% to 0.036%		
	2 mV to 20 mV			
	20 mV to 200 mV			0.036% to 0.017%
	200 mV to 2 V			0.017% to 0.013%
2 V to 200 V	0.01%			

Avijit Das
Program Manager

Neeraj Verma
Convenor

Laboratory Calibration & Standards Laboratory, Controllerate of Quality Assurance (Electronics), JC Nagar Post, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration Issue Date 30.07.2014

Certificate Number C-0838 Valid Until 29.07.2016

Last Amended on - Page 2 of 5

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks	
AC Voltage	50 Hz to 1 kHz 200 V to 750 V	0.011% to 0.038%	Using Fluke5720A By Direct Method	
	1 kHz to 50 kHz 2 mV to 20 mV	0.63% to 0.10%		
	50 kHz to 300 kHz 200 mV to 20 V	0.12% to 0.042%		
	300 kHz to 1 MHz 2 V to 20 V	0.26% to 0.25%		
4. AC Current	40 Hz to 1 kHz 10 μ A to 200 μ A 200 μ A to 100 mA 100 mA to 1 A 1 A to 2 A	0.14% to 0.025% 0.025% 0.025% to 0.04% 0.04% to 0.035%	Using Fluke5720A By Direct Method	
	1 kHz 2 A to 20 A	0.35% to 0.12%		Using Multi Function calibrator, Fluke 5720A , Trans conductance Amplifier by Indirect Method

Avijit Das
Program Manager

Neeraj Verma
Convenor

Laboratory Calibration & Standards Laboratory, Controllerate of Quality Assurance (Electronics), JC Nagar Post, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration Issue Date 30.07.2014

Certificate Number C-0838 Valid Until 29.07.2016

Last Amended on - Page 3 of 5

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
5. Resistance	1 Ω to 1.9 Ω 1.9 Ω to 10 Ω 10 Ω to 19 k Ω 19 k Ω to 1 M Ω	0.012% 0.012% to 0.0027% 0.0027% to 0.0011% 0.0011% to 0.0046%	Using DMM/ Keithley 2002, Std. Resistor L & N Tinsley 0.01 Ω by Indirect Method
6. Oscilloscope			
Vertical channel			
a) DC Voltage	3mV to 50V (1M Ω) 3mV to 5V (50 Ω)	1.20% 1.20%	Using Oscilloscope Calibrator Fluke 9500 by Direct Method
	1 kHz		
b) Square wave	6mV to 60V (1M Ω) 6mV to 5V (50 Ω)	1.20% 1.20%	
Time Base	10 ns to 1 s	46ppm to 126ppm	
Band Width	10MHz to 1GHz	4% to 5%	
MEASURE			
7. DC Voltage ^s	190 mV to 1.9 V 1.9 V to 19 V 19 V to 1000 V	0.0039% to 0.0015% 0.0015% 0.0015% to 0.0028%	Using DMM Keithley 2002 by Direct Method
8. DC Current ^s	190 μ A to 1.9 mA 1.9 mA to 190 mA 190 mA to 1.9 A	0.045% to 0.043% 0.045% to 0.069% 0.069% to 0.11%	Using DMM Keithley 2002 by Direct Method

Avijit Das
Program Manager

Neeraj Verma
Convenor

Laboratory Calibration & Standards Laboratory, Controllerate of Quality Assurance (Electronics), JC Nagar Post, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 30.07.2014

Certificate Number C-0838 **Valid Until** 29.07.2016

Last Amended on - **Page** 4 of 5

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
DC Current ^{\$}	2 A to 100 A	0.017% to 0.031%	DMM / Keithley 2002, Std Resistor L & N 0.01 Ω , 0.001 Ω , 0.0001 Ω by Indirect Method
9. AC Voltage ^{\$}	50 Hz to 1 kHz 190 mV to 19 V 19 V to 190 V 190 V to 750 V	0.05% to 0.10% 0.10% 0.10% to 0.14%	Using DMM, Keithley 2002 by Direct Method
	1 kHz to 50 kHz 190 mV to 19 V	0.08% to 0.10%	
	5kHz to 50 kHz 19 V to 100 V	0.010% to 0.12%	
10. AC Current ^{\$}	40 Hz to 100 Hz 190 μ A to 1.9 mA 1.9 mA to 190 mA 190 mA to 1.9 A	0.43% to 0.37% 0.37% 0.37% to 0.43%	Using DMM, Keithley 2002 & DMM Fluke 289 by Direct Method
	100 Hz to 1 kHz 1.9 mA to 190 mA	0.20%	

Avijit Das
Program Manager

Neeraj Verma
Convenor

Laboratory Calibration & Standards Laboratory, Controllerate of Quality Assurance (Electronics), JC Nagar Post, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration Issue Date 30.07.2014

Certificate Number C-0838 Valid Until 29.07.2016

Last Amended on - Page 5 of 5

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
AC Current ^{\$}	50 Hz 2 A to 10 A 10 A to 20 A	1.05% to 0.99% 0.12% to 0.098%	Using DMM, Keithley 2002 by Indirect Method
11. Resistance ^{\$}	1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1M Ω	0.041% to 0.007% 0.007% to 0.0051% 0.0051% to 0.0019% 0.0019% to 0.002% 0.002% to 0.0051% 0.0051% to 0.010%	Using DMM, Keithley 2002 by Direct Method
12. DC Current *	1 A to 100 A	0.15% to 0.035%	Using DMM Fluke 289, Std Resistor L & N 0.01 Ω , 0.001 Ω & 0.0001 Ω ,

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

^{\$}Only in Permanent Laboratory

*Only for Site Calibration

Avijit Das
Program Manager

Neeraj Verma
Convenor