

Laboratory Calsytech, No. 38, North Mada Street, Nandambakkam, Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2193

Page 1 of 7

Validity 10.10.2018 to 09.10.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>ELECTRO-TECHNICAL CALIBRATION</u>				
I.	SOURCE			
1.	DC Voltage [§]	1 mV to 10 mV 10 mV to 100 mV 100 mV to 1000 V	0.43 % to 0.05 % 0.05 % to 0.01 % 0.01 % to 0.005 %	Using Multi Product Calibrator by Direct Method
2.	AC Voltage [§]	40 Hz to 1 kHz 20 mV to 1 V 1 V to 1000 V	0.7 % to 0.1 % 0.1 % to 0.09 %	Using Multi Product Calibrator by Direct Method
3.	AC Voltage [§]	1 kHz to 10 kHz 100 mV to 10 V 10 V to 700 V	2 % to 0.2 % 0.2 % to 0.34 %	Using Multi Product Calibrator by Direct Method
4.	DC Current [§]	1 μ A to 10 μ A 10 μ A to 100 μ A 100 μ A to 1 A 1 A to 20 A 20 A to 1000 A	2.8 % to 0.24 % 0.24 % to 0.34 % 0.34 % to 0.02 % 0.02 % to 0.1 % 0.8 % to 0.36 %	Using Multi Product Calibrator by Direct Method Using Multi Product Calibrator & Current Coil by Direct Method
5.	AC Current [§]	45 Hz to 1 kHz 20 μ A to 100 μ A 50 Hz to 60 Hz 1 A to 20 A 500 Hz 2 A to 20 A	4 % to 0.88 % 0.88 % to 0.19 % 0.14 %	Using Multi Product Calibrator A Direct Method

Sangeeta Kunwar
Convenor

Avijit Das
Program Manager

Laboratory Calsytech, No. 38, North Mada Street, Nandambakkam, Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2193

Page 2 of 7

Validity 10.10.2018 to 09.10.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
		50 Hz to 60 Hz 20 A to 1000 A	0.77 % to 0.38 %	Using Multi Product Calibrator & Current Coil by Direct Method
6.	Frequency [§]	45 Hz to 10 MHz 10 MHz to 600 MHz	0.004 % 0.004 %	Using Multi Product Calibrator by Direct Method
7.	Capacitance [§]	1 kHz 1 nF to 100 μ F	0.4 % to 1 %	Using Multi Product Calibrator by Direct Method
8.	AC Resistance [#]	1 kHz 100 Ω , 1 k Ω 10 k Ω	0.07 %	Using Standard Resistor by Direct Method
9.	DC Resistance [§] 4 Wire/2 Wire	100 $\mu\Omega$ 1 m Ω 10 m Ω 100 m Ω to 1 Ω 1 Ω to 1 k Ω 1 k Ω to 1 M Ω 1 M Ω to 1 G Ω 1 G Ω to 100 G Ω	0.46 % 0.46 % 0.46 % 0.15 % to 0.13 % 0.13 % to 0.005 % 0.005 % to 0.012 % 0.012 % to 2.4 % 2.4 % to 6.2 %	Using Standard Resistor by Direct Method Multi Product Calibrator by Direct Method Using Multi Product Calibrator 2041 & HV Resistance Box by Direct Method
10.	Oscilloscope [§] Amplitude Time Base Bandwidth	2 mV/Div to 20 V/Div (12 mV to 120 V) 2 ns to 5 sec 10 MHz to 600 MHz Ref: 50 kHz	0.1 % to 0.013 % 0.03 % to 0.012 % 1.3 dB	Using Multi Product. Calibrator by Direct Method

Laboratory Calsytech, No. 38, North Mada Street, Nandambakkam, Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2193

Page 3 of 7

Validity 10.10.2018 to 09.10.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
11.	DC Power #	10 V to 600 V 1 A to 20 A 10 W to 12 kW	0.038 % to 0.09 %	Using Multi Product Calibrator by Direct Method
12.	Phase Angle #	0° to 330°	0.3°	Using Multi Product Calibrator by Direct Method
13.	Inductance#	1 kHz 1 mH to 10 H	0.61 %	Using Inductance Box by Direct Method
	Temperature Simulation #			
	RTD	(-) 100 °C to 800 °C	0.15 °C	Using Multifunction Calibrator by Direct Method
	Thermocouple "K & J" Type	(-) 140 °C to 1200 °C	0.54 °C	
	Thermocouple "B" Type	600 °C to 1800 °C	1.5 °C	
	Thermocouple "E" Type	0 °C to 800 °C	0.36 °C	
	Thermocouple "N" Type	(-) 180 °C to 1300 °C	0.25 °C to 0.5 °C	
	Thermocouple "R & S" Type	0 °C to 1700 °C	0.7 °C to 1.17 °C	
	Thermocouple "T" Type	0 °C to 400 °C	0.26 °C	
15.	Energy # 1 ϕ and 3 ϕ (0.5PF to UPF)	50 Hz 63.5 V to 300 V 1 A to 15 A	0.27 %	Using Energy Cal Source by Direct Method
16.	AC Power @ 50 Hz ^s 1 ϕ & 3 ϕ (0.5 PF to UPF)	110 V to 300 V 1 A to 20 A 110 W to 6000 W 300 V to 600 V 20 A	0.24 %	Using Energy Cal Source by Direct Method

Sangeeta Kunwar
Convenor

Avijit Das
Program Manager

Laboratory Calsytech, No. 38, North Mada Street, Nandambakkam, Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2193

Page 4 of 7

Validity 10.10.2018 to 09.10.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
		6000 W to 12000 W 110 V to 300 V 1 A to 20 A 110 W to 6000 W	0.24 % 0.24 %	
17.	Power Factor #	0.5 PF to 1 PF	0.1 %	Using Energy Cal Source by Direct Method
18.	DC Voltage*	1 mV to 20 mV 20 mV to 200 mV 200 mV to 1000 V	0.47 % to 0.04 % 0.04 % to 0.008 % 0.008 %	Using Multi Product Calibrator by Direct Method
19.	AC Voltage *	50 Hz to 1 kHz 20 mV to 200 V 200 V to 1000 V	0.26 % to 0.07 % 0.07 %	Using Multi Product Calibrator by Direct Method
20.	DC Current *	1 μ A to 100 μ A 100 μ A to 200 mA 200 mA to 20 A 20 A to 1000 A	2.32 % to 0.06 % 0.06 % to 0.02 % 0.02 % to 0.05 % 0.77 % to 0.36 %	Using Multi Product Calibrator by Direct Method
21.	AC Current *	50 Hz to 1 kHz 25 μ A to 20 mA 20 mA to 2 A 2 A to 20 A 50 Hz 20 A to 1000 A	1.4 % to 0.5 % 0.5 % to 0.1 % 0.1 % 0.1 % to 0.3 %	Using Multi Product Calibrator by Direct Method
22.	DC Resistance*	100 μ Ω 1 m Ω 10 m Ω 100 m Ω 1 Ω 10 Ω to 100 k Ω 100 k Ω to 10 M Ω	0.46 % 0.46 % 0.46 % 0.5 % 0.15 % 0.12 % to 0.06 % 0.06 % to 0.07 %	Using Multi Product Calibrator by Direct Method

Sangeeta Kunwar
Convenor

Avijit Das
Program Manager

Laboratory Calsytech, No. 38, North Mada Street, Nandambakkam, Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2193

Page 5 of 7

Validity 10.10.2018 to 09.10.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
		10 M Ω to 100 M Ω 100 M Ω to 1 G Ω 1 G Ω to 100 G Ω	0.07 % to 0.6 % 0.6 % to 2.3 % 2.3 % to 6.3 %	
23.	Frequency*	100 Hz to 10 MHz 10 MHz to 250 MHz	0.002 % to 0.0023 % 0.0023 % to 0.004 %	Using Multi Product Calibrator by Direct Method
24.	Oscilloscope* Amplitude Time Base Bandwidth	2 mV/Div to 20 V/Div 20 ns to 1 s 10 MHz to 250 MHz ref 50 kHz	0.1 % to 0.013 % 0.03 % to 0.012 % 1.3 dB	Using Multi Product Calibrator by Direct Method
25.	AC Power@ 50 Hz * 1 ϕ UPF	100 V to 300 V 1 A to 20 A 100 W to 6 kW	0.35 % to 1.2 %	Using Multi Product Calibrator by Direct Method
26.	Capacitance*	1 kHz 1 nF to 100 μ F	1.7 % to 5.78 %	Using Multi Product Calibrator by Direct Method
II.	MEASURE			
	DC Voltage #	1 mV to 100 mV 100 mV to 1000 V	0.054 % to 0.008 % 0.008 % to 0.002 %	Using 8½ Digital Multimeter by Direct Method
	DC High Voltage #	1 kV to 35 kV	2.7 %	High Voltage Probe with 8½ Digital Multimeter by Direct Method
	AC Voltage #	50 Hz to 1 kHz 50 mV to 100 V 100 V to 1000 V 1 kHz to 10 kHz 100 mV to 700 V	0.2 % to 0.04 % 0.04 % to 0.051 % 0.08 %	Using 8½ Digital Multimeter by Direct Method

Sangeeta Kunwar
Convenor

Avijit Das
Program Manager

Laboratory Calsytech, No. 38, North Mada Street, Nandambakkam, Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2193

Page 6 of 7

Validity 10.10.2018 to 09.10.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	AC High Voltage #	50 Hz 1 kV to 28 kV	6 %	High Voltage Probe with 8½ Digital Multimeter by Direct Method
3.	DC Current #	1 μ A to 10 μ A 10 μ A to 100 μ A 100 μ A to 100 mA 100 mA to 1 A 1 A to 20 A 20 A to 100 A	0.046 % to 0.05 % 0.05 % to 0.002 % 0.002 % to 0.005 % 0.005 % to 0.02 % 0.02 % to 0.09 % 0.7 %	Using 8½ Digital Multimeter by Direct Method Using Precision Shunt with Digital Multimeter by Direct Method
4.	AC Current #	50 Hz to 1 kHz 20 μ A to 1 A 1 A to 20 A 50 Hz 20 A to 100 A	0.15 % to 0.058 % 0.058 % to 0.16 % 0.93 %	Using 8½ Digital Multimeter by Direct Method Using Precision Shunt with Digital Multimeter by Direct Method
5.	DC Resistance #	1 Ω to 100 Ω 100 Ω to 1 M Ω 1 M Ω to 100 M Ω 100 M Ω to 100 G Ω	0.003 % to 0.0012 % 0.0012 % to 0.0016 % 0.0016 % to 0.015 % 0.05 % to 0.5 %	Using 8½ Digital Multimeter by Direct Method
6.	Timer #	100 ms to 1000 s	0.34 %	Using Timer by Direct Method
	Temperature #			
	Thermocouple "K" Type	(-) 180 °C to 1200 °C	0.51 °C	
	Thermocouple "J" Type	(-) 140 °C to 1340 °C	0.61 °C	
	Thermocouple "B" Type	300 °C to 1800 °C	0.52 °C	
	Thermocouple			

Sangeeta Kunwar
Convenor

Avijit Das
Program Manager

Laboratory Calsytech, No. 38, North Mada Street, Nandambakkam, Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2193

Page 7 of 7

Validity 10.10.2018 to 09.10.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	"E" Type Thermocouple	0 °C to 800 °C	0.31 °C	
	"N" Type Thermocouple	(-) 180 °C to 1300 °C	0.32 °C	
	"R" Type Thermocouple	0 °C to 1700 °C	1.2 °C	
	"S" Type Thermocouple	0 °C to 1700 °C	0.58 °C	
	"T" Type Thermocouple	0 °C to 400 °C	0.41 °C	
	RTD	(-) 100 °C to 800 °C	0.07 °C	
8.	AC Power@ 50 Hz [§] UPF 1 ϕ	10 V to 600 V 1 A to 20 A 10 W to 12 kW	0.42 % to 0.34 %	Using Digital Power Meter by Direct Method
9.	Frequency [§]	0.1 kHz to 200000 kHz	0.84 % to 0.004 %	Using Universal Counter by Direct Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

[§]Only in Permanent Laboratory

*Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Sangeeta Kunwar
Convenor

Avijit Das
Program Manager