

Laboratory Centre of Excellence for Electronic Products (CEEP), Plot No. 638,
GIDC Makarpura, Vadodara, Gujarat

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2952

Page 1 of 6

Validity 16.02.2019 to 15.02.2021

Last Amended on 14.03.2019

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>ELECTRO TECHNICAL CALIBRATION</u>				
I	MEASURE			
1.	Voltage Dips & interruption[#]			
	a) Generator Output Voltage at NO Load	0 to 100%	2.60%	Using Oscilloscope with HV probe and Digital Multimeter as per IEC 61000-4-11 Ed 2.0
	b) Dips/ interruption Time	10 ms to 5 s	3.00%	Using Oscilloscope with HV probe as per IEC 61000-4-11 Ed 2.0
2.	Combination Wave Surge Test[#]			
	Voltage Amplitude	± 0.5 kV to ± 8.0 kV	3.1%	Using Oscilloscope and HV Differential Probe as per IEC 61000-4-5 Ed 3.0
	Voltage Front Time	1.2 μ s $\pm 30\%$ 10 μ s $\pm 30\%$	7.8% 2.28%	
	Voltage Pulse Duration	50 μ s $\pm 20\%$ 700 μ s $\pm 20\%$	3.10% 1.10%	
	Current Amplitude	± 0.25 kA to ± 4 kA ± 12.5 A to ± 200 A	3.10%	Using Oscilloscope and Current Clamp as per IEC 61000-4-5 Ed 3.0
	Current Front Time	8 μ s $\pm 20\%$ 5 μ s $\pm 20\%$	3.12% 3.48%	
	Current Pulse Width	20 μ s $\pm 20\%$ 320 μ s $\pm 20\%$	3.41% 1.0%	

Shally Sharma
Convenor

Battal Singh
Program Manager

Laboratory Centre of Excellence for Electronic Products (CEEP), Plot No. 638, GIDC Makarpura, Vadodara, Gujarat

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2952

Page 2 of 6

Validity 16.02.2019 to 15.02.2021

Last Amended on 14.03.2019

"In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020"

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
3.	Ring Wave Generator 12 Ω and 30 Ω[#]			
	Voltage Rise Time (Open Circuit)	0.5 μ s \pm 30%	1.0%	Using CRO and HV Differential Probe as per IEC 61000-4-12
	Current Rise Time (Short Circuit)	0.2 μ s to 1.0 μ s	1.0%	Using CRO and Current Probe as per IEC 61000-4-12
	Frequency	100 kHz \pm 10%	2.11%	Using CRO and HV Differential Probe as per IEC 61000-4-12
	Decay in Voltage	0.4 < Ratio of Pk2 to Pk1 < 1.1 0.4 < Ratio of Pk3 to Pk2 < 0.8 0.4 < Ratio of Pk4 to Pk3 < 0.8	4.42%	
	Repetition rate	1/ min or faster	2.0%	
	Open Circuit Voltage	250 V to 6 kV \pm 10%	3.05%	
	Short Circuit Current	20.8 A to 500 A \pm 10% @ 12 Ω 8.3 A to 200 A \pm 10% @ 30 Ω	2.58%	Using CRO and Current Probe as per IEC 61000-4-12
4.	Electrical Fast Transient Test System, With Cdn And Capacitive Coupling Clamp[#]			
	Pulse Amplitude	\pm 0.25 kV to \pm 5.0 kV \pm 10%	3.29%	Using CRO & Dividers (50 Ω & 1000 Ω) as per IEC 61000-4-4
	Pulse Rise Time	5 \pm 1.5 ns	3.93%	
	Pulse Width	50 \pm 15 ns	2.44%	
	Repetition Frequency	5 kHz \pm 20 %, 100 kHz \pm 20 %, 1MHz \pm 20 %	1.0%	
	Burst Duration	15 \pm 3 ms	1.0%	
	Burst Period	300 \pm 60 ms	1.0%	
	Pulse Amplitude in 1000 Ω	\pm 0.25 kV to \pm 5.0 kV \pm 20%	3.29%	

Shally Sharma
Convenor

Battal Singh
Program Manager

Laboratory Centre of Excellence for Electronic Products (CEEP), Plot No. 638, GIDC Makarpura, Vadodara, Gujarat

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2952

Page 3 of 6

Validity 16.02.2019 to 15.02.2021

Last Amended on 14.03.2019

"In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020"

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	Pulse Width in 1000 Ω	50ns (-15 ns to +100 ns)	2.44%	
	Pulse Amplitude Capacitive coupling Clamp	2.0 kV \pm 200V	3.29%	
	Peak voltage -CDN	\pm 0.25 kV to \pm 5.0 kV \pm 10%	3.29%	
	Pulse Width-CDN	(45 \pm 15) ns.	2.44%	
	Pulse Rise Time-CDN	(5.5 \pm 1.5) ns	3.93%	
5.	Electrostatic Discharge Generator^s			
	Contact Discharge			
	First Peak Current	\pm 7.5 A to \pm 112.5 A (\pm 15 %) (\pm 2 kV to \pm 30 kV)	3.78%	Using CRO & ESD Target Attenuator Cable Chain as per IEC 61000-4-2
	Rise Time	0.8 ns \pm 25 %	8.21%	
	Current at 30 ns	\pm 4.0A to \pm 60 A (\pm 2 kV to \pm 30 kV)	3.79%	
	Current at 60 ns	\pm 2.0 A to \pm 30 A (\pm 2 kV to \pm 30 kV)	3.79%	
	Air Discharge			
	Generator DC output voltage	\pm 2 kV to \pm 30 kV (\pm 5%)	3.44%	Using DMM with Divider as per IEC 61000-4-2
6.	Target – Attenuator Cable Chain^s			
	Low Frequency Transfer Impedance	0.02 Ω to 0.2 Ω \pm 1%	1%	Using DMM or VNA as per IEC 61000-4-2
	DC Resistance	< 2.1 Ω	0.73%	Using DMM as per IEC 61000-4-2
7.	Impedance^s			
	Line Impedance Stabilization Network- LISN, Coupling & Decoupling-CDN	9 kHz to 250 MHz 1 Ω to 300 Ω	5.0% to 10.1%	Using Vector Network Analyzer by Direct Method as per CISPR 16-1-2: Ed. 2.0 2017-11, CISPR 32

Shally Sharma
Convenor

Battal Singh
Program Manager

Laboratory Centre of Excellence for Electronic Products (CEEP), Plot No. 638,
GIDC Makarpura, Vadodara, Gujarat

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2952

Page 4 of 6

Validity 16.02.2019 to 15.02.2021

Last Amended on 14.03.2019

"In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020"

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
8.	Phase Angle^s LISN	9 kHz to 250 MHz (-) 90° to $+90^{\circ}$	3.1°	Using Vector Network Analyzer by Direct Method as per CISPR 16-1-2, CISPR 32
9.	Power Frequency Magnetic Field Immunity*			
	Current : Continuous & Short Duration- 1s to 3s, Current Values Standard Coil:1 m X 1m,1mX2.6m	1 A/m to 1000 A/m $\pm 2.0\%$	2%	Using Current clamp with DMM and Clamp on Meter As per IEC 61000-4-8
10.	Impulse Magnetic Field[#]			
	Short Circuit Peak Current	100 A to 1000 A $\pm 10\%$	2.38%	Using CRO & Current Probe As per IEC 61000-4-9
	Current Front Time:	$8 \mu s \pm 20\%$	2.02%	
	Duration	$20 \mu s \pm 20\%$	1.0%	
11.	Frequency^s	9kHz to 15 GHz	1.17ppm	Using Vector Network Analyzer By Direct Method
12.	RF Parameters^s			
	Attenuation, Voltage Division Factor / Coupling Decoupling Factor/Isolation,	9 kHz to 15GHz 0.1 dB to 70 dB	0.4 dB	Using Vector Network Analyzer By Direct Method
	Insertion loss/ Bulk Current Injection, Directional Coupler, Attenuator, Adapter LISN, Impedance Stabilization Network , ESD Target, Resistive HV Dividers, EM Clamp, Cable	9 kHz to 3.2 GHz 0.1 dB to 70 dB	0.25 dB	Using Signal Generator and Power meter By Direct Method

Shally Sharma
Convenor

Battal Singh
Program Manager

Laboratory Centre of Excellence for Electronic Products (CEEP), Plot No. 638,
GIDC Makarpura, Vadodara, Gujarat

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2952

Page 5 of 6

Validity 16.02.2019 to 15.02.2021

Last Amended on 14.03.2019

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks

13.	RF Power Amplifier^s			
	Gain	9 kHz to 6 GHz 5 dB to 65 dB	0.4 dB to 0.56 dB	Using Signal Generator, Power Meter By Direct Method
	Harmonics Level	9 kHz to 15GHz 5 dBc to 80 dBc	0.52dB to 0.7 dB	Using Vector Network Analyzer With Spectrum By Direct Method
14.	Amplitude Modulation ^s	9kHz to 6GHz +15dBm to -60dBm	0.38 dB to 0.56 dB	Using Signal Generator Power Sensor By Direct Method
		9kHz to 6GHz +60dBm	0.56 dB	Using Signal Generator Power Sensor with Attenuator By Direct Method
14.	Amplitude Modulation ^s	9kHz to 3.2 GHz 10% to 90%	3.30%	Using CRO By Direct Method
		3.2 GHz to 15 GHz 10% to 90%	4.20%	Using Vector Network Analyzer With Spectrum By Direct Method

Shally Sharma
Convenor

Battal Singh
Program Manager

Laboratory Centre of Excellence for Electronic Products (CEEP), Plot No. 638,
GIDC Makarpura, Vadodara, Gujarat

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2952

Page 6 of 6

Validity 16.02.2019 to 15.02.2021

Last Amended on 14.03.2019

"In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020"

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
15.	Return Loss (VSWR)/ Attenuator, Antenna, Amplifier, Termination, RF Load, Directional Coupler, Adapter Line	9 kHz to 15GHz 1 dB to 46 dB	0.40dB	Using Vector Network Analyzer By Direct Method
		9 kHz to 3.2 GHz 1 to 46 dB	0.35dB	Using Signal Generator , Power meter / Vector Network Analyzer With Spectrum & Directional Coupler By Direct Method
16.	Antenna Factor (Antenna)[§]			
	2M Loop Antenna, Monopole/ ROD Antenna	9 kHz to 30 MHz	1.0 dB	Using Signal Generator, Vector Network Analyzer With Spectrum as per CISPR 16-1-4
17.	Transfer Impedance – RF Current Clamp [§]	10 kHz to 500 MHz	0.4dB	Using Vector Network Analyzer By Direct Method
II.	SOURCE			
1.	Frequency [#]	20 Hz to 3.2GHz	3.1 ppm to 1.16ppm	Using Signal Generator by Direct Method
2.	Power Measurement [#]	9kHz to 3.2GHz +15dBm to -60dBm	0.25 dB to 0.35 dB	Using Signal Generator and Power Sensor by Comparison method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

§ Only in Permanent Laboratory

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Shally Sharma
Convenor

Battal Singh
Program Manager