

Laboratory	Chetas Water Flow Laboratory, Shed No. 6 & 7, Pandhari Industrial Estate, Shivane, Pune, Maharashtra		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Fluid Flow Calibration	Issue Date	04.10.2014
Certificate Number	C-0861	Valid Until	03.10.2016
Last Amended on	20.10.2014	Page	1 of 1

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
I. FLOW MEASUREMENT			
1. Flow Rate^{\$} (Water)	0.4 m ³ /hr to 310 m ³ /hr	0.4% rdg	Using 2 Ton & 6 Ton Gravimetric System and Test Rings as per ISO 4185

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

^{\$}Only in Permanent Laboratory

Shally Sharma
Convenor

Avijit Das
Program Manager