

Laboratory CLICK-IN Calibration Laboratory, S. No. 44/3, Flat No. 11, Satyam Plaza,
Sinhagad College Campus, Vadgaon Bk., Pune, Maharashtra

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 19.07.2014

Certificate Number C-0839 **Valid Until** 18.07.2016

Last Amended on - **Page** 1 of 3

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>SOURCE</u>			
1. DC Voltage [#]	1 mV to 200 mV 200mV to 20 V 20V to 1000 V	3.77% to 0.12% 0.12% to 0.14% 0.14% to 0.26%	Using Zeal 5½Digit Multifunction Calibrator by Direct Method
2. AC Voltage [#]	50 Hz 40 mV to 200 mV 200 mV to 20 V 20V to 990 V	1.0% to 0.44 % 0.44% to 0.19 % 0.19% to 0.21 %	Using Zeal 5½Digit Multifunction Calibrator by Direct Method
3. DC Current [#]	0.1mA to 20 mA 20 mA to 2A 2A to 10 A 10 A to 990 A	2.3% to 0.37% 0.37% to 0.24% 0.24% to 0.20% 0.73%	Using Zeal 5½Digit Multifunction Calibrator with current coil by Direct Method
4. AC Current [#]	50 Hz 0.1mA to 20 mA 20 mA to 2A 2A to 10 A 10 A to 990 A	5.25% to 0.40% 0.40% to 0.34% 0.34% to 1.40% 2.60%	Using Zeal 5½Digit Multifunction Calibrator with current coil by Direct Method
5. Capacitance [#]	1 kHz 500 pF to 1 μ F	2.90%	Using AGRONIC CDB 6 Capacitance Box by Direct Method
6. Resistance [#]	1 Ω to 1000 Ω 1000 Ω to 1 M Ω	1.47% 1.16%	Using AGRONIC Decade Resistance Box RDB 6 by Direct Method

Avijit Das
Program Manager

Sangeeta Kunwar
Convenor

Laboratory **CLICK-IN Calibration Laboratory, S. No. 44/3, Flat No. 11, Satyam Plaza, Sinhagad College Campus, Vadgaon Bk., Pune, Maharashtra**

Accreditation Standard **ISO/IEC 17025:2005**

Discipline **Electro-Technical Calibration** Issue Date **19.07.2014**

Certificate Number **C-0839** Valid Until **18.07.2016**

Last Amended on **-** Page **2 of 3**

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
7. Temperature Indicator/ Recorder / Controller[#]			
RTD (Input)	-200°C to 850°C	0.59°C to 0.82°C	Using Universal calibrator (Radix Model-Microcal) by Simulation Method
T/C J Type	-200 °C to 700°C	0.6°C to 1.6°C	
T/C K Type	-200 °C to 1200°C	0.78°C to 1.6°C	
T/C R Type	150 °C to 1700°C	0.6°C to 1.6°C	
T/C S Type	150 °C to 1700°C	0.6°C to 1.6°C	
<u>MEASURE</u>			
8. DC Voltage[#]			
	1 mV to 100 mV	0.85% to 0.01%	Using 6½ Digital Multimeter by Direct Method
	100 mV to 10 V	0.10%	
	10 V to 1000 V	0.10%	
9. DC High Voltage[*]			
	0.5 kV to 10kV	3.5%	Using Zeal HV Probe with 4½ Digital Multimeter by Direct Method
10. AC Voltage[#]			
	50 Hz		Using 6½ Digital Multimeter by Direct Method
	100 mV to 10 V	0.12% to 0.11%	
	10 V to 1000 V	0.20%	
11. AC High Voltage[*]			
	0.5 kV to 10 kV	3.5%	Using Zeal HV Probe with 4½ Digital Multimeter by Direct Method
12. DC Current[#]			
	0.1mA to 20 mA	0.58%	Using 6½ Digital Multimeter by Direct Method
	20 mA to 2A	0.58% to 0.19%	
	2A to 10 A	0.19% to 0.22%	

Laboratory CLICK-IN Calibration Laboratory, S. No. 44/3, Flat No. 11, Satyam Plaza,
Sinhagad College Campus, Vadgaon Bk., Pune, Maharashtra

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 19.07.2014

Certificate Number C-0839 **Valid Until** 18.07.2016

Last Amended on - **Page** 3 of 3

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
13. AC Current [#]	50 Hz		
	0.2mA to 20 mA	0.3% to 0.17%	Using 6½ Digital Multimeter by Direct Method
	20 mA to 2A	0.17% to 0.29%	
2A to 10 A	0.29%		
14. Resistance [#]	1Ω to 10 Ω	0.69%	Using 6½ Digital Multimeter by Direct Method
	10 Ω to 10 kΩ	0.69% to 0.20%	
	10 kΩ to 1 MΩ	0.20%	
15. Frequency [#]	10 Hz to 3 kHz	0.09% to 0.019%	Using 6½ Digital Multimeter by Direct Method
16. Time [#]	10 sec to 3600 sec	0.32 sec to 3.0 sec	Using Race make Digital Stop Watch by Comparison Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

*Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used