

Laboratory **CLICK-IN Calibration Laboratory, S. No. 44/3, Flat No. 11, Satyam Plaza,
Sinhagad College Campus, Vadgaon Bk., Pune, Maharashtra**

Accreditation Standard **ISO/IEC 17025:2005**

Discipline **Thermal Calibration** Issue Date **19.07.2014**

Certificate Number **C-0841** Valid Until **18.07.2016**

Last Amended on **-** Page **1 of 1**

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
TEMPRATURE			
1. Temperature # (RTD, Thermocouple, with / without Indicator/ Dial Type Thermometer)	50°C to 200°C	1.0°C	Using RTD Sensor, Universal Calibrator & Tem. Bath by Comparison Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used