

Laboratory Dr. Reddy's FTO-3 Calibration Laboratory, Survey No: 41, Bachupally,
 Qutubullapur, R. R. Dist., Telangana
Accreditation Standard ISO/IEC 17025:2005
Discipline Thermal Calibration **Issue Date** 19.08.2014
Certificate Number C-1103 **Valid Until** 18.08.2016
Last Amended on - **Page** 1 of 1

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
I. TEMPERATURE			
1. Temperature Indicator with Sensor (RTD)[#]	-15 °C to 150 °C	0.14 °C	Using Temperature Indicator with RTD Sensor and Dry Block Calibrator by Comparison Method
	150 °C to 400 °C	0.83 °C	

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

[#] The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.