

Laboratory	Electrometer Corporation , 34, Main Patel Nagar Road, New Delhi		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	01.12.2016
Certificate Number	C-1478	Valid Until	30.11.2018
Last Amended on	09.12.2016	Page	1 of 4

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>SOURCE</u>			
1. D C VOLTAGE[§]	1 mV to 100 mV	1.37% to 0.027%	Using Fluke 5080A Calibrator by Direct Method
	100 mV to 1 V	0.027% to 0.015%	
	1 V to 10 V	0.015% to 0.016%	
	10 V to 100 V	0.016% to 0.024%	
	100 V to 1000 V	0.024% to 0.016%	
2. D C CURRENT[§]	10 μ A to 100 μ A	1.22% to 0.20%	Using Fluke 5080A Calibrator by Direct Method
	100 μ A to 1 mA	0.20% to 0.10%	
	1 mA to 100 mA	0.10% to 0.076%	
	100 mA to 1 A	0.076% to 0.20%	
	1 A to 10 A	0.20% to 0.15%	
D C HIGH CURRENT[§]	10 A to 20 A	0.15% to 0.59%	with 50 turn coil
	100 A to 1000 A	0.59% to 1.44 %	
3. RESISTANCE[§]	1 Ω to 10 Ω	2.15% to 0.17%	Using Fluke 5080A Calibrator by Direct Method
	10 Ω to 100 Ω	0.17% to 0.046%	
	100 Ω to 1 k Ω	0.046% to 0.028%	
	1 k Ω to 100 k Ω	0.028% to 0.044%	
	100 k Ω to 1 M Ω	0.044% to 0.048%	
	1 M Ω to 10 M Ω	0.048% to 0.113%	
	10 M Ω to 19 M Ω	0.113% to 0.177%	
4. A C VOLTAGE[§]	50 Hz		Using Fluke 5080A Calibrator by Direct Method
	30 mV to 100 mV	0.612% to 0.25%	
	100 mV to 1 V	0.25% to 0.17%	
	1 V to 10 V	0.17% to 0.16%	
	10 V to 100 V	0.16% to 0.19%	
100 V to 1000 V	0.19%		
5. FREQUENCY[§]	45Hz to 1000Hz	0.061%	Using Fluke 5080A Calibrator by Direct Method

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory	Electrometer Corporation , 34, Main Patel Nagar Road, New Delhi		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	01.12.2016
Certificate Number	C-1478	Valid Until	30.11.2018
Last Amended on	09.12.2016	Page	2 of 4

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
6. A C CURRENT ^s	50 μ A to 1mA 1 mA to 100 mA 100 mA to 1 A 1 A to 10 A	2% to 0.59% 0.59% to 0.37% 0.37% to 0.42% 0.42% to 0.61%	Using Fluke 5080A Calibrator by Direct Method
A C HIGH CURRENT ^s	100 A to 1000 A	1.63% to 1.33%	with 50 turn coil
7 TEMPERATURE SIMULATION ^s (INDICATOR/ CONTROLLER / RECODER)			
K Type	(-)200 °C to 0 °C 0 °C to 1300 °C	0.83 °C to 0.30 °C 0.30 °C to 0.55 °C	Using Fluke 5080A Calibrator by Direct Method
J Type	(-)200 °C to 0 °C 0 °C to 1200 °C	0.44 °C to 0.24 °C 0.24 °C to 0.38 °C	
E Type	(-)200 °C to 0 °C 0 °C to 1000 °C	0.81 °C to 0.2 °C 0.2 °C to 0.31 °C	
T Type	(-)200 °C to 0 °C 0 °C to 400 °C	1.03 °C to 0.3 °C 0.3 °C to 0.24 °C	
N Type	(-)200 °C to 0 °C 0 °C to 1300 °C	0.15 °C to 0.44 °C 0.44 °C to 0.51 °C	
R Type	0 °C to 1700 °C	2.27 °C to 1.1 °C	
S Type	0 °C to 1700 °C	2.27 °C to 1.17 °C	
B Type	600 °C to 1700 °C	1.93 °C to 1.11 °C	

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory	Electrometer Corporation , 34, Main Patel Nagar Road, New Delhi		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	01.12.2016
Certificate Number	C-1478	Valid Until	30.11.2018
Last Amended on	09.12.2016	Page	3 of 4

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>MEASURE</u>			
1. D C VOLTAGE[§]	1 mV to 100 mV	0.477% to 0.011%	Using Data Acquisition Switch Unit by Direct Method
	100 mV to 1 V	0.011% to 0.009%	
	1 V to 10 V	0.009% to 0.01%	
	10 V to 100 V	0.01% to 0.007%	
	100 V to 300 V	0.007% to 0.09%	
2. D C CURRENT[§]	1 mA to 10 mA	0.28% to 0.08%	Using Data Acquisition Switch Unit by Direct Method
	10 mA to 100m A	0.08% to 0.22%	
	100 mA to 1 A	0.22% to 0.32%	
3. RESISTANCE[§]	2 Ω to 100 Ω	2.15% to 0.017%	Using Data Acquisition Switch Unit by Direct Method
	100 Ω to 1 k Ω	0.017% to 0.046%	
	1 k Ω to 10 k Ω	0.046% to 0.028%	
	10 k Ω to 100 k Ω	0.028% to 0.044%	
	100 k Ω to 1 M Ω	0.044% to 0.049%	
4. A C VOLTAGE[§]	50 Hz		Using Data Acquisition Switch Unit by Direct Method
	1 mV to 100 mV	4.72% to 0.17%	
	100 mV to 1 V	0.17% to 0.12%	
	1 V to 10 V	0.12%	
	10 V to 100 V	0.12%	
100 V to 300 V	0.12% to 0.166%		
5. FREQUENCY[§]	45Hz to 1000Hz	0.037% to 0.016%	Using Data Acquisition Switch Unit by Direct Method
6. A C CURRENT[§]	1 mA to 10 mA	0.64% to 0.41%	Using Data Acquisition Switch Unit by Direct Method
	10 mA to 100 mA	0.41% to 0.73%	
	100 mA to 1 A	0.73% to 0.59%	

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory	Electrometer Corporation , 34, Main Patel Nagar Road, New Delhi		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	01.12.2016
Certificate Number	C-1478	Valid Until	30.11.2018
Last Amended on	09.12.2016	Page	4 of 4

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
7 TEMPERATURE SIMULATION^{\$} (Indicator/ Controller / Recorder)			
K Type	0°C to 1200°C	0.58°C to 0.63°C	Using Data Acquisition Switch Unit by Direct Method
J Type	0°C to 1200°C	0.58°C	
T Type	0°C to 400°C	0.58°C	
R Type	0°C to 1700°C	1.22°C to 0.9°C	
S Type	0°C to 1700°C	1.22°C to 0.93°C	
Pt-100	(-200°C to 600°C	0.06°C to 0.20°C	

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

^{\$} Only in Permanent Laboratory

Vishal Shukla
Convenor

Avijit Das
Program Manager